



अखिल भारतीय आयुर्विज्ञान संस्थान, रायपुर (छ.ग.)
All India Institute of Medical Sciences, Raipur (Chhattisgarh)
Tatibandh, GE Road, Raipur-492 099 (CG)

No. AIIMS/R/CS/Pul_Med/18/035/OT
Tender ID- 2019_IMSRP_477497_1

Raipur, Dtd:-10/07/2019

Tender Name – Supply of Broadbased QR for Body Plethysmograph with Diffusion Study for Dept of Pulmonary Medicine at AIIMS Raipur.

Corrigendum

| Subheading of Technical Specification on Annexure - I | Existing Sentence | To be read as |
|---|---|--|
| Body Box | The seat must be ergonomic, comfortable and adjustable in height to bear weight up to 150 kg. | The seat must be ergonomic, comfortable and adjustable/fixed in height to bear weight up to 150 kg. |
| Flow meter | Flowmeters: Pneumotachograph with bidirectional precision. | Pneumotachograph and / or ultrasonic with Bidirectional precision. |
| Essential Accessories | Refilling Of DLCO Gas : 02 no's | To be deleted. |
| Additional to be added in warranty and CMC | | The Company Should Provide the rate of consumables separately and should be valid for warranty period. |

Note :-

1. Technical specification remains the same as earlier except for above corrigendum.
2. All other terms and condition will remain unchanged.

**Stores Officer
AIIMS Raipur (CG)**