



**Date: - 10/07/2019**

**Corrigendum**

**विषय/Sub:** Corrigendum in Notice Inviting Tender for “Supply of Electro-Hydraulic OT Table” for Urology Department at AIIMS, Raipur

**संदर्भ/Ref:** Tender No. AIIMS/R/CS/URO/18/31/OT/A

In context to following corrigendum is issued against subject tender:

Sr. No	Page/Clause/ Point	Existing Sentence	To be read as/ Addendum
01	Page 17, Point No. 18	<b>A. General operating table features:</b> 13. It should have a stable mobile base with lockable antistatic castors. Table should be locked at the head/ foot end via a central foot pedal/ Hand control. It should be able to perform 360-degree rotation on castors.	13. It should have a stable mobile base with hydraulic locking system through hand held controller & antistatic castors for easy maneuverability. Table should be locked at the head/ foot end via a central foot pedal/ Hand control. It should be able to perform 360-degree rotation on castors.
		Nil	<b>Added:</b> 15. Should have kidney elevation bridge of 70mm of more between back plate & Seat Plate.
		Nil	<b>Added:</b> 16. All adjustments of electrohydraulic powered motions are controlled electrically from outside intervention area via cable connected hand control with backlit and/or display.
02	Page 18, Point No. 9	<b>B. Table Top Adjustments:</b> 9. Head plate adjustment (raising/ lowering), detachable: - at least 60 degree up, at least 90 degree down	9. Head plate adjustment (raising/ lowering), detachable: - at least 45 degree up, at least 90 degree down

Note: All other terms & condition will be remain unchanged.

**Stores Officer  
AIIMS Raipur (CG)**