



AIIMS/R/CS/Neuro/ 75/18/OT /corrigendum/

दिनांक:- 26.09.2018

Corrigendum-2

Tender ID No.: 2018_IMSRP_367432_1

With Reference to above tender ID No., the following amendment is being issued:-

Page No./Clause no/point no.	Existing sentence	To be read as/Addendum
Page No. 07 point No. 10, Annexure-I & II	5 year warranty and 5 years AMC should be included with provision of supply of spare parts items for next 5 years.	5 year warranty should be included (AMC Clause has been removed from NIT Document)
Page No. 11 L-1 Point	L-1 = Total Price of set (Both Unit) + AMC of set (Both Unit) (L1 will be decided on Total Price of Set FoR AIIMS Raipur + AMC of 5 years)	L-1 = Total Price of set (Both Unit) (L1 will be decided on Total Price of Set FoR AIIMS Raipur)
Page No. 12	PRICE BID (a) Price bid in the form of BOQ_XXXX.xls (b) Annual Maintenance contract – AMC.pdf (c) Custom Duty Tariff – CDT.pdf Note:- Bidder are supposed to quote for all items of above said set.	PRICE BID (a) Price bid in the form of BOQ_XXXX.xls (b) Custom Duty Tariff – CDT.pdf Note:- Bidder are supposed to quote for all items of above said set.
Technical Specification of Retractor system (Annexure-I)		
Point 2	System should contain two frames – one for medial/lateral & one for cranial/caudal. The same should be marked on the frames to avoid any confusion. And should have 2 possible connections to flex arm.	System should contain two frames – one for medial/lateral & one for cranial/caudal. The same should be marked on the frames to avoid any confusion. And should have possible connections to flex arm.
Point 5	System should have cranial/caudal blades from 40 – 100mm in 10mm increments and width of blades should be 20mm. Also blades should have Right and left inscriptions to avoid and confusion and waste of time	System should have cranial/caudal blades from 40 – 80mm in 10mm increments and M/L blade narrow 5, 7, 9cm (2 each) M/L blade wide 5, 7, 9 cm (2 each) left blade & right blade 40, 50, 60, 70, 80mm (each). Narrow width of blades should be 20mm. Also blades should have Right and left inscriptions to avoid and confusion and waste of time
Point 6	Cranial/caudal blades should form a tube of 19mm in diameter	Cranial/caudal blades should form a tube of 19-22 mm in diameter
Point 8	System should achieve higher distraction depending on the length of cranial/caudal blade for e.g. 50 mm of blades can achieve maximum expansion of 115mm and 80mm of blade can achieve 145 mm of expansion.	On full expansion the proximal working opening should be at least 50mm and distal opening to be at least 70mm and max up to 120mm



Point 9	System should have medial/lateral blades from 40 – 110mm in 10mm increments and width of blades should be 40mm (for wide) & 25mm (for narrow). Also blades should have size and width inscriptions to avoid and confusion and waste of time the medical/lateral blades should have teeth to avoid tissue slipping	System should have medial/lateral blades from 50 – 90mm in 20mm increments and width of blades should be 40mm (for wide) & 25mm (for narrow). Also blades should have size and width inscriptions to avoid and confusion and waste of time the medical/lateral blades should have teeth to avoid tissue slipping
Point 13	System should achieve higher distraction depending on the length of medial/lateral blade for e.g. 40 mm of blades can achieve maximum expansion of 90mm and 90mm of blade can achieve 140 mm of expansion	On full expansion the proximal working opening should be at least 50mm and distal opening to be at least 70mm and max up to 120mm
Point 14	All the blades in the system should be radiolucent, black & have non-glare coating.	All the blades in the system should be non-glare
Changes proposed in accessories (Annexure-I)		
Point 2	2 retractor handle	1 retractor handle and 1 medio lateral frame
Point 13-17	Medial/lateral blade, 40 mm 1 Medial/lateral blade, 50 mm 1 Medial/lateral blade, 60 mm 1 Medial/lateral blade, 70 mm 1	Removed Medial/lateral blade, narrow, 50 mm 2 Medial/lateral blade, narrow, 60 mm 2 Medial/lateral blade, narrow, 70 mm 2
Point 32	Instrument tray	Modular Tray 1

Annexure II

General specifications and terms for Cervical Retractor and Distractor

Page No./Clause no/point no.	Existing sentence	To be read as/Addendum
Page no. 20 Point 2	5 year warranty and 5 years AMC should be included with provision of supply of spare parts items for next 5 years.	5 year warranty should be included (AMC Clause has been removed from NIT Document)
Technical Specification for Cervical Retractor and Distractor System		
Page no. 20 Sr. no. 2	Ports with inner diameters of 12 mm, 15 mm, 18 mm, 21 mm and 24 mm and lengths ranging from 30 mm to 110 mm in 10 mm increments. The tip configuration for these ports should be straight and anatomic (beveled) in every size	5 Tube retractor with inner diameters ranging from 12-25 with increment of 2-3 between each retractor lengths ranging from 30 mm to 90 mm in 10 mm increments. The tip configuration for these tube retractors should be straight and anatomic (beveled) in every size