



AIIMS/R/CS/Nephro/340/19/OT/corrigendum

दिनांक:- 16.01.2020

Corrigendum

Tender ID No.: 2019_IMSRP_528260_1

With Reference to above tender ID No., the following amendment is being issued for the tender supply of **“Biomedical Liquid Waste Treatment System”** for the Department of Nephrology at AIIMS Raipur:-

Annexure I

S.no.		Existing sentence	To be read as/Addendum
1.	Annexure-I Point No. 20 of Page no. 19 of NIT	The equipment have efficacy test reports by Shriram Institute for Industrial Research, New Delhi with 99.99% efficacy.	Deleted
2.	Annexure-I Point No. 21 of Page no. 19 of NIT	European CE and EN ISO 13485:2016, ISO 9001:2015 & OHSAS ISO 18001:2007 Certified.	CE/BIS/ISO standard applicable
3.	Page No. 8 Point. No. 5	The Supplier should arrange for regular weekly visit to the AIIMS, Raipur campus by its technical team and assist in maintenance of the item/equipment within warranty period.	During the warranty period the visit shall be in six months or as and when required during breakdown from time to time.
4.	Added	NIL	Equipment capacity shall be 150 Ltr

Sr. Administrative Officer
AIIMS Raipur