

अखिल भारतीय आयुर्विज्ञान संस्थान ,रायपुर (छत्तीसगढ़) All India Institute of Medical Sciences, Raipur (Chhattisgarh)

No. AIIMS/R/CS/Nu.Med./18/993/OT

Tatibandh, GE Road, Raipur-492 099 (CG) <u>www.aiimsraipur.edu.in</u> Raipur, Dtd:-16.02.2019

Corrigendum

Tender ID.No.2019_IMSRP_434275_1, Dated:25.01.2019.

With reference to above Tender ID.No., the following amendment is issued:

S.N 0.	Page no in the NIT	Existing Tender Specification	To be read as
1.	Page.No.5 Point.No.27.		Earnest money by means of a Bank Demand Draft/ FD/ BG , a scanned copy to be enclosed. It is also clarified that the bids submitted without earnest money will be summarily rejected. The DD/FD may be prepared in the name of "All India Institute of Medical Sciences, Raipur (AIIMS RAIPUR)". The EMD cost must reach at office of the Stores Officer Gate no. 5, Medical College Building, 2nd Floor, AIIMS, Raipur before opening of tender.
2.	Page no 7 Point.No.3	Delivery& Installation (if applicable) : The successful bidder should strictly adhere to the following delivery schedule supply, installation & commissioning should be effected within 45 days from the date of purchase order and this clause should be strictly adhere to failing which administrative action as deemed fit under rules will be taken against the defaulter. Otherwise Liquidation Damages will be imposed as per clause no. 4. Purchase order will be placed as required by consignee	Delivery& Installation (if applicable) : The successful bidder should strictly adhere to the following delivery schedule supply, installation & commissioning should be effected within 90 days from the date of opening of Letter of Credit else LD charges shall be imposed.
3.	Page.No.18 Point.No.1. C	Separate CE Certificate for Workflow Information System.	No changes in the existing specification.
4.	Page no.18 Point.No.8	The XDS registry & repository should be of the by same vendor who is providing WS PACS S/W. No freeware XDS registry/repository should be offered with the proposed upgrade solution.	No changes in the existing specification.
5.	Page no 18 Point.No.9	Non-Dicom data should be stored in native format only. Data format correction/conversion is not acceptable. This is in view of upgrading the existing solution to a Clinical Repository for Non-DICOM Data.	No changes in the existing specification.
6.	Page.No.20 Point.No.b	The vendor should have installed and managed a WS& PACS Site for more than 10 Years for at least two hospitals in India	The vendor should have installed and managed a WS& PACS Site for more than 7 Years for at least two hospitals in India
7.	Page no 20 Point.No.g	All the proposed IT solutions included in this tender should be from single OEM. Third party clinical OEM products used for 3D should not be quoted.	No changes in the existing specification.

8.	Page.No.31 Point.No.4	3D Post processing License for 150k Exams per year for and below application available on all Radiation Workstations	3D Post processing License for 250k Exams per year for and below application available on all Radiation Workstations
		-Pre Processing -Volume Rendering -Auto Bone Removal - Multi Modality Fusion	-Pre Processing -Volume Rendering -Auto Bone Removal - Multi Modality Fusion
9.	Page.No.31 Point.No.4	Exact CT angio data from CT Perfusion Exams 4 concurrent user licenses	Exact CT angio data from CT Perfusion Exams 3 concurrent user licenses
8.	Page.No.31 Point.No.4	Advanced 3D post processing Tools needed on Concurrent user basis which are embedded with PACS-WS Viewer	No changes in the existing specification.
9.	Page no.35 Point.No.8	Speech Magic Voice Recognition Software to be quoted separately- Nuance Speech Magic - 3 Concurrent User and 9 Voice Profiles with 3 Speech Mics with 3 Speech Mics	No changes in the existing specification
10.	Page no.36 Point.No.9.I	Specification for the Healthcare Digital Solution Infrastructure:- VRTX Blade Servers Chassis 4 Blades Server: 2x12 Core, 128 GB RAM per Blade, 25x1.2 TB 10K HDD, Dual Port FC on Chassis, CMC Management, 2xQuad Port PCI 1G NIC Card, 4 Power Supplies	No change in the existing specification.
11.	Page.No.37 Point.No.10	Imaging reporting workstation with 4MP FDA Approved Monitor and Single Clinical Monitor	Imaging displays should include online calibration and QA software.
12.	Page.No.37	Additional Points to be added	The price freeze for the hardware has to come with a warranty of 365 days or should except renegotiation if the fluctuation on dollar value goes above 5% in the time of frame.
13.	Page.No.30	Additional Points to be added on Software Licenses	Texad software with 03 concurrent users to be integrated with PACS

All other terms and condition will remain unchanged.

Dr.Karan Peepre Professor & Head, AIIMS,Raipur (C.G.)

> भंडार अधिकारी अखिल भारतीय आयुर्विज्ञान संस्थान, रायपुर (छ.ग.)