



अखिल भारतीय आयुर्विज्ञान संस्थान ,रायपुर (छत्तीसगढ़)  
**All India Institute of Medical Sciences, Raipur (Chhattisgarh)**  
Tatibandh, GE Road,  
Raipur-492 099 (CG)  
[www.aiimsraipur.edu.in](http://www.aiimsraipur.edu.in)

No. AIIMS/R/CS/NU.Med./18/236/OT

Raipur, Dtd:-16.04.2020

**Corrigendum**

**Tender ID.No.2020\_IMSRP\_548946\_1, Dated:29.02.2020.**

With reference to above Tender ID.No., the following amendment is issued:

S. No.	Item Name	Existing Tender Specification	To be read as
1	68Ge-68Ga Generator	2.f.Elution yield : Not less than 80%	Elution Yield: 80% with +/- 10% variation
2.	TLC Scanner for Quality Control	6.h.System should be capable of monitoring two detectors simultaneously.	System should have two detectors.
3	Warranty & CMC	Should have resident service engineer available in Raipur within 24hrs to solve the complaints	Onsite support as and when required for the first year.

All other terms and condition will remain unchanged.

Vetted by

Dr.Karan Peepre  
Professor and Head  
Nuclear Medicine Department  
AIIMS Raipur (C.G.)

Sr.Administrative Officer  
AIIMS Raipur (C.G.)