



No. AIIMS/R/CS/3649/OBGY/19/OT

Date - 16/07/2019

Corrigendum

Sub: Corrigendum in Notice Inviting Tender for “Supply of Electro Surgical Unit with Vessel Sealing” for Department of OBGY, AIIMS Raipur.

Tender Id: - 2019_IMSRP_479726_1 date 20/06/2019

In context to following corrigendum is issued against subject:

S.no.	Page No/ Clause no/ Point no.	Existing Specification	To be read as
1	Page no 19 Point no.1	It should be microprocessor controlled with atleast 90 user programmes with 300 Watt Electro Surgical Unit	It should be microprocessor controlled with wide range of user programmes with 300 Watt Electro Surgical Unit
2	Page no 19 Point no.3	It should have the facility of Dual application of hand and foot with 4 OUTPUTS i.e. 2 MONOPOLAR (cut and coagulation) AND 2 BIPOLAR (cut and coagulation) outputs	It should have the facility of Dual application of hand and foot with 4 OUTPUTS i.e. 1 MONOPOLAR (cut and coagulation) AND 1 BIPOLAR (cut and coagulation) outputs ,1 vessel sealing output and 1 neutral socket
3	Page no 19 Point no.5	There should be two sided neutral electrode plate for patient control system which continually monitors the neutral electrode for proper application	There should be Reusable two sided neutral electrode plate for patient control system which continually monitors the neutral electrode for proper application & It should from same OEM make quoted ESU Generator
4	Page no 19 Point no.8	It should have at least 25 total different types of currents for monopolar and bipolar application.	It should have wide range of different types of currents and monopolar and bipolar application.
5	Page no 19 Point no.11	It should have automatic HF power cut off by auto coagulation stop.	It should have automatic HF power cut off by auto coagulation stop/ Start/Both.
6	Page no 19 Point no.13	It should have reusable vessel sealing forceps required to seal up to 6-7 mm vessel for open surgery 23cm cvd (4nos.).	It should have reusable vessel sealing forceps required to seal up to 7 mm vessel for open surgery 21cm – 23cm cvd (4nos.) & It should be from same OEM Make of the Quoted ESU generator. The 7mm vessel sealing should be approved by a dignified certifying authority such as BIS / European CE / USFDA”.
7	Page no 19 Point no.14	Should offer reusable laparoscopy vessel Seal instrument 5mm. The instrument should have seal with inbuilt cutting facility (2 nos). The Laparoscopy Vessel Seal instrument should be of same parent company as the main unit.	Should offer reusable laparoscopy vessel Seal instrument 5mm. The instrument should have seal Capacity upto 7mm (2nos). The Laparoscopy Vessel Seal instrument should be of same OEM as the main unit.”
8	Page no 19 Point no 15	Should have reusable bipolar scissor for open surgery and laparoscopy surgery (4 each) and should be of same parent company as main unit.	Should have reusable bipolar scissor for open surgery and laparoscopy surgery (4 each) and should be of same parent company as main unit OEM.
9	Page no 19 Point no 16	The maximum power of different type outputs should be as:- · Monopolar Cutting Max - 300W- 400W, Monopolar Coagulation 120W or more	The maximum power of different type outputs should be as:- · Monopolar Cutting Max - 300W- 400W, Monopolar Coagulation 120W or more · Bipolar Cutting 150 W or more,

		10· Bipolar Cutting 150 W or more, Bipolar coagulation 120 W or more · Endo Mode Max – 100 W at 200 Ohms · Contact Coagulation – Max 250 W at 200 Ohms · Spray Coagulation – Max 120 W at 300 Ohms	Bipolar coagulation 120 W or more .Endo Mode Max – 100 W at 200 Ohms . Contact Coagulation – Max 250 W at 200 Ohms /Endocut effect Mode. . Spray Coagulation – Max 120 W at 300 Ohms.
11	Page no 19 Point no 18	The system should be supplied with all the standard reusable accessories such as <ul style="list-style-type: none"> • Reusable Pencil with finger switch – 5 • Dual pedal footswitch -1 • Electrode set of needle/Ball/pin point/Flat tip long blade - 10 Nos. each • Reusable Earth Pad – 5. • Bipolar Forceps -5 • Bipolar Cable – 5 • Bipolar Footswitch - 1 	The system should be supplied with all the standard reusable accessories and It should be from same OEM make of the Quoted ESU generator such as <ul style="list-style-type: none"> • Reusable Pencil with finger switch – 5 • Dual pedal footswitch -1 • Electrode set of needle/Ball/pin point/Flat tip long blade - 10 Nos. each • Reusable Earth Pad – 5. • Bipolar Forceps -5 • Bipolar Cable – 5 • Bipolar Footswitch - 1
12	Page no 20 Point no 21	Should have cart system for electro surgery unit with following accessories:- - One power unit - It should have five year warranty from the date of installation and should be quoted with CMC rate for 5 year after warranty period. - It should be provided with trolley 1 no.	Should have cart system for electro surgery unit with following accessories:- - One power unit - It should have five year warranty from the date of installation and should be quoted with CMC rate for 5 year after warranty period. It should be provided with trolley 1 no. and cart which may from same OEM make of the Quoted ESU generator.

Note:

- a) All other terms & condition will be remain unchanged.
b) Corrigendum must be uploaded along with technical bid as acceptance.

भंडार अधिकारी
अखिल भारतीय आयुर्विज्ञान संस्थान,
रायपुर (छ.ग.)