



AIIMS/R/CS/Neuro/ 289/18/OT /corrigendum/

दिनांक:- 17.02.2020

Corrigendum

Tender ID No.: 2018_IMSRP_535761_1

With Reference to above tender ID No., the following amendment is being issued for the tender “Supply of Radiofrequency ablation” for the Department of Neurosurgery at AIIMS Raipur:-

Annexure I

S.N.	Technical specifications Existing	Technical specifications to be read as
1	Have touch screen interface.	Should be touch screen/ feather touch button interface
3	Have thermal & pulsed RF treatment on 1, 2, 3 or 4 electrodes.	Multi RF cable for the same should be supplied
12	E-dose functionality	Multi Electrode customized parameter setting
15	Should have capability of intradiscal electrode.	Should have capability of long intradiscal electrode.
16	Should have compatible echogenic RF Cannulae	Should have compatible RF Cannulae
18	Also have spinal cannulae (Any Size) Qty. - 15 boxes	Need clarification on usage of spinal needles so that appropriate accessory can be quoted
20	Should have compatible disposable electrodes.	Should be supplied with disposable electrodes 50 nos from the same manufacturer
	Added:-	Should have option to upgrade the same system to a cooled radio frequency system as and when required in the future and price for the same should be quoted separately
		Have capability of continuous impedance measurement
		Have capability of continuous temperature measurement
		Have capability of voltage control, temperature control and width control in pulsed RF mode
		Have graphic display of temperature-gradient
		Have inbuilt capability to store up to 05 preset parameters
		Have remote control facility
		Have two RF channels with the capability of independent treatment of two lesions at the same time

Sr. Administrative Officer
AIIMS Raipur