



No. AIIMS/R/CS/1856/PHY/19/OT

Date – 17/02/2020

Corrigendum

Sub: Corrigendum in Notice Inviting Tender for “Supply of Body Composition Analyser” for the Department of Physiology, AIIMS Raipur.

Tender Id: 2019_IMSRP_510064_2.

In context to following corrigendum is issued against subject:

S.no.	Page/Clause/Point	Existing	To be read as
1.	Page 17 Clause – Annexure – I Technical Specification Condition	Output : Body Composition Intracellular Water, , Body Fat, Soft Lean Mass, Fat Free Mass, Weight, Skeletal Muscle Mass, Body Fat Mass, Percent Body Fat, BMI, Segmental Lean Analysis, Segmental Water Analysis, Total and segmental water ratio(ECW/TBW), BCM(Body Cell Mass), BMC(Bone Mineral Extracellular Water, Total Body Water, Protein Mineral Content), AC(Arm circumference), AMC(Arm Muscle Circumference), Waist Circumference, Visceral Fat Area, Basal Metabolic Rate(BMR), TBW/FFM, Body Water History(12times accumulated results), Impedance of Each Segments & Frequencies(Impedance, Reactance, Phase angle) Also Body Water result sheet should be obtained separately.	Output : Body Composition Intracellular Water, Extracellular Water, Total Body Water, Protein Mineral, Body Fat, Soft Lean Mass, Fat Free Mass, Weight, Skeletal Muscle Mass, Body Fat Mass, Percent Body Fat, BMI, Segmental Lean Analysis, Segmental Water Analysis, Total and segmental water ratio(ECW/TBW), BCM(Body Cell Mass), BMC(Bone Mineral Content), Visceral Fat Area, Basal Metabolic Rate(BMR), TBW/FFM, Body Water History (multiple times accumulated results) , Impedance of Each Segments & Frequencies(Impedance, Reactance, Phase angle). Desirable: AC (Arm circumference), AMC(Arm Muscle Circumference), Waist Circumference, Skeletal Muscle Mass Index. Also Body Water result sheet should be obtained separately.
2.	Page 17 Clause – Annexure – I Technical Specification Condition Point – Functional specifications.	Portability Indoor :- With cart (to be provided), Outdoor - with portable bag (to be provided) Data Back-Up : Possible to back up data through USB Storage Device and to restore the data to the In Body Power Output Display: DC 12V, 3.4A	Portability: Indoor - With cart (to be provided) / in built wheels , Outdoor - with portable bag (to be provided). Data Back-Up: Possible to back up data through USB Storage Device and to restore the data to the dedicated software. Power Output: DC 12V, 3.4A / Direct power input.
3.	Page 18 Clause – Annexure – I Technical Specification Condition Point - Other Specifications	Display Type: Minimum 800 x 480 Touch color LCD External Interface: RS-232C 1EA, USB Slave 1EA, USB Host 1EA Compatible Printer: Laser: PCL 3 or above and SPL or SPL6, Thermal Printer (Optional) Operation Environment: 10 ~ 40°, 30 ~ 80%RH Storage Environment: -20 ~ 60°, 10 ~ 95%RH, 50 ~ 106kPa (No condensation)	Display Type : Minimum 800 x 480 Touch color LCD / Latest Touch screen laptop / TFT LCD touch panel with at least 8 GB RAM and 1 TB HDD External Interface : RS-232C 1EA, USB Slave 1EA, USB Host 1EA / USB Type- A, Type-B & LAN port Compatible Printer: Compatible laser printer (preferably Laser : PCL 3 or above and SPL or SPL6), Thermal Printer(Optional) Operation Environment: 10 ~ 40° (±5), 30 ~ 80% RH Storage Environment : -20 ~ 60°(±10), 10 ~ 95% RH, 50 ~ 106kPa (No condensation)

Note:

- a) All other terms & condition will be remain unchanged.*
- b) Corrigendum must be uploaded along with technical bid as acceptance.*

**Sr. Administrative Officer
AIIMS Raipur**