



No. AIIMS/R/CS/Suture/19/RC/Corrigendum/

Date: 17-08-2019

Corrigendum

Tender reference no: AIIMS/R/CS/Suture/19/RC

Tender ID: 2019_IMSRP_488206_1

S. No.	Page No/ Clause no/ point no	Existing	To be read as
1	As mentioned in Critical date sheet	Bid Submission End Date : 19/08/2019 at 3:00 PM	Bid Submission End Date : 26/08/2019 at 3:00 PM
2	As mentioned in Critical dates sheet	Bid Opening Date : 21/08/2019 at 3:30 PM	Bid Opening Date : 28/08/2019 at 3:30 PM
3	Page no. 5, Point no. 27.	Earnest money of ₹10,74,000.00 by means of a Bank Demand Draft/ FD, a scanned copy to be enclosed. It is also clarified that the bids submitted without earnest money will be summarily rejected. The DD/FD may be prepared in the name of "All India Institute of Medical Sciences, Raipur (AIIMS RAIPUR)".	Earnest money of ₹10,74,000.00 by means of a Bank Demand Draft/FD/BG, a scanned copy to be enclosed. It is also clarified that the bids submitted without earnest money will be summarily rejected. The DD/FD/BG may be prepared in the name of "All India Institute of Medical Sciences, Raipur (AIIMS RAIPUR)".
4	Addendum as point no. 24 in Page no. 9	-	Sample Evaluation: Sample of all the suture items should be provided for evaluation to ensure quality after technical suitability when called for.
5	Addendum as point no. 25 in Page no. 9	-	Minimum shelf life of product should be for 4 years.
6	Addendum as point no. 26 in Page no. 9	-	Grade of Stainless Steel suture should be 316.
7	Addendum as point no. 'P' in Page no. 14	-	Performance Certificate: Considering the quality and continuity of the product, the bidder should also have manufactured at least 3 commercial batches of the quoted items every years in the last 5 consecutive years.
8	Addendum as point no. 'Q' in Page no. 14	-	It should provide with WHO GMP certificate for Indian Manufacturer.
9	Addendum as point no. 'R' in Page no. 14	-	It should prove with ISO 13485 certificate.
10	Addendum as point no. 'S' in Page no. 14	-	It should provide with European CE/BIS/USFDA Certificate.
11	Addendum as point no. 'T' in Page no. 14	-	Valid Drug License of the bidder & Manufacturer should be provided.

Specification to be read as (Annexure-I of tender and BOQ):-

Schedule 2

Sl.no.	Suture material	USP	Needle type	Size	Dimension	Anticipated annual consumption
9	Polypropylene	6-0	Double needle round body (75cm)	3/8 circle	13mm	60 box
10	Polypropylene	5-0	Double needle round body (90cm)	½ circle	17mm	60box
11	Polypropylene	4-0	Double needle round body (90cm)	½ circle	17mm	180box
Schedule 3						
12	Polypropylene	5-0	Taper point , double needle (75cm)	3/8 circle	13mm	120 box
13	Polypropylene	4-0	Taper point , double needle (75cm)	3/8 circle	26mm	120 box
14	Polypropylene	3-0	Taper point , double needle (75cm)	1/2 circle	22mm	120 box
15	Polypropylene	2-0	Taper point , double needle (90cm)	1/2 circle	26mm	120 box
Schedule 13						
70	Polypropylene	7-0	Taper point , double needle (75cm)	3/8 circle	9mm/9.3mm	120 box
71	Polypropylene	8-0	Taper point , double needle (60cm)	3/8 circle	8mm	60box

Specifications of other items will remain same.

Stores officer
AIIMS, Raipur