



क्रमांक: AIIMS/R/CS/Pedia S./18/257/OT/Corrigendum/

दिनांक: 18/09/2018

Tender ID No.: 2018_IMSRP_374370_1

With reference to above tender ID No., the following amendment is being issued:-

Corrigendum -01

Page/Clause/Point	Existing	To be read as
Page no. 18, point no. 06	System should have separate monopolar, bipolar & Vessel Sealing foot pedal.	System should have separate monopolar, bipolar & Vessel Sealing foot pedal/ monopolar and bipolar foot pedal which can be used for vessel sealing also.
Page no. 18, point no. 25	System should have both reusable open surgical instruments for Vessel Sealing purposes.	System should have both reusable open and laparoscopic surgical instruments for Vessel Sealing purpose and for dissection.
Page no. 18, point no. 27	System should be US FDA/European CE/BIS approve (Provided BIS standards are available)	System and all accessories should be US FDA/European CE/BIS approve (Provided BIS standards are available) and from same manufacturer.

All other terms and condition will remain unchanged.

(सुशील सोनबेर)
भंडार अधिकारी
अखिल भारतीय आयुर्विज्ञान संस्थान,
रायपुर (छ.ग.)