



क्रमांक AIIMS/R/CS/Pul-Med/17/4977/OT

रायपुर दिनांक :-20/11/2017

Corrigendum

Tender ID.No.2017_IMSRP_260155_1, Dated: 26.10.2017.

With reference to above Tender ID.No., the following amendment is issued:

Page/clause/point	Existing sentence	To be read as
NIT PAGE 18/2.1	System complete with all hardware and software is required.	ADD: Polysomnography System must have integrated ZRIP driver allows to setup the patient faster and easier by reducing the no. of connection during hook-up.
NIT PAGE 19/3.4.k	Data replay software operates on Microsoft windows XP, windows vista and windows 7 professional. Computer with 8GB RAM with 500GB hard drive with colour laser printer.	Data replay with software operates on latest Microsoft windows professional. Computer with 8GB RAM with 1024 GB hard drive with colour laser printer.
NIT PAGE 19/3.4.t	Titration device should have modes like: CPAP, AUTO CPAP, BIPAP, adaptive ventilation (same make)	Titration device should be as follows: Modes: CPAP, Auto CPAP, S, Auto Bi-level, S/T, T, PC, AutoSV, AVAPS-AE. Weight Approximately: less than 2.5 kg Altitude compensation: Automatic Pressure ranges CPAP : 4- 20 cm H2O Auto CPAP : 4 -20 cm H2O S Mode :EPAP: 4-25 cm H2O, IPAP: 4 – 30 cm H2O Auto Bi-level : EPAP: 4 -25 cm H2O,IPAP 4-25 cm H2O S/T, T, PC:EPAP:4- 25 cm H2O, IPAP:4 – 30 cm H2O AutoSV :EPAP: 4-30 cm H2O, Pressure Support: 0- 26 cm H2O AVAPS-AE: 4-30 cm H2O, Pressure Support 2 – 26 cm H2O It should have integrated Humidifier It should be USFDA approved.
NIT PAGE 20/9.9.2	One double bed one , cubud	<ul style="list-style-type: none"> One complete wooden double bed 6 feet width x 6 feet length with suitable double bed mattress (sleepwell/ kurlon) with side table One wooden double door cupboard with 70cm length x 35 cm width with multiple internal racks and one locker.

NIT PAGE 20/9.9.5	False flooring as well as POP with LED lighting	False roofing with all necessary POP work with adequate LED lighting (Philips)
ENTER CLAUSE 10	NONE	ADD 10: Additional Accessories: Actigraphy-1 LoFlo (Etco2) Starter Kit- 1 Standard set of sensor kit-2
ENTER CLAUSE 11	MENTIONED BELOW CLAUSE 9 IN NIT PAGE 20 : FIVE YEAR WARRANTY/ CMC WITH PROVISION OF CONSUMABLES SUPPLIED FREE OF COST.	TERMS AND CONDITIONS: <ul style="list-style-type: none"> • 5 years comprehensive warranty after installation. • CMC for 5 years after completion of 5 years warranty of the equipment, spare parts/ accessories used. • The company should give the certificate that the model quoted is the latest and not obsolete; and spares will be easily available for next 7 years.

All other terms and condition will remain unchanged.

Vetted by TSC Members.

(सुशील सोनबेर)
भंडार अधिकारी
अ.भा.आ.सं.- रायपुर (छ.ग.)