



No. AIIMS/R/CS/Gen.Sur./16/008/OT

Raipur, Dtd:-22/07/2017

Corrigendum

Tender ID.No.2017_IMSRP_213958_1, Dated:16.06.2017.

With reference to above Tender ID.No., the following amendment is issued:

S. No.	Item Name	Existing Tender Specification	To be read as
1	Diagnostic Gastrovideoscope	Should have Chrome endoscopy imaging (NBI/FICE/ I scan) and dual focus	Should have Chrome endoscopy imaging (NBI/FICE-BLI/ I scan-OE/m BLU/S technology) and preferably dual focus Capacity for detailed mucosal study
		Fully immersible in disinfectant solution (No need to attach water resistant cap) and one touch connectivity	Fully immersible in disinfectant solution.
		The scope should be the latest available in Indian market	The scope should be the latest high end model available in Indian market
		1.D.Depth of field: Normal 4/5- 100 mm or better	Depth of field: Near 2-3mm or better and Far 4-100 mm or better
2.	Accessories	Accessories	Accessories: Authorization certificate by the parent company of the endoscope should be attached for compatibility.
3	Colonovideoscope	Should have Chrome endoscopy imaging (NBI/FICE/ I scan) and dual focus detailed mucosal study	Should have Chrome endoscopy imaging (NBI/FICE-BLI/ I scan-OE/m BLU/S technology) for detailed mucosal study.
		The scope should be the latest available in Indian market	The scope should be the latest high end model available in Indian market.
		Field of view: In normal focus ≥ 140 degrees and in near focus ≥ 160 degrees	Field of view: In normal focus ≥ 140 degrees or better and in near focus ≥ 140 degrees or better.
		Depth of field: Normal 4/5- 100 mm and near 2-6 mm or better	Depth of field: Far-4-100 mm or better and near 2-3 mm or better.
4	Accessories	Accessories	Accessories: Authorization certificate by the parent company of the endoscope should be

			attached for compatibility.
5	Video Processor unit along with the light source	Should contain electronics to operate dual focus for clear visibility of near and far objects	Should contain electronics for clear visibility of near and far objects.
		Optical chrome endoscopy imaging such as NBI/FICE/I scan and HD plus videoscope	Optical chrome endoscopy imaging such as NBI/FICE-BLI/I scan-OE/mBLU/S technology and HD plus videoscope.
		Automatic IRIS control and automatic white balance	Automatic IRIS control and white balance.
		Should have inbuilt light source or separate light source with NBI/FICE/I scan imaging capacity/ HD plus video	Should have inbuilt light source or separate light source with NBI/FICE-BLI/I scan - OE/m BLU/S technology imaging capacity/ HD plus video.
		Compatible for waterproof one touch connector	As removed from the specification
6.	High definition LCD monitor	High definition LCD monitor	High definition LED monitor (From the same parent manufacturer of the scope) including the separate 32" LED monitor for teaching purpose
7.	Comprehensive Warranty	Comprehensive warranty is for 2 years	Comprehensive warranty is for 5 years
8.	Additional added points	1.Requirement of Almirah was not mentioned	An almirah either customized or made to order should be provided for housing the scopes should be provided.
		2.Demonstration of the quoted model of the endoscope at AIIMS Raipur was not mentioned in the specification	Demonstration of the quoted model of the endoscope at AIIMS Raipur is mandatory for technical evaluation and acceptability.

All other terms and condition will remain unchanged.

Vetted by

Dr.Debajyoti Mohanty
Associate Professor
General Surgery

Stores Officer
AIIMS Raipur (CG)



अखिल भारतीय आयुर्विज्ञान संस्थान ,रायपुर (छत्तीसगढ़)
All India Institute of Medical Sciences, Raipur (Chhattisgarh)
Tatibandh, GE Road,
Raipur-492 099 (CG)
www.aiimsraipur.edu.in

No. AIIMS/R/CS/Gen.Sur./16/008/OT

Raipur, Dtd:- 22/07/2017

Corrigendum

Tender ID.No.2017_IMSRP_213958_1, Dated:16.06.2017.

With reference to above Tender ID.No., the following amendment is issued:

Page/Clause/ Point	Existing Sentence	To be read as
As mentioned in Tender Document	Last Date of Bid Submission 22.07.2017.	Last Date of Bid Submission 10.08.2017 at 3:30PM
As mentioned in Tender Document	Bid Opening Date:27.07.2017	Bid Opening Date:09.08.2017 at 4:40PM.

All other terms and condition will remain unchanged.

**Stores Officer
AIIMS Raipur (CG)**