



अखिल भारतीय आयुर्विज्ञान संस्थान ,रायपुर) छत्तीसगढ़
(All India Institute of Medical Sciences, Raipur (Chhattisgarh)
Tatibandh, GE Road, Raipur-492 099 (CG)
www.aiimsraipur.edu.in

No. AIIMS/R/CS/21150/OPHTH/18/OT /

Date - 21/08/2020

Corrigendum

Sub: Corrigendum in Notice Inviting Tender for “Vitreous Retina Unit with Integrated Phacoemulsification for the department of Ophthalmology, AIIMS Raipur.

Tender Id: - 2019_IMSRP_497959_2 date 18/07/2020

In context to following corrigendum is issued against subject:

S.no.	Page/Clause/Point	Existing	To be read as/Addenda
1.	Page 19	Anterior segment - Submodes -- •Burst, pulsed, continuous •Hand piece with Longitudinal and torsional movement.	Anterior segment - Submodes -- • Burst, pulsed, continuous • Hand piece with Longitudinal / torsional movement.
2	Page 20	<u>Endoillumination</u> - Xenon light source preferably with mirrors for uniform co-illumination with 2 or more than 2 ports. - System should have ability to monitor the bulb life	<u>Endoillumination</u> - Xenon / LED light source preferably with mirrors for uniform co-illumination with 2 or more than 2 ports. - System should have ability to monitor the bulb life (Optional).

Note :-

*All other terms & condition will be remain unchanged.
Corrigendum must be uploaded along with technical bid as acceptance.*

**Store Officer
AIIMS Raipur**