



अखिल भारतीय आयुर्विज्ञान संस्थान ,रायपुर) छत्तीसगढ़
(All India Institute of Medical Sciences, Raipur (Chhattisgarh)
Tatibandh, GE Road, Raipur-492 099 (CG)
www.aiimsraipur.edu.in

No. AIIMS/R/CS/102/PHY/19/OT

Date – 23/07/2019

Corrigendum

Sub: Corrigendum in Notice Inviting Tender for “Supply of CPR Mannequin” for Department of Physiology, AIIMS Raipur.
Tender Id: - 2019_IMSRP_479336_1

In context to following corrigendum is issued against subject:

S.no.	Page/Clause /Point	Existing Specification	To be read as
1.	Page 16 Point No. 2. (a)	Initial state of manikin with pupillary dilation along with absence of carotid artery pulse	The change in size of the pupil and appearance of the carotid pulse with any wired / wireless remote or any device.
	Page 16 Point No. 2. (b)	Should be able to produce passive carotid artery pulse and the pulse rate in synchrony with compression operation.	Should be able to produce passive carotid artery pulse and the pulse rate in synchrony with compression operation.
3.	Page 16 Point No. 4.	To perform artificial breathing, with electronic monitoring of artificial breath rate and cardiac compression rate	To perform artificial breathing, with electronic monitoring of artificial breath rate and cardiac compression rate and depth
4.	Page 16 Point No. 5.	Should have at least three operational features: CPR training, assessment of simulation at least up to complete 5 cycles CPR operation according to the 2010 international CPR standard.	Should have at least three operational features: CPR training, assessment of simulation at least up to complete 5 cycles CPR operation according to the 2015 international CPR standard.
5.	Page 16 Point No. 11.	Desirable : Remote controller for teacher to control CPR operation, pupillary diameter, pulse and breathing activity, etc.	Desirable : Remote controller for teacher to control CPR operation, pupillary diameter, pulse and breathing activity, etc.

Note:

- All other terms & condition will be remain unchanged.
- Corrigendum must be uploaded along with technical bid as acceptance.

भंडार अधिकारी
अखिल भारतीय आयुर्विज्ञान संस्थान,
रायपुर (छ.ग.)