

## अखिल भारतीय आयुर्विज्ञान संस्थान ,रायपुर (छतीसगढ़)

## All India Institute of Medical Sciences, Raipur (Chhattisgarh)

Tatibandh, GE Road, Raipur-492 099 (CG) www.aiimsraipur.edu.in

Raipur, Dtd:-23.10.2018

No. AIIMS/R/CS/Micro/18/49/OT

## **Corrigendum**

## Tender ID.No.2018\_IMSRP\_388416\_1, Dated:26.09.2018.

With reference to above Tender ID.No., the following amendment is issued:

S.N o.	Page no in the NIT	<b>Existing Tender Specification</b>	Points raised by bidders	To be read as
1.	Point.No.	Guarantee/Warranty, Service, Maintena nce: The Bidders should submit their quote for subsequent 3 years on site CMC for Refrigerated Centrifuge but it should not be more than 7% per year of quoted unit price other wise offer mar summarily rejected. Failure to comply this condition will entail the rejection of the Bids. The price comparsion shall be made taking into account on basic price and post warranty CMC. The Rate Contracting Authority reserves the right to award CMC (with spare parts) shall be quoted for equipment. So the price of CMC should be quoted according to the cost of equipments. The amount of CMC would be released to the supplier on successful completion of the maintenance of that particular year duly certified by the user department.	1.M/s.KasliwalBrothers Raipur: The CMC charges after 3 years of warranty shall ber more than 7% of the basic value since there can be wear & tear.As such it is requested to delete the clause.  2.M/s.SuranaEnterprises Raipur:Firstly they have suggested to remove this 7% bounding as CMC prices would be given as a present currency status within that limit.  Secondly the CMC prices would be given as per present currency status which would be freezed for the 4 <sup>th</sup> ,5 <sup>th</sup> & 6 <sup>th</sup> year and as per the present currency fluctuations scenario there is more than 10% in a duration of one month.	Not accepted ,stand good.
2.	Page.No.1 9.	System should be UL listed CSA certified.	1.M/s.Kasliwal Brothers: They requested to please delete the clause –Sysyem should be UL listed & CSA certified. Since already you have asked for many certification.	No changes in the existing specification.
3.	Page no 20.	Additional points to be added		1.The buckets and rotor sealing lids must be certified for biocontainment by a 3 <sup>rd</sup> party lab of worldwide recognition. 2.System should be CE certified & IVD approved.

All other terms and condition will remain unchanged.

Dr.Ujjwala Nitin Gaikwad Additional Professor AIIMS,Raipur (C.G.)

> (सुशील सोनबेर) भंडार अधिकारी अखिल भारतीय आयुर्विज्ञान संस्थान, रायपुर (छ.ग.)