



AIIMS/R/CS/Neuro/18/307/OT /corrigendum

दिनांक:- 25.06.2019

Corrigendum

Tender ID No.: 2019_IMSRP_473218_1

With Reference to above tender ID No., the following amendment is being issued for the tender “Neuro – Ventilator for Adult Paediatric and Neonatal Patients” for Neurosurgery Department at AIIMS Raipur:-

Annexure I

	Existing		To be read as
S. no.	Specification Sr. No.		
1	1	Should be microprocessor controlled ventilator with 12 inch color TFT touch screen integrated graphics and easy to use rotary knob operation providing support to adult, pediatric and infant/neonate patient range.	Should be microprocessor controlled ventilator with tiltable 12 inch or more color TFT touch screen integrated graphics and easy to use rotary knob operation providing support to adult, pediatric and infant/neonate patient range.
2	12	Peak flow settings from 0 to 240 L/min or better	Peak flow settings from 0 to 200 L/min or better
3	13	Trigger sensitivity – flow 1 to 20 L/min	Trigger sensitivity – flow 0.2 to 15 L/min
4	17	Inspiratory time (T1) 1 to 12 sec	Inspiratory time (T1) 1 to 10 sec
5	28	Should have visual representation of ventilator dependency grouped into oxygenation CO2 elimination and patient activity	Should have visual / numeric representation of ventilator dependency grouped into oxygenation CO2 elimination and patient activity
6	36	Nebulizer kit – 20 NO	Disposable Nebulizer kit – 20 NO

भंडार अधिकारी
अखिल भारतीय आयुर्विज्ञान संस्थान,
रायपुर (छ.ग.)