



अखिल भारतीय आयुर्विज्ञान संस्थान, रायपुर (छत्तीसगढ़)

All India Institute of Medical Sciences, Raipur (Chhattisgarh)

Tatibandh, GE Road,

Raipur-492 099 (CG)

www.aiimsraipur.edu.in

क्रमांक: AIIMS/R/CS/H. Admin/18/390/OT/A/Corrigendum/

दिनांक: 26/04/2019

Tender ID No.: 2018_IMSRP_421834_2

With reference to above tender ID No., the following amendment is being issued:-

Corrigendum

Page/Clause/Point	Existing	To be read as
Page no. 25 under Medical Equipment item no. 1	Automatic Collapsible Stretcher cum Trolley – it Should be EN 1789 certified	Automatic Collapsible stretcher cum trolley should be EN 1789 and EN 1865 certified.
Page no. 25 under Medical Equipment item no. 2	Scoop stretcher should be EN1865 certificate	Scoop stretcher should be EN 1865 certified as per AIS 125
Page no. 25 under Medical Equipments item no. 3	Spine Board with straps with head immobilizer – Fixtures device and compliance with EN 1789	Spine Board with Straps with head immobilizer – Fixture device should be compliance with EN 1789 and EN 1865

All other terms and condition will remain unchanged.

भंडार अधिकारी
अखिल भारतीय आयुर्विज्ञान संस्थान,
रायपुर (छ.ग.)