



अखिल भारतीय आयुर्विज्ञान संस्थान, रायपुर (छ.ग.)
All India Institute of Medical Sciences, Raipur (Chhattisgarh)
Tatibandh, GE Road, Raipur-492 099 (CG)

No. AIIMS/R/CS/Pul_Med/18/012/OT
Tender ID- 2019_IMSRP_475985_1

Raipur, Dtd:- 26/06/2019

Tender Name – Supply of Impulse Oscillometry with Essential accessories for Dept of Pulmonary Medicine at AIIMS Raipur.

Corrigendum

Group/ Point no of Technical Specification on Annexure - I	Existing Sentence	To be read as
Point - 02	The Forced Oscillation should be based on Signal Optimized Pseudorandom Noise (PRN) between 4Hz and 48Hz	The Forced Oscillation should be based on Signal Optimized Pseudorandom Noise (PRN) between 4Hz and 48Hz. It should have single frequency mode and multiple frequency mode.

Note :-

1. The above changes have no financial implication.
2. Technical specification remains the same as earlier except for above corrigendum.
3. All other terms and condition will remain unchanged.

**Stores Officer
AIIMS Raipur (CG)**