

अखिल भारतीय आयुर्विज्ञान संस्थान, रायपुर (छत्तीसगढ़) All India Institute of Medical Sciences, Raipur (Chhattisgarh)

Tatibandh, GE Road, Raipur-492 099 (CG) www.aiimsraipur.edu.in

No. AIIMS/R/HS/MOT/21326/2020/RC

Dated: - 27-11-2020

Technical Corrigendum Notice

Tender ID. No. 2020_IMSRP_595240_1 **Dated**: 05/11/2020

With reference to above Tender ID. No., the following amendment is issued:

S. No.	Page No. Details	Existing Tender Specification	To be read as
1	Page No.12 [c]	c) In case of Distributor/Dealer/Supplier must be upload Manufacture/OEM's annual turnover & balance sheet of last three year duly certified by CA as mentioned in tender document should be uploaded.	c) In case of OEM/Manufacturer/Distributor/ Dealer/ Supplier must be upload Bidders annual turnover & balance sheet of last three year duly certified by CA as mentioned in tender document should be uploaded.
2	Page No. 4 [12]	12. Manufacture/OEM should be registered and should have average annual turnover at least ₹ 36,45,000.00/- in the last three financial years.	12. Manufacture/OEM/Bidders should be registered and should have average annual turnover at least ₹ 36,45,000.00/- in the last three financial years.

In addition to above amendments, the following specifications of items that are needed to be revised which are as under:

S. No.	Page No. Details	Title under which amendments made	Revised additional specifications
1	Page No.13 [Annexure -1]	Technical Specification Of Pouch Roll For Plasma Sterilizer For Major Operation Theatre	 Seal strength test, microbial barrier properties, burst test and peel open test should demonstrated acceptable material compatibility characteristics of the Pouches/Rolls after processing in the Plasma Sterilizer; also sterilant penetration and package integrity. The Pouch Roll should maintain sterility for 12 months post processing in plasma sterilizer.

All other terms and condition will remain unchanged.

Stores Officer (H) AIIMS Raipur (C.G)