



No. AIIMS/R/CS/300/BPS/19/OT

Date - 28/08/2019

## Corrigendum

**Sub: Corrigendum in Notice Inviting Tender for “Supply of High End Defibrillator” for Department of Burns & Plastic Surgery, AIIMS Raipur.**

**Tender Id: - 2019\_IMSRP\_492631\_1 date 06/08/2019**

S. No	Page /Point No.	Existing	To be read as/Addenda
1	Page no. 19/ Point no. 3/ Clause 3.7	Adult and paediatric should be available	Adult and paediatric & <b>Neonatal</b> should be available
2	Page no. 19/ Point no. 4/ Clause 4.2	Paddles Adult & Pediatric -01 each	Paddles Adult & Pediatric & <b>Neonatal</b> - 01 each
3	Page no. 20/ Point no. 7/ Clause 7.2	Electrical safety conforms to acceptable standards (provide certificate)	(To Be Added) <ul style="list-style-type: none"><li>• Must meet IP44 Level for water resistance and protection against harmful ingress of dust.</li><li>• Should conform to latest safety standards, such as IEC-60601-2-4, IEC-60601-1-2, ISO 14971:2007, EN 1789:2007 etc.</li></ul>
4	Page no. 20/ Point no. 7/ Clause 7.5	To Be Added	7.5 The defibrillator should be up grad able for SPO2, NIBP, & ET, CO2.
5	Page no. 20/ Point no. 8/ Clause 8.7,8.8.,9,8.10,8.11	To Be Added	8.7. Must submit user list and performance report within last 5 years from major hospitals/government hospitals 8.8 Service manual in English 8.9 Certificate of calibration and inspection from factory. 8.10 Log book with instruction for daily, weekly monthly and quarterly maintenance check list. The job description of the hospital technician and company service engineer should be clearly spelt out. 8.11 List of Equipment available for providing calibration and routine maintenance support as per manufacturer documentation in service / technical manual.
6	Page no. 20/ Point no. 2	Temporary External Dual Chamber Pacemaker System	To be Deleted From point no 1 to 15

In context to following corrigendum is issued against subject:

**Note:**

- a) *All other terms & condition will be remain unchanged.*
- b) *Corrigendum must be uploaded along with technical bid as acceptance.*

भंडार अधिकारी  
अखिल भारतीय आयुर्विज्ञान संस्थान,  
रायपुर (छ.ग.)