



Date: 28-09-2018

**CORRIGENDUM**

Tender No. AIIMS/R/CS/CTVS/18/179/OT (Tender ID: 2018\_IMSRP\_375496\_1)

**Tender for “Supply & Installation of 16 High End Multipara Modular Monitor System with Central Monitoring Station” for Department of Cardiovascular & Thoracic Surgery and Department of Cardiology at AIIMS-Raipur.**

Sr. No.	NIT Page No./ Clause No./ Para No.	Existing sentence	To be read as/Addenda
	Page. No. 15, Item Sr. No.01, Point No. 1-1	Multipara monitor 12” TFT color display with touch screen facility with resolution of 800 X 600 dots. It should have HL7 gateway to connect the HIS & CIS.	Multipara monitor having screen size of atleast 18” TFT color display and above with touch screen facility with resolution of 1920 X 1080 dots. All the monitors should come with integrated transport module to avoid data loss. It should have ECG, SPO2, NIBP, Dual Temp., Respiration, Dual IBP & ETCO2 so that there is no loss in data in case the patient needs to be shifted from one bed to another. Monitor should have HL7 connectivity to connect to the HIS & CIS.
	Page. No. 15, Item Sr. No.01, Point No. 1-5	Monitor must be capable of simultaneously monitoring the following parameters which should be present as standard in all monitors: ECG, NIBP, SpO2, respiration, dual temperature, HR, Dual IBP, non-invasive cardiac output, invasive cardiac output through thermodilution.	Monitor must be capable of simultaneously monitoring the following parameters which should be present as standard in all monitors: ECG, NIBP, SpO2, respiration, dual temperature, HR, Dual IBP. a. Five of the monitors out of 16 should have Sepsis protocol along with either non-invasive cardiac output or invasive Cardiac output. In case a vendor is providing invasive cardiac Output, he has to provide all the necessary accessories to perform 750 cases. (NOTE: In case of non-invasive cardiac Output, no additional accessories are required for monitoring. Hence, those who are providing invasive cardiac output will have to supply accessories for 750 cases free of cost. Delivery should be 150 set of accessories every year for five years)
	Page. No. 15, Item Sr. No.01, Point No. 1-14	Monitor should have facility to monitor 12 leads ECG through 10 lead ECG cable	Monitor should have facility to monitor 12 leads ECG through 10 lead ECG cable along with 12 lead ECG analysis & Report.
	Page. No. 15, Item Sr. No.01, Point No. 1-15	All automatic arrhythmia detection & ST analysis should be available.	All 23 automatic arrhythmia detection & ST analysis should be available.
	Page. No. 16, Item Sr. No.01, Point No. 1-36	Include laser printer and dual channel strip recorder.	Include 2 laser printer and 2 dual channel strip recorder
	Page. No. 16, Item Sr. No.01, Point No. 1-38	To provide suitable facility for sending and receiving DICOM compatible radiological image like ultrasound X-Ray etc to and from the monitoring network to and from hospital information system radiology information system etc for integration of various information	<b>DELETED POINT – No. 38</b>
	Page. No. 16, Item Sr. No.01, Point No. 1-42	Following accessories need to be provided per 16 units of monitor 1. 16 unit ECG cable 2. 16 units of SPO2 sensors 3. 16 units of NIBP cuff 4. 16 units of NIBP tubing 5. 32 units of IBP cable 6. 16 units of skin temperature probe 7. 16 units of rectal temperature probe 8. 5 units of ETCO2 sensors	Following accessories need to be provided per 16 units of monitor 1. 16 unit ECG cable 2. 16 units of reusable adult finger clip & 5 units neonatal & 5 units child SPO2 sensors 3. 16 units of adult NIBP cuff & 5 units of child NIBP cuff & 5 units of infant NIBP cuff 4. 16 units of NIBP tubing 5. 32 units of IBP cable 6. 16 units of skin temperature probe

			7. 16units of rectal temperature probe 8. 5 units of ETCO2 sensors
	Page. No. 16, Item Sr. No.02, Point No. 2-12	The system should have facility to view the data remotely on iPad or iPhone. It should be supplied with at least one i-pad. Please quote the option separately.	The system should have Web Browsing, should provide monitoring information of multiple patients on an i-Pad or on android based tablets with minimum 9" screen size 128 GB of internal memory (Wi-fi & cellular). Full range of patient information should be able to access. Should be able to review the patient's condition and trends in near real-time from a conference room, physician's office, home or even outside the country and discuss treatment with the caregivers who are at the bedside. Should have complete review capability giving patient data that includes waveforms, 12-lead ECG analysis, full disclosure, arrhythmia and ST recall, trends, Bedside Monitor's Waveform display Window, Graphical Window, Hemodynamic window and other information, just like on the bedside monitor. Confidentiality and controlled access to patient data are important in a network. The administrator should be able to create multiple user accounts and control who can review which patients.
	Page. No. 16, Item Sr. No.02, Point No. 2-13	CNS should be BIS/ European CE or FDA approved.	CNS should be BIS/European CE & USFDA approved.
	Page No. 02 Clause No. 11	Tenderer must provide evidence of having supplied government hospital/ reputed private hospital organizations in India similar nature of items of at least ₹98,34,000.00 for Supply and Installation of 16 High End Multipara Modular Monitor System with Central Monitoring Station in the last three years and the copy of the same should be uploaded.	Tenderer must provide evidence of having supplied government hospital/ reputed private hospital organizations in India similar nature of items of at least ₹1,96,68,000.00 for Supply and Installation of 16 High End Multipara Modular Monitor System with Central Monitoring Station in the last three years and the copy of the same should be uploaded.
	Page No. 02 Clause No. 12	The firm should be registered and should have the average annual turnover at least ₹1,96,68,000.00 of the bidder in the last three financial years. Copies of authenticated balance sheet for the past three financial years should be uploaded.	The firm should be registered and should have the average annual turnover at least ₹3,93,35,000.00 of the bidder in the last three financial years. Copies of authenticated balance sheet for the past three financial years should be uploaded.
	Page No. 02 Clause No. 13	Total estimated value of tender: ₹1,96,67,200.00.	Total estimated value of tender: ₹3,93,34,400.00.
	Page No. 02 Clause No. 28	<b>Earnest Money:</b> Earnest money of ₹5,91,000.00 by means of a Bank Demand Draft/FD, a scanned copy to be enclosed.	<b>Earnest Money:</b> Earnest money of ₹11,81,000.00 by means of a Bank Demand Draft/FD, a scanned copy to be enclosed.
	Page No. 1	Bid Document Download/ Sale End Date : 01-10-2018, 03.00PM	Bid Document Download/ Sale End Date : 08-10-2018, 03.00PM
		Bid Submission End Date : 01-10-2018, 03.00PM	Bid Submission End Date : 08-10-2018, 03.00PM
		Bid Opening Date : 03-10-2018, 03.30PM	Bid Opening Date : 10-10-2018, 03.30PM

**Note:**

1. All other terms & Conditions will remain unchanged.
2. Corrigendum must be uploaded along with technical bid as acceptance.

Stores Officer,  
Central Store,  
AIIMS, Raipur