



अखिल भारतीय आयुर्विज्ञान संस्थान, रायपुर (छ.ग.)  
All India Institute of Medical Sciences, Raipur (Chhattisgarh)  
Tatibandh, GE Road, Raipur-492 099 (CG)

No. AIIMS/R/CS/PuI\_Med/18/009/OT

Raipur, Dtd:- 29/09/2018

**Corrigendum**

Group/ Point no of Technical Specification on Annexure - I	Existing Sentence	To be read as
1.1	Hi support – automatically optimize the B- mode and Doppler image.	Should automatically optimize the B-mode and Doppler image with Real time tissue latest elastography.
1.6. IX	Computer –Intel i7 3.2 GHz processor or better (3 <sup>rd</sup> generation), at least 2 Terabyte hard disk , at least eight GB DDR3 RAM , DVD writer at least 19” TFT monitor ,at least four USB 3.0 ports, original Window 7 , UPS with power backup of at least 20 minutes.	Computer – Intel i7 3.2 GHz processor or better (latest generation) , at least 2 Terabyte SSD hard disk , at least eight GB DDR3 RAM, DVD writer, at least 24 “ TFT monitor resolution of at least 1280 x 1280, at least four USB 3.0 ports, original Window 10 , UPS with power backup of at least 20 minutes.
1.6. XI	NONE	ADDITION OF POINT XI – Should provide high quality customized case for storage and easy transport of bronchoscope for each of the scope. With 2 spare cases.

Note :-

1. Technical specification remains the same as earlier except for above corrigendum.
2. All other terms and condition will remain unchanged.

**Stores Officer  
AIIMS Raipur (CG)**