



No. AIIMS/R/CS/GM-GS/293-50/19/OT/B

Date:30/05/2020

Corrigendum

Subject:- Supply and Installation of High End ICU Ventilator for Department of General Surgery & General Medicine at AIIMS Raipur.

Tender ID No. 2019_IMSRP_451585_3

NIT Page No	Exiting	To be read as
Page 16 of 21/ Sr. 3	Should be based on reliable flow measuring technology with proximal/distal flow sensor which ensures the most precise flow and pressure measurements for better patient assessment .The sensor should also be covered under warrenty of 5 years/ or the vendor should provide flow	Should be based on reliable flow measuring technology with proximal/distal flow sensor which ensures the most precise flow and pressure measurements for better patient assessment .The sensor should also be covered under warrenty of 5 years / or the vendor should provide flow sensor so that machine should function uninturuptted
CONTROLS		
Page 16 of 21/ Sr. 18	Tidal volume should be minimum 20 ml to 2000 ml in Volume Control Mode or better	Tidal volume should be minimum 50 ml to 2000 ml in Volume Control Mode or better
Ventilator Should be supplied with the following accessories:		
Page 17 of 21/ Sr. 32	d.) 2 -reusable autocleavable flow sensors per ventilator with life time warranty	To be deleted

Stores Officer
AIIMS Raipur (CG)