



No. AIIMS/R/CS/MS/Defibrillator/18/711/OT

Dtd: 31/07/2019

Corrigendum

Subject:- Supply and Installation of Biphasic Defibrillator with ECG Machine for AIIMS Raipur.

Page	/Clause/ Point No.	Existing	To be read as
Page No. 16 of 20	2	The Machine must be upgradeable to vital sign parameter such as mainstream EtCO ₂ , SpO ₂ and NIBP on installation site	The machine must be upgradeable to vital sign parameter such as Mainstream / Side Steam/ Micro Stream EtCO ₂ , SpO ₂ and NIBP on installation site
Page No. 16 of 20	6	Should have display of atleast 4 waveforms of various monitoring parameters and their numerical data.	The Unit should have 3 Waveform of various monitoring parameter & their numerical data.

Store Officer
AIIMS Raipur (CG)