



Date: 31-08-2017

CORRIGENDUM

Tender No. AIIMS/R/CS/Hospital/17/705/OT
Tender for “Supply & Installation of ICU / ICCU Bed”
for Hospital at AIIMS-Raipur.

S. NO.	Page No./ Clause No./ Para No.	Existing sentence	To be read as/Addenda
1	Page No.15 Line No. 3 of Comfort	User friendly separate patient and attendant control panel.	It should have integrated controls on both inside and outside of railing for patient and attendant control along with one additional handset control.
2	Page No.16 Line No. 1 of Mattress	High density mattress (40kg/m ³), X ray translucent, latex, Height 14cm or more, cover material polyurethane coated polyester, Fire retardant mattress system (BS EN 597-1:2015 & BS EN 597-2:2015)	High density mattress (40kg/m ³), X ray translucent, latex free, Height 14cm or more, cover material polyurethane coated polyester, Fire retardant mattress system (BS EN 597-1:2015 & BS EN 597-2:2015). Cover should be water proof and stand chemical disinfection.
3	Page No. 1	Bid Document Download/ Sale End Date : 31-08-207, 03.00PM	Bid Document Download/ Sale End Date : 05-09-207, 03.00PM
4		Bid Submission End Date : 31-08-207, 03.00PM	Bid Submission End Date : 05-09-207, 03.00PM
5		Bid Opening Date : 01-09-207, 03.30PM	Bid Opening Date : 06-09-207, 03.30PM

Note:

1. All other terms & Conditions will remain unchanged.
2. Corrigendum must be uploaded along with technical bid as acceptance.

Stores Officer,
Central Stores,
AIIMS Raipur