



AIIMS/R/CS/ENT/19/879/RC /Corrigendum

Date:- 31.12.2019

**Corrigendum**

Tender ID No.: 2019\_IMSRP\_518704\_1

With Reference to above tender ID No., the following amendment is being issued for the tender  
 “**Rate Contract for supply of Implantable Hearing Device**” for the Department of ENT & HNS at  
 AIIMS Raipur:-

**Annexure I**

Page No/ Clause no/ point no	Existing sentence	To be read as
Point no.1	<b>COCHLEAR IMPLANT SYSTEM</b>	<b>COCHLEAR IMPLANT SYSTEM-BASIC MODEL</b>
Point no. 3 A	Options of Behind the Ear (BTE) Speech Processor with the ability to store Multiple Programmes	Options of Behind the Ear (BTE) Speech Processor OR Single unit Processor with the ability to store Multiple Programmes
Point no 16 .B	B. External Components, including Speech Processor & battery packs, cables & coils: 3years from the date of switch on.	External component: 3 year warranty for speech processor, coil, battery frames, remote control, electronic dehumidifier- 2 year, rechargeable battery and charger- 1 year, consumable excluded from warranty.
Point no. 16. C	Consumable such as batteries chargers, rechargeable batteries - 2 years from the date of switch on.	Nil
Point no. 18	At least one year of speech and rehabilitative therapy, which would include mapping etc?	At least one year of speech and rehabilitative therapy training support to the clinicians , which would include mapping etc.
nil	nil	Point. 20 Accessories should include 4-coil cables. 10 packets implant plus batteries (non-rechargeable)
nil	nil	Point. 21 For single unit the accessories should have 2-coils cable ( 28 cm) and 2 short cable ( 12 cm)
nil	nil	Point 22 Huggy ( 2 in number)
<b>ADVANCED COCHLEAR IMPLANT SYSTEM</b>		
1A	Nil	“The implant and speech processor quoted in this category should be advanced and different from the ones quoted in the basic category.”
Point 16 B	External Components, including Speech Processor & battery packs, cables & coils: 3years from the date of switch on.	External Components: 3 year warranty for Speech Processor & battery frame, Electronic dehumidifier-n 2 year; rechargeable battery and charger-1 year. Consumable excluded from warranty.



Point 18	At least one year of speech and rehabilitative therapy, which would include mapping etc.	At least one year of speech and rehabilitative therapy, training support to the clinicians which would include mapping etc.
<b>1. BONE CONDUCTION IMPLANT SYSTEMS</b>		
Point no 8	Gain: 45dB	nil

Sr. Administrative Officer  
AIIMS Raipur