



Bid Number/बोली क्रमांक (बिड संख्या):
GEM/2023/B/3339193
Dated/दिनांक : 18-05-2023

Bid Document/ बिड दस्तावेज़

Bid Details/बिड विवरण	
Bid End Date/Time/बिड बंद होने की तारीख/समय	02-06-2023 11:00:00
Bid Opening Date/Time/बिड खुलने की तारीख/समय	02-06-2023 11:30:00
Bid Offer Validity (From End Date)/बिड पेशकश वैधता (बंद होने की तारीख से)	170 (Days)
Ministry/State Name/मंत्रालय/राज्य का नाम	Ministry Of Health And Family Welfare
Department Name/विभाग का नाम	Department Of Health And Family Welfare
Organisation Name/संगठन का नाम	All India Institute Of Medical Sciences (aiims)
Office Name/कार्यालय का नाम	Raipur
Total Quantity/कुल मात्रा	16
Item Category/मद केटेगरी	High End Multipara Monitor (Q2)
Minimum Average Annual Turnover of the bidder (For 3 Years)/बिडर का न्यूनतम औसत वार्षिक टर्नओवर (3 वर्षों का)	60 Lakh (s)
OEM Average Turnover (Last 3 Years)/मूल उपकरण निर्माता का औसत टर्नओवर (गत 3 वर्षों का)	400 Lakh (s)
Years of Past Experience Required for same/similar service/उन्हीं/समान सेवाओं के लिए अपेक्षित विगत अनुभव के वर्ष	3 Year (s)
MSE Exemption for Years Of Experience/अनुभव के वर्षों से एमएसई छूट/ and Turnover/टर्नओवर के लिए एमएसई को छूट प्राप्त है	Yes
Startup Exemption for Years Of Experience/अनुभव के वर्षों से स्टार्टअप छूट/ and Turnover/ टर्नओवर के लिए स्टार्टअप को छूट प्राप्त है	Yes
Document required from seller/विक्रेता से मांगे गए दस्तावेज़	Experience Criteria,Past Performance,Bidder Turnover,Certificate (Requested in ATC),OEM Authorization Certificate,OEM Annual Turnover,Additional Doc 1 (Requested in ATC) *In case any bidder is seeking exemption from Experience / Turnover Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer

Bid Details/बिड विवरण	
Past Performance/विगत प्रदर्शन	50 %
Bid to RA enabled/बिड से रिवर्स नीलामी सक्रिय किया	Yes
RA Qualification Rule	50% Lowest Priced Technically Qualified Bidders
Comprehensive Maintenance Charges Required	Yes
Type of Bid/बिड का प्रकार	Two Packet Bid
Time allowed for Technical Clarifications during technical evaluation/तकनीकी मूल्यांकन के दौरान तकनीकी स्पष्टीकरण हेतु अनुमत समय	2 Days
Inspection Required (By Empanelled Inspection Authority / Agencies pre-registered with GeM)	No
Evaluation Method/मूल्यांकन पद्धति	Total value wise evaluation
Financial Document Required/वित्तीय दस्तावेज की आवश्यकता है।	Yes

EMD Detail/ईएमडी विवरण

Advisory Bank/एडवाइजरी बैंक	Bank of India
EMD Amount/ईएमडी राशि	373000

ePBG Detail/ईपीबीजी विवरण

Advisory Bank/एडवाइजरी बैंक	Bank of India
ePBG Percentage(%) / ईपीबीजी प्रतिशत (%)	5.00
Duration of ePBG required (Months) / ईपीबीजी की अपेक्षित अवधि (महीने).	62

(a). EMD EXEMPTION: The bidder seeking EMD exemption, must submit the valid supporting document for the relevant category as per GeM GTC with the bid. Under MSE category, only manufacturers for goods and Service Providers for Services are eligible for exemption from EMD. Traders are excluded from the purview of this Policy./जेम की शर्तों के अनुसार ईएमडी छूट के इच्छुक बिडर को संबंधित केटेगरी के लिए बिड के साथ वैध समर्थित दस्तावेज प्रस्तुत करने हैं। एमएसई केटेगरी के अंतर्गत केवल वस्तुओं के लिए विनिर्माता तथा सेवाओं के लिए सेवा प्रदाता ईएमडी से छूट के पात्र हैं। व्यापारियों को इस नीति के दायरे से बाहर रखा गया है।

(b). EMD & Performance security should be in favour of Beneficiary, wherever it is applicable./ईएमडी और संपादन जमानत राशि, जहां यह लागू होती है, लाभार्थी के पक्ष में होनी चाहिए।

Beneficiary/लाभार्थी :

AIIMS Raipur

Raipur, Department of Health and Family Welfare, All India Institute of Medical Sciences (AIIMS), Ministry of Health and Family Welfare

(Director)

Splitting/विभाजन

Bid splitting not applied/बोली विभाजन लागू नहीं किया गया.

MII Purchase Preference/एमआईआई खरीद वरीयता

MII Purchase Preference/एमआईआई खरीद वरीयता	Yes
--	-----

MSE Purchase Preference/एमएसई खरीद वरीयता

MSE Purchase Preference/एमएसई खरीद वरीयता	Yes
---	-----

1. If the bidder is a Micro or Small Enterprise as per latest definitions under MSME rules, the bidder shall be exempted from the requirement of "Bidder Turnover" criteria and "Experience Criteria" subject to meeting of quality and technical specifications. If the bidder is OEM of the offered products, it would be exempted from the "OEM Average Turnover" criteria also subject to meeting of quality and technical specifications. In case any bidder is seeking exemption from Turnover / Experience Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer.
2. If the bidder is a Startup, the bidder shall be exempted from the requirement of "Bidder Turnover" criteria and "Experience Criteria" subject to their meeting of quality and technical specifications. If the bidder is OEM of the offered products, it would be exempted from the "OEM Average Turnover" criteria also subject to meeting of quality and technical specifications. In case any bidder is seeking exemption from Turnover / Experience Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer.
3. The minimum average annual financial turnover of the bidder during the last three years, ending on 31st March of the previous financial year, should be as indicated above in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the bidder is less than 3-year-old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria.
4. Experience Criteria: In respect of the filter applied for experience criteria, the Bidder or its OEM {themselves or through reseller(s)} should have regularly, manufactured and supplied same or similar Category Products to any Central / State Govt Organization / PSU / Public Listed Company for number of Financial years as indicated above in the bid document before the bid opening date. Copies of relevant contracts to be submitted along with bid in support of having supplied some quantity during each of the Financial year. In case of bunch bids, the category of primary product having highest value should meet this criterion.
5. OEM Turn Over Criteria: The minimum average annual financial turnover of the OEM of the offered product during the last three years, ending on 31st March of the previous financial year, should be as indicated in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the OEM is less than 3 year old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria.
6. Preference to Make In India products (For bids < 200 Crore):Preference shall be given to Class 1 local supplier as defined in public procurement (Preference to Make in India), Order 2017 as amended from time to time and its subsequent Orders/Notifications issued by concerned Nodal Ministry for specific Goods/Products. The minimum local content to qualify as a Class 1 local supplier is denoted in the bid document. If the bidder wants to avail the Purchase preference, the bidder must upload a certificate from the OEM regarding the percentage of the local content and the details of locations at which the local value addition is made along with their bid, failing which no purchase preference shall be granted. In case the bid value is more than Rs 10 Crore, the declaration relating to percentage of local content shall be certified by the statutory auditor or cost auditor, if the OEM is a company and by a practicing cost accountant or a chartered accountant for OEMs other than companies as per the Public Procurement (preference to Make-in -India) order 2017 dated 04.06.2020. Only Class-I and Class-II Local suppliers as per MII order dated 4.6.2020 will be eligible to bid. Non - Local suppliers as per MII order dated 04.06.2020 are

not eligible to participate. However, eligible micro and small enterprises will be allowed to participate .In case Buyer has selected Purchase preference to Micro and Small Enterprises clause in the bid, the same will get precedence over this clause.

7. Purchase preference to Micro and Small Enterprises (MSEs): Purchase preference will be given to MSEs as defined in Public Procurement Policy for Micro and Small Enterprises (MSEs) Order, 2012 dated 23.03.2012 issued by Ministry of Micro, Small and Medium Enterprises and its subsequent Orders/Notifications issued by concerned Ministry. If the bidder wants to avail the Purchase preference, the bidder must be the manufacturer of the offered product in case of bid for supply of goods. Traders are excluded from the purview of Public Procurement Policy for Micro and Small Enterprises. In respect of bid for Services, the bidder must be the Service provider of the offered Service. Relevant documentary evidence in this regard shall be uploaded along with the bid in respect of the offered product or service. If L-1 is not an MSE and MSE Seller (s) has/have quoted price within L-1+ 15% (Selected by Buyer)of margin of purchase preference /price band defined in relevant policy, such Seller shall be given opportunity to match L-1 price and contract will be awarded for 25%(selected by Buyer) percentage of total QUANTITY.

8. Past Performance: The Bidder or its OEM {themselves or through re-seller(s)} should have supplied same or similar Category Products for 50% of bid quantity, in at least one of the last three Financial years before the bid opening date to any Central / State Govt Organization / PSU / Public Listed Company. Copies of relevant contracts (proving supply of cumulative order quantity in any one financial year) to be submitted along with bid in support of quantity supplied in the relevant Financial year. In case of bunch bids, the category related to primary product having highest bid value should meet this criterion.

9. Reverse Auction would be conducted amongst first 50% of the technically qualified bidders arranged in the order of prices from lowest to highest. Number of sellers eligible for participating in RA would be rounded off to next higher integer value if number of technically qualified bidders is odd (e.g. if 7 bids are technically qualified, then RA will be conducted amongst L-1 to L-4). In case number of technically qualified bidders are 2 or 3, RA will be between all without any elimination. If Buyer has chosen to split the bid amongst N sellers, then minimum N sellers would be taken to RA round. In case Primary products of only one OEM are left in contention for participation in RA based on lowest 50% bidders qualifying for RA, the number of sellers qualifying for RA would be increased to get at least products of one more OEM (directly participated or through its reseller) if available. Further, if bid(s) of any seller(s) eligible for MSE preference is / are coming within price band of 15% of Non MSE L-1 or if bid of any seller(s) eligible for Make in India preference is / are coming within price band of 20% of non MII L-1, then such MSE / Make in India seller shall also be allowed to participate in the RA process.

High End Multipara Monitor (16 pieces)

(Minimum 50% and 25% Local Content required for qualifying as Class 1 and Class 2 Local Supplier respectively/क्रमशः श्रेणी 1 और श्रेणी 2 के स्थानीय आपूर्तिकर्ता के रूप में अर्हता प्राप्त करने के लिए आवश्यक)

Brand Type/ब्रांड का प्रकार	Registered Brand
-----------------------------	------------------

Technical Specifications/तकनीकी विशिष्टियाँ

* As per GeM Category Specification/जेम केटेगरी विशिष्टि के अनुसार

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
Performance Parameters	Type of Monitor	Modular
	EtCO2	Yes
	Cardiac output	NA
	Transport monitor modulefor transporting within the hospital	No
	Screen size of transport module in inches	NA

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
	Battery backup for transport monitor in hrs	NA
	AGM (Anesthetic Gas Module)	Up-gradable
	NMT (Neuro Muscular Transmission)	Upgradable
	EEG (Electroencephalogram)	Upgradable
	AEEG (Amplitude Integrated Electroencephalography)Trend Graph	No
	Channels in EEG	8
	BIS monitoring	Up-gradable
	Entropy Monitoring	No
	Spirometry	Up-gradable
	Non invasive hemoglobin	No
	PVI (Pleth variability index)	No
	Perfusion Index	Yes
	Transcutaneous PO₂ / PCO₂	Up-gradable
	ScVO₂	Up-gradable
	S T Analysis	Yes
ECG Parameters	ECG Provided	Yes Or higher
	Number of leads of ECG	5
	12-Lead interpretation for ECG	Yes
	Accuracy	± 2 bpm or better
Respiration Parameters	Respiration rate measurement range	0-150 bpm or more
Pulse Oximetry Parameters	Measurement technique	Others
	Probe type	Clip
	Probe for neonates	Reusable
	SPO2 measurement range	70-100%
	Accuracy of SPO2 measurement	± 3% or better
NIBP Parameters	Measurement technique for NIBP	Linear Inflation
	Dual NIBP	Yes

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
	Number of channels for NIBP	2
	Blood pressure monitoring range Systolic	40-270 or more
	Diastolic BP	10-200 Or more
	Gradation	±1 mm/Hg or better
Temperature Parameters	Number of channels for temperature	2
	Type of probe	skin
IBP Parameters	Number of channels for IBP	2
ETCO ₂ Parameters	Sensor type	Mainstream
	Patient type	Both intubated and non intubated
	CO₂ Measuring range in mmHg	0 to 99 or better
Other Parameters	Length of time of trending in hr	≥72
Alarms	Activating silence of alarm	Yes
	Duration of alarm silence in sec	120
	Suspension of all alarms	Yes
	Latching of alarm	Yes
	Cuff leakage alarm	Yes
	Cuff disconnect alarm	Yes
	Alarm for inflation/deflation errors	Yes
	Indicator for low-battery	Yes
	EtCO₂ - low and high	Yes
	Escalation of Alarm Priority	Yes
	Sensor/ wire/probe disconnection alarm	Yes
Networking Parameters	Wireless	No
	Connectivity with Network printer	Yes
	Connectivity provision with Central Nursing Station	Yes
Display Parameters	Type of display	TFT
	Display Size,inch	12
	Traces of display	>14

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
	User Interface of display	Touchscreen
	"Inbuilt thermal Recorder (Printer)"	No
	Display must allow easy viewing at all ambient light levels	No
Standard Accessories	Number of reusable adult probe	1
	Number of reusable pediatric probe	1
	Number of reusable neonatal probe	1
	Number of Disposable neonate probes	NA
	Number of ECG 3/5 lead cable	1
	Disposable NIBP cuff for neonatal	NA
	EtCO2 accessories	2 each for adult , neonatal and pediatric
	Accessories to wall mount the monitor	Yes
Additional Accessories	ECG cables	No
	ECG cables for neonates	No
	IBP monitoring	No
	ETCO2 Monitoring	No
Power Requirements	Mains cable	3 m or more with Indian plug
	UPS	No
	Back up time of UPS in minutes	NA
	Type of battery	Ni-MH
	Operating time of battery in min	60
Miscellaneous Parameters	Availability of toll free facility for technical support maintained by OEM or authorized agencies	Yes
	Warranty (Option of comprehensive warranty is available through bidding only, which if opted will supersede normal warranty in the catalogue)	5

Comprehensive Maintenance	
Warranty of required product	5 Year
Comprehensive Maintenance Duration (Post Warranty)	5 Year

*Warranty displayed under the AMC/CMC Details section will supersede the warranty displayed under the catalog specification

Additional Specification Parameters - High End Multipara Monitor (16 pieces)

Specification Parameter Name	Bid Requirement (Allowed Values)
Specification as per ATC	Bidder must upload product Brouchure & item quoted must comply with the specification added in ATC.

* Bidders offering must also comply with the additional specification parameters mentioned above.

Consignees/Reporting Officer/परेषिती/रिपोर्टिंग अधिकारी and/ तथा Quantity/मात्रा

S.No./क्र. सं.	Consignee Reporting/Officer/ परेषिती/रिपोर्टिंग अधिकारी	Address/पता	Quantity/मात्रा	Delivery Days/डिलीवरी के दिन
1	Uttam Kant Sharma	492001,AllMS Raipur Tatibandh	16	15

Special terms and conditions-Version:2 effective from 21-10-2022 for category High End Multipara Monitor

1. All Provisions of Drugs and Cosmetics Act, 1940 and Rules made there under as amended till date will always be applicable. This will include all notifications issued by *Central Drugs Standard Control Organisation(CDSCO)*, Ministry of Health & Family Welfare (MoHFW) and Department of Pharmaceuticals (DOP), Ministry of Chemicals & Fertilizers time to time in this regard.
2. The sellers are registered on GeM based on self-declaration of valid Drug License, product certification, test reports etc. However, buyers must check and validate the details at their end for all applicable licenses and certifications e.g., validity and authenticity/genuineness of drug license, product certification, manufacturer certification/licenses, test reports etc.
3. The price offered by the seller shall not, in any case exceed the DPCO controlled price, if any, fixed by the Central/State Government, the Maximum Retail Price (MRP) and the selling price. The seller must reduce the prices if there is any reduction in DPCO ceiling price, if any.
4. **Comprehensive warranty:** Comprehensive warranty shall include preventive maintenance including calibration as per technical/ service /operational manual of the manufacturer, service charges and spares. During the warranty period commencing from date of the successful completion of warranty period, Service personnel shall visit each consignee site as recommended in the manufacturer's technical/ service /operational manual, at least once in six months. warranty shall not be including the consumables. Further there will be 98% uptime warranty during warranty period on 24 (hrs) X 7 (days) X 365 (days) basis, with penalty, to extend warranty period by double the downtime period.

5. **Service centres:** Details of Service outlets in India to render services for equipment to be furnished to buyer/consignees with complete address, telephone numbers, e mails etc at time of making the supplies. It shall be the responsibility of seller to ensure that authorized service centres are available to cater to the areas where supplies are made within reasonable distance from where the service calls can be handled. Details of toll-free numbers for service call and online registration of service requests also to be provided buyer/consignee at the time of supplies.
6. **Source of supply:** It shall be responsibility of seller to provide Documents regarding source of equipments such as copy of Performa invoice or any other documents to establish that the products supplied are manufactured by OEM indicated and sourced from them.
7. **Packing and Marking:** Medical equipments being very delicate and sensitive packing for the goods should be strong and durable enough to withstand transit including transshipment (if any), rough handling, open storage etc. without any damage, deterioration etc. .The size, weights and volumes of the packing cases, remoteness of the final destination of the goods, availability or otherwise of transport and handling facilities at all points during transit up to final destination,. Quality of packing, the manner of marking within & outside the packages and provision of accompanying documentation shall take into consideration the type of medical equipments being supplied. The accessories shall be suitably labelled and packed. Each of the package shall be marked on three sides with indelible paint of proper quality: indicating contract number and date, brief description of goods including quantity, Packing list reference number, country of origin of goods and any other relevant details.
8. **Spare Parts:** Seller shall provide materials, information etc. pertaining to spare parts manufactured and supplied by the OEM. It shall be ensured that the required spares are available for purchase at least for 10 years from date of supplies. In case due to any reasons the production of the spare parts is discontinued sufficient advance notice should be given to the buyer/consignee before such discontinuation to provide adequate time to purchase the required spare parts etc. Further, OEM and their service centres/dealers shall carry sufficient inventories to assure ex-stock supply of consumables and spares for the equipments so that the same are available.
OEM or reseller shall always accord most favoured client status to the buyer/consignee and shall give the most competitive price for spares and consumables of its machines/equipments supplied.
9. **Installation, Training, Manuals:** Seller shall be responsible to carry out Installation & commissioning, Supervision and Demonstration of the goods. They shall provide required jigs and tools for assembly, minor civil works for the completion of the installation and Training of Consignee's representatives for operating and maintaining the equipment and supplying required number of operation & maintenance manual for the goods. In case the category parameters are specifying any requirements regarding the installations, training and manuals the same shall also be applicable.
10. **Electrical safety checking:** Sellers are required to make sure that they furnish the list of equipments for carrying out routine and preventive maintenance to buyer/consignee .They should make sure to periodically check the electrical safety aspects as per BIS Safety Standards or equivalent .In case they do not have required equipment for such testing should ensure that the equipments checked for electrical safety compliance through labs with facilities for such checking during every preventive maintenance call.
11. **Software:** All software updates should be provided free of cost during warranty period.
12. Any other Terms and Conditions which is not included or at variance with the conditions specified in STC/GTC, may be added by the buyer through Additional Terms and Conditions in the bid to ensure items are procured from authentic/validated source with appropriate and applicable quality. The above terms and conditions are in reverse order of precedence i.e. ATC shall supersede specific STC which shall supersede General Terms and Conditions ("GTC"), whenever there are any conflicting provisions.

Buyer Added Bid Specific Terms and Conditions/क्रेता द्वारा जोड़ी गई बिड की विशेष शर्तें

1. Experience Certificate for the supply of the same to any Govt/ PSU/ any renowned private organisation along with Supply/ Purchase Order.
2. If the agency is registered under MSME or NSIC, then EMD exemption certificate needs to be enclosed.
3. Make in india specific authorisation certificate needs to be enclosed.
4. **Buyer Added Bid Specific ATC**

Buyer Added text based ATC clauses

Specification for 16 units High End Multipara Modular Monitor along with Central Monitor Having Remote Viewing Facility

1. Multipara Monitor should be advanced Modular type monitor 15" TFT color Display with touch screen facility with resolution of 800 X 600 dots. It should have HL7 gateway to connect the HIS & CIS.
2. Monitor must be capable of simultaneously monitoring the following parameters which should be present as standard in all monitors: ECG, NIBP, SpO₂, Respiration, Dual temperature, HR.
3. Ports should be available in all the monitors for the following parameters: Dual IBP, Mainstream ETCO₂, BIS & Cardiac Output.
4. All the monitors should be upgradeable to connect EEG- 8 channel, NMT, Transcutaneous tcPO₂, Regional Oxygen Saturation (rSO₂).
5. Both invasive & non-invasive method of ETCO₂ measurement should be available.
6. Non-invasive continuous cardiac output along with Hemodynamic graph to view Hemodynamic condition of patient based on Frank-Starling principle and EGDT protocol should be measured through basic parameters like ECG, SpO₂ and NIBP/IBP, there should be no need of dedicated sensor or catheter for its measurement.
7. Drug calculation, Lung function, Hemodynamic data and Oxy CRG screen should be available as standard.
 8. Monitor should have facility to display at least 12 waveforms
 9. Patient modes Adult, Pediatric, & Neonate
 10. ECG 3/5L Input : Isolated and floating 3/5 leads. Protect against surges produced by Electro-surgical and Defibrillator potentials and should have built-in advance arrhythmia monitoring on all leads.
 11. Monitor should have facility to monitor 12 Lead ECG through 10 Lead ECG cable along with ECG analysis & report.
 12. Should have facility to detect advance arrhythmias (at least 23 different types) and should show number of arrhythmias occurred with ECG waveform for last 72 hours.
 13. Monitor should have facility to store minimum 1600 arrhythmia episode.
 14. Monitor should have facility for Full Disclosure of ECG and 4 other parameters of last 24 hours.
 15. Monitor should have facility of ST Recall
 16. Monitor should have Facility of Interbed display up to 20 beds.
 17. Heart Rate Range Adult : 30-200 bpm, Child and neonate - 30 to 250 beats/min.
 18. PR Source : Auto/SpO₂/NIBP

19. Respiration Range : 0-150 breaths/min.
20. Temperature – Measurement range : 1 C-45 degree C, Unit : C or F, user selectable
21. SpO₂: Measurement should be possible in low perfusion with any of the following technologies: Masimo SET/Nihon Kohden/Fast/ NellcoreOximax. (Clinical papers for low perfusion measurement to be submitted)
 - a. Measurement Method : Two wavelength light absorption method.:-
 - b. Measurement range :0-100% Accuracy Adults : 70-100% +/- 2 digit , 0-69% unspecified. Accuracy Neonates : 70-100% +/- 3 digits, 0-69% unspecified, Pulse rate range : 30 to 300 bpm.
22. Monitor should display perfusion Index(PI%) from SPO₂ as an indication of pulse strength at the sensor site.
23. NIBP Method :Oscillometric, Display : Systolic, Diastolic and Mean, Modes : Manual, Auto,Stat&Venius puncture mode. Auto mode intervals : 2,4,5,10,15,30,60,90,120,240 and 360 mins.Unit : mmHg or kPa, Range : 0-300 mmHg, Accuracy : +/- 3 mmHg
24. Monitor should have facility to take blood pressure measurement automatically in case of sudden blood pressure change in between periodic measurements of blood pressure.
 25. Alarms : Equipment alarms : Audio (Alarm Beep) Visual (Flashing Blue LED), Patient alarms : Audio (Alarm Beep), Red LED (High priority), Yellow LED (Medium Priority)
 26. Alarms suspend : Continuous RED LED with display of alarm crossed bell.
27. Machine must have facility for alarm escalation in case nursing staff miss the alarm.
28. Monitor shall provide the capability to interact with alarms at remote bedside monitor.
29. Alarm limit status (ON/OFF) must be indicated on-screen for each parameter and actual parameter alarm settings must be displayed on-screen when alarms are on.
 30. Invasive Blood Pressure (IBP) – Calculation – CPP,PPV,SPV,CVP-ET
 - a. Auto zero balancing range +/- 200 mmHg Auto
 - b. Zero balancing accuracy: +/- 1mmHg
31. Monitor should have facility to measure dynamic preload parameter such as Pulse pressure variation, or Systolic pressure variation.
 32. ETCO₂ – Main Stream – CO₂ measuring range – 0 to 100mmHg
 33. CO₂ value display update cycle every 4 sec or when alarm is generated.
34. Must be able to store & display the 72 hours beat by beat waveforms for 4 or more selected parameter by the user.

35. Trends-Data storage : 24 Hrs up to 6 parameters can be user selectable f or 3 separate graphical windows.

36. Graphical Trend :

37. Tabular Trend : 30 sec. 1min, 2min, 4min, 8min, 15min, 30min & 60min

38. Alarm Trend (Recall) : minimum 8000 events

39. Monitor should be upgradable to provide remote viewing of real time wa veforms throughinternet.

40. Battery Back up – Minimum 1Hour or higher

41. Monitor should have following certifications –

- Monitor should be “European CE/USFDA/BIS/ISO” approved. certific ate to be enclosed in the technical bid.
- IEC60601-1-2:2001 certificate to be enclosed in the technical bid.

42. Demonstration of the equipment is a must.

43. All modules should be compatible with all monitors quoted.

44. Warranty: 5 years from the date of installation.

45. CMC: 5 Years after the end of the warranty.

46. Following Accessories of the monitor to be supplied per 16 unit of monito r.

A. ECG Cable	16 units
B. SPO2 sensor (adult)	16units
C. SPO2 sensor (neo)	4 units
D. NIBP Cuff (adult)	16 units
E. NIBP Cuff (Pedia)	16 units
F. NIBP Cuff (Neo)	5 units
G. IBP Cable	32 units
H. Skin Temp. probe	16 units
I. Rectal Temp. Probe	8 units
J. Mainstream EtCO2	8 units
K. EEG- 8 Channel	1units
L. BIS	4 units

M. Non-Invasive C.C.O. along with Hemodynamic Graphs

1units

5. **Warranty**

Bidder / OEM has to give an undertaking that after expiry of warranty period, it will provide Comprehensive Maintenance Service for next 5 years for the offered products at the rate not more than 5 % of contract price per annum. Buyer reserves the right to enter into a CMC agreement with the Successful Bidder / OEM after expiry of the Warranty period at above mentioned rate and the payment for the CMC charges would be made Annually after rendering of the CMC Services of the relevant CMC period. Performance Security of the successful bidder shall be forfeited if it fails to accept the CMC contract when called upon by the buyer. CMC would include cost of

31,04,080

(Upload the undertaking). The original Performance Security of contract will be returned only after submission and verification of AMC Performance Security for 3% of total CMC value valid up to CMC period plus 2 months (if there is no other claim).

6. **Inspection**

Nominated Inspection Agency: On behalf of the Buyer organization, any one of the following Inspection Agency would be conducting inspection of stores before acceptance:
Pre-dispatch Inspection at Seller Premises (applicable only if pre-dispatch inspection clause has been selected in ATC):

NA

Post Receipt Inspection at consignee site before acceptance of stores:
By user department i.e CCU, Anaesthesiology AIIMS Raipur

7. **Scope of Supply**

Scope of supply (Bid price to include all cost components) : Supply Installation Testing Commissioning of Goods and Training of operators and providing Statutory Clearances required (if any)

8. **Service & Support**

Availability of Service Centres: Bidder/OEM must have a Functional Service Centre in the State of each Consignee's Location in case of carry-in warranty. (Not applicable in case of goods having on-site warranty). If service center is not already there at the time of bidding, successful bidder / OEM shall have to establish one within 30 days of award of contract. Payment shall be released only after submission of documentary evidence of having Functional Service Centre.

9. **Service & Support**

Dedicated /toll Free Telephone No. for Service Support : BIDDER/OEM must have Dedicated/toll Free Telephone No. for Service Support.

10. **Service & Support**

Escalation Matrix For Service Support : Bidder/OEM must provide Escalation Matrix of Telephone Numbers for Service Support.

Disclaimer/अस्वीकरण

The additional terms and conditions have been incorporated by the Buyer after approval of the Competent Authority in Buyer Organization, whereby Buyer organization is solely responsible for the impact of these clauses on the bidding process, its outcome, and consequences thereof including any eccentricity / restriction arising in the bidding process due to these ATCs and due to modification of technical specifications and / or terms and

conditions governing the bid. Any clause(s) incorporated by the Buyer regarding following shall be treated as null and void and would not be considered as part of bid:-

1. Definition of Class I and Class II suppliers in the bid not in line with the extant Order / Office Memorandum issued by DPIIT in this regard.
2. Seeking EMD submission from bidder(s), including via Additional Terms & Conditions, in contravention to exemption provided to such sellers under GeM GTC.
3. Publishing Custom / BOQ bids for items for which regular GeM categories are available without any Category item bunched with it.
4. Creating BoQ bid for single item.
5. Mentioning specific Brand or Make or Model or Manufacturer or Dealer name.
6. Mandating submission of documents in physical form as a pre-requisite to qualify bidders.
7. Floating / creation of work contracts as Custom Bids in Services.
8. Seeking sample with bid or approval of samples during bid evaluation process.
9. Mandating foreign / international certifications even in case of existence of Indian Standards without specifying equivalent Indian Certification / standards.
10. Seeking experience from specific organization / department / institute only or from foreign / export experience.
11. Creating bid for items from irrelevant categories.
12. Incorporating any clause against the MSME policy and Preference to Make in India Policy.
13. Reference of conditions published on any external site or reference to external documents/clauses.
14. Asking for any Tender fee / Bid Participation fee / Auction fee in case of Bids / Forward Auction, as the case may be.

Further, if any seller has any objection/grievance against these additional clauses or otherwise on any aspect of this bid, they can raise their representation against the same by using the Representation window provided in the bid details field in Seller dashboard after logging in as a seller within 4 days of bid publication on GeM. Buyer is duty bound to reply to all such representations and would not be allowed to open bids if he fails to reply to such representations.

Additional Clause For Comprehensive Maintenance Charges

- 1.CMC shall include preventive maintenance including calibration as per technical/ service /operational manual of the manufacturer, service charges and spares, after satisfactory completion of Warranty. During the CMC period commencing from date of the successful completion of warranty period, Service personnel shall visit each consignee site as recommended in the manufacturer's technical/ service /operational manual, at least once in six months or as per user requirement. Cost of consumables shall not be included in CMC.Further there will be 98% uptime warranty during CMC period on 24 (hrs) X 7 (days) X 365 (days) basis, with penalty, to extend CMC period by double the downtime period.
- 2.CMC charges to be indicated as percentage of cost of equipment quoted for each year after the warranty period.
- 3.GST shall be included in the CMC Charges quoted.
- 4.Cost of CMC will be added for Ranking/Evaluation purpose with depreciation formula.A 10% discounting rate per year shall be applied on CMC Charges for price evaluation on Net Present Value.
- 5.The payment of CMC will be made on quarterly basis after satisfactory completion of said period, duly certified by end user.
- 6.While creating a bid or RA, buyers shall indicate whether CMC is required against Yes/No" options. If CMC Charges are included, an option for number of years for CMC required after the warranty period shall be available.Under this option up to 10 years can be chosen for CMC charges beyond warranty period.
- 7.In case the bid has a provision for CMC, the warranty of the product will also be deemed to have been converted into Comprehensive warranty including preventive maintenance and calibration as per technical/ service /operational manual of the manufacturer, service charges and spares, during the Warranty Period also. Sellers are therefore advised to include the cost of Comprehensive Warranty including spares (excluding consumables) also in product Cost.
- 8.The CMC functionality shall be available in bid only and no direct RA shall be applicable.In case of bid to R/A decrement rules shall be applicable on total price inclusive of CMC charges. Bunching of products shall not be available while creating bids with CMC charges.

8.1. Buyer shall indicate number of years of warranty by selecting different options available in the field depending on warranty parameter applicable in category parameters for the equipment. No. of years of warranty indicated here shall supersede the warranty period indicated elsewhere in bid or product specifications. The Seller while participating in Bid/RA will get fields to indicate CMC charges as percentage depending on number of years of CMC selected by Buyer. The following shall be applicable, if 5 year CMC selected:

- CMC charges for 1st year after warranty period- Percentage to be indicated- A1
- CMC charges for 2nd year after warranty period- Percentage to be indicated- A2
- CMC charges for 3rd year after warranty period - Percentage to be indicated- A3
- CMC charges for 4th year after warranty period - Percentage to be indicated- A4
- CMC charges for 5th year after warranty period - Percentage to be indicated- A5

Similarly, A6 to A10 are to be indicated for 6th to 10th year of CMC if applicable.

8.2. The calculation of CMC Charges shall take into account the number of years of warranty and duration of CMC as specified while creating bid.

8.3. In the price evaluation, the system shall provide function to calculate the cost of each equipment by formula indicated below including CMC and then show the inter-se-ranking of the bidders. The following are the variables

- (i) Number of years for which CMC required.
- (ii) Number of years of product warranty

The formula for calculating total cost including CMC charges shall be as under:

Total Cost for evaluation=

$C + C * \{ (A1/100)/(1.10^n) + (A2/100)/(1.10^{n+1}) + (A3/100)/(1.10^{n+2}) + (A4/100)/(1.10^{n+3}) + (A5/100)/(1.10^{n+4}) \}$ and so on

C - Cost for equipment quoted and n shall be number of years of product warranty specified.

If 2 year warranty specified, n shall be 2 and if 5 year warranty specified, n shall be 5. A1, A2, A3, A4 & A5 shall depend on how many years CMC selected. For 3 year CMC, only A1, A2 and A3 factors are to be taken into account and A4 and A5 will not be applicable.

8.4. CMC charges offered for each subsequent year should be same or higher than preceding year.

8.5. The CMC charges shall be offered within range of 3 to 10% of cost of equipment.

9. Since CMC charges are to be paid only later for each year during CMC period, applicable performance guarantee amount after placement of contract shall be based on the cost of equipment excluding the cost of CMC Charges.

10. Performance bank guarantee applicable for CMC is to be submitted at start of the CMC and shall be applicable between 2.5% to 5% as specified in bid on total CMC Charges. The PBG submitted after award of contract shall be released only after new PBG for the CMC period is submitted and accepted by buyer/consignee after due verification. Bank guarantee for CMC is to remain valid till completion of CMC period plus one year. The bank guarantee for CMC shall be submitted to buyer directly. In case, seller fails to submit the PBG or does not provide services for the CMC contract after expiry of warranty period then PBG of equipment shall be forfeited.

11. In case of splitting of order quantity, equipment cost and CMC charges offered by L1 bidder shall be matched by higher quoting eligible bidders on one-to-one basis. The equipment cost and CMC charges (year to year) shall be matched individually.

12. The CMC Contract shall be an offline contract to be handled by buyer. The payment of CMC will be made on quarterly basis after satisfactory completion of said period, duly certified by end user and scope of CMC will be as per para 1 above.

13. CMC Charges are inclusive of all the charges for Transportation, Lodging, Boarding, all insurances including third party insurance and all other incidental charges. The same shall include GST. The prices also include cost of spares and damaged parts. Purchaser does not have any liability, whatsoever, over and above the cost of CMC. It also includes for arranging hand tools & tackles, special tools etc. required to carry out the work.

[This Bid is also governed by the General Terms and Conditions/ यह बिड सामान्य शर्तों के अंतर्गत भी शासित है](#)

In terms of GeM GTC clause 26 regarding Restrictions on procurement from a bidder of a country which shares a land border with India, any bidder from a country which shares a land border with India will be eligible to bid in this tender only if the bidder is registered with the Competent Authority. While participating in bid, Bidder has to undertake compliance of this and any false declaration and non-compliance of this would be a ground for immediate termination of the contract and further legal action

in accordance with the laws./जेम की सामान्य शर्तों के खंड 26 के संदर्भ में भारत के साथ भूमि सीमा साझा करने वाले देश के बिडर से खरीद पर प्रतिबंध के संबंध में भारत के साथ भूमि सीमा साझा करने वाले देश का कोई भी बिडर इस निविदा में बिड देने के लिए तभी पात्र होगा जब वह बिड देने वाला सक्षम प्राधिकारी के पास पंजीकृत हो। बिड में भाग लेते समय बिडर को इसका अनुपालन करना होगा और कोई भी गलत घोषणा किए जाने व इसका अनुपालन न करने पर अनुबंध को तत्काल समाप्त करने और कानून के अनुसार आगे की कानूनी कार्यवाई का आधार होगा।

---Thank You/धन्यवाद---

