



अखिल भारतीय आयुर्विज्ञान संस्थान, रायपुर (छत्तीसगढ़)  
**All India Institute of Medical Sciences, Raipur (Chhattisgarh)**

Tatibandh, GE Road,  
Raipur-492 099 (CG)  
[www.aiimsraipur.edu.in](http://www.aiimsraipur.edu.in)

No: AIIMS/R/CS/ENT/1111/20/PAC

Date: 06.04.2021

**Call for Objection**

**Subject:** Inviting comments/objection, if any before declaring proprietary article for procurement of "Microdebrider Handpiece" for Department of ENT & HNS AIIMS, Raipur.

ENT & HNS Department AIIMS, Raipur has to procure Microdebrider Handpiece through Proprietary Article basis.

The proposal submitted by M/s Medtronic Inc USA who is sole manufactures and M/s Bagree Enterprises Raipur is the Local Agent of this item along with Proprietary Article Certificate are attached & uploaded on Institute website.

The above documents are being uploaded for open information to submit objections, comments if any from any manufacturer/supplier before declaring proprietary article of the said equipment/items to be procured, within 15 days (i.e. up to 21.04.2021.) from the date of issuance/uploading of the notification.

The objection should be raised in the technical compliance sheet as enclosed, if any Firm claiming suitability of their product with respect to specification mentioned.

The comments should be sent to the office of Central Store Office on above address at AIIMS Raipur in a sealed envelope with above reference on or before 21.04.2021 up to 05:00 PM from the date of uploading on institutional website, failing which it will be presumed that any other manufacture/vendor is having no comment to offer and case will be decided on merits.

**Stores Officer  
AIIMS Raipur (CG)**

**भंडार अधिकारी (के क्रय)  
Stores Officer (CP)  
एन्स, रायपुर (छ.ग.)  
AIIMS Raipur (C.G.)**

Enclosure:-

1. PAC Certificate by Manufacturer.
2. PAC Certificate by Department.
3. Technical Specification.



(21)

अखिल भारतीय आयुर्विज्ञान संस्थान, रायपुर (छ.ग.)  
All India Institute of Medical Sciences, Raipur (Chhattisgarh)  
Tatibandh, GE Road,  
Raipur-492 099 (CG)  
[www.aiimsraipur.edu.in](http://www.aiimsraipur.edu.in)

स्वामित्व प्रमाण पत्र  
Proprietary Article Certificate

फाइल संख्या और संदर्भ File Number and Reference		microckbrides metal prep
1	सामाग्री का विवरण Description of article	straight shot m.s micro-drill bit, style drill & sheet
2	पूर्वानुमानित मात्रा / वार्षिक आवश्यकता Forecast of quantity/annual requirement	each-1
3	उपरोक्त मात्रा हेतु अनुमानित मूल्य Approximate estimated value for above	
4	निर्माता का नाम एवं पता Maker's name and address	Indig Medhonic Pvt Ltd. 1241, Solitaire park B.n-12, 4th floor Link Road, Andheri (E), Mumbai, m.h.
5	अधिकृत डीलर/स्टाकिस्ट का नाम Name(s) of authorised dealers/stockists	Bagree Enterprise Near City Kotwadi Chhatrapati Raipur (C.G.)
6	<p>मैं पी ए सी के आधार पर उपरोक्त खरीद को स्वीकार करता हूँ और यह प्रमाणित करता हूँ कि:</p> <p>नोट- (बी), (सी-1) या (सी-2) में से केवल एक को बनाए रखने के लिए टिक करें, जो भी लागू हो और दूसरो को काट दें। कृपया (ए) टिक कर पुष्टि करें इसके बिना पीएसी प्रमाण पत्र अवैध होगा</p> <p>I approve the above purchase on PAC basis and certify that:-</p> <p>Note- Tick to retain only one out of (b), (c-1) or (c-2) whichever is applicable and cross out others.</p> <p>Please do confirm (a) by ticking it - without which PAC certificate will be invalid.</p>	
6 (a)	यह एकमात्र फर्म है जो इस मद का निर्माण/संग्रहण कर रहा है। और This is the only firm who is manufacturing /stocking this item. AND	✓
6 (b)	किसी अन्य फर्म द्वारा समरूप मद निर्मित/विक्रय नहीं किया जाता है, जिसका उपयोग इसके बदले किया जा सकता है। अथवा A similar article in not manufacturing/sold by any other firm, which could be used in lieu OR	✓
6 (c-1)	कोई अन्य मेक/ब्रांड निम्नलिखित कारणों (जैसे ओईएम/वारंटी के) के लिए उपयुक्त नहीं होगा। अथवा No other make/brand will be suitable for following tangible reasons (like OEM/warranty spares): OR	✗

	<p>पिछले खरीद में भी दिया गया था, तो कृपया इसके बाद से अधिक स्रोतों का पता लगाने के लिए प्रयास करें) तथा</p> <p>No other make/brand will be suitable for following intangible reasons (if PAC was also given in the last procurement cycle, please also bring out efforts made since then to locate more sources): OR</p>	
7	<p>प्रस्ताव के लिए वित्त शाखा की सहमति का संदर्भ (कार्रवाई भंडार और लेखा विभाग द्वारा की जायेगी)</p> <p>Reference of concurrence of finance wing to the proposal (Action will be taken by stores &amp; Account Department)</p>	

पिछले तीन सालों में इस मद की पीएससी खरीद का इतिहास नीचे दिया जा सकता है (यदि कोई हो) History of PAC purchase of this item for past three years may be given below (if any)

प्रदायक का नाम Name of the Supplier			
आदेश/निविदा संदर्भ और दिनांक Order/Tender reference & Date	आदेशित मात्रा Quantity Ordered	आदेश पर मूल दर (₹) Basic Rate on order (Rs.)	प्रतिकूल प्रदर्शन रिपोर्ट अगर कोई हो Adverse Performance Reported if any

अनुमोदन करने वाले प्राधिकारी का हस्ताक्षर

*Rupesh Mehta*

डॉ. रुपेश मेहता

Dr. Rupesh Mehta

सह-अध्यापक (नोक, कान एवं गला विभाग)

Associate Professor (ENT & HNS)

राष्ट्रिय भारतीय अयुर्विज्ञान संस्थान रायपुर (छ.ब.)

AIIMS Raipur (C.G.)

दिनांक 29/1/2020

अधिकारी का पदनाम

# Medtronic

India Medtronic Pvt. Ltd.  
Atty. (201) 10044700/27004701 (3)  
R. 43, Solitaire Corporate Park  
Bldg No. 12, 4th Floor  
Andheri (East) Mumbai - 400 093 India  
www.medtronic.co.in  
for +91 22 33074700  
fax +91 22 33074704

AIIMS Raipur

20  
2

Date: 16th November 2019

To  
The Director  
All India Institute of Medical Sciences  
Raipur

Dear Sir

## OEM & PROPRIETARY ARTICLE CERTIFICATE (PAC)

This is to certify that the items listed below Medtronic Straight Shot M5 Micro-Dibridder; Stylus Drill & Skeeter are the proprietary items of Medtronic Inc. with address at 7000 Central Avenue NE, Minneapolis MN 55432, USA. They are being manufactured at Medtronic Xomed, Inc. 6743, Southpoint Drive North, Jacksonville, FL-32216, USA

S. No.	Catalogue number	Description
1	1899200	Straight Shot M5 Micro-Dibridder
2	EM-200	Stylus Drill
3	3055601	Skeeter

We also certify that only Drills, Debridder, Sketter manufactured by Medtronic are compatible with Medtronic IPC Console and no other manufacturer's Drills, Debridder or Sketter are Compatible with the above console.

Yours faithfully  
For and on behalf of India Medtronic Private Limited

Digitally signed  
by Krishna  
Balijepalli  
Date:  
2019.11.18  
11:28:17 +05'30'

Krishna Balijepalli  
Sr. Manager – Commercial Operations - RTG  
Cell: +919620155514  
e-mail: [krishna.balijepalli@medtronic.com](mailto:krishna.balijepalli@medtronic.com)

Q

## Technical specification for "Microdebrider Handpiece" for Department of ENT & HNS

### Specifications for M5 Handpiece :

- Able to work up to the speed of 30000 RPM in forward rotation and 5000RPM in oscillation mode.
- Have fingertip control to rotate only the tip of the blade up to 360 deg.
- Have straight suction path to reduce clogging and allow efficient tissue removal.
- Have integrated blade locking system to lock the blade tip rotation.
- Have integrated side grooves and cable clips to provide better tubing management.
- Have Titanium body to avoid rusting.
- Should come with Zero, 12, 40,60,90 and 120 degree blades
- Storage box for all items.
- For use in FESS & Anterior skull base surgery (Hand piece with straight & angled attachment)

### Specifications for electric drill Handpiece for Skull base surgeries

- 1 Should be compatible with existing Medtronic IPC console
- 2 Drill handpiece should be ergonomically designed and easy to hold while operating.
- 3 The system should have electric motor which can provide torque upto 43mNM.
- 4 Motor Speed should be variable from 10000 to 75000 RPM.
- 5 Weight of the drill should not be more than 100grams and length should be less than 8.0 cm with diameter not exceeding 1.70 cm.
- 6 Should have integrated cable to connect to console.
- 7 Motor should not require any lubrication for movement.
- 8 Handpiece should have quick release and lock system for tools and attachment.
- 9 The handpiece should be provided with three straight and three angled attachment required for lateral skull base surgeries - 7.5 cm preferred.
- 10 The handpiece should be compatible to trans-nasal skull base burrs which should be attachable directly to motor.
- 11 High quality store box for all items.

### Stapes Drill

- The burs should be color coded
- Technical Specs of drill
- Weight: 56.8 Grams
- Diameter: 2mm and angled at 15° from the plane of the handpiece.
- Power: (3) 9-Volt batteries housed in the accelerator foot control.
- Speed : Upto 12,000 r/min
- Skeeter burs -3 each ( 2mm, 2.5 mm, 3mm)
- Storage box for all



**Note:**

- 5 year CMC and 5 years warranty is required including all accessories
- System should be BIS/ISO/European CE/USFD Approved certificate of both standards should be attached with tender.
- All the equipment including all accessories should be provided with detailed data sheet with the original brochure of the quoted model (Photocopies/Print outs are not accepted)
- List of major installation (minimum 5) in last two years should be provided.
- Training for user department, doctors and staff should be provided. As per satisfaction of buyer.
- Storage box for all

Objection should be submitted in following format:

S. no.	Item Specification as given	Specification offered by firm	Deviation if any	Remark
	<b><u>Specifications for M5 Handpiece :</u></b> <ul style="list-style-type: none"> <li>• Able to work up to the speed of 30000 RPM in forward rotation and 5000RPM in oscillation mode.</li> <li>• Have fingertip control to rotate only the tip of the blade up to 360 deg.</li> <li>• Have straight suction path to reduce clogging and allow efficient tissue removal.</li> <li>• Have integrated blade locking system to lock the blade tip rotation.</li> <li>• Have integrated side grooves and cable clips to provide better tubing management.</li> <li>• Have Titanium body to avoid rusting.</li> <li>• Should come with Zero, 12, 40,60,90 and 120 degree blades</li> <li>• Storage box for all items.</li> <li>• For use in FESS &amp; Anterior skull base surgery (Hand piece with straight &amp; angled attachment)</li> </ul>			
	<b><u>Specifications for electric drill Handpiece for Skull base surgeries</u></b> <ul style="list-style-type: none"> <li>• Should be compatible with existing Medtronic IPC console</li> <li>• Drill handpiece should be ergonomically designed and easy to hold while operating.</li> <li>• The system should have electric motor which can provide torque upto 43mNM.</li> <li>• Motor Speed should be variable from 10000 to 75000 RPM.</li> <li>• Weight of the drill should not be more than 100grams and length should be less than 8.0 cm with diameter not exceeding 1.70 cm.</li> <li>• Should have integrated cable to connect to console.</li> <li>• Motor should not require any lubrication for movement.</li> <li>• Handpiece should have quick release and lock system for tools and attachmnet.</li> <li>• The handpiece should be provided</li> </ul>			1

	<p>with three straight and three angled attachment required for lateral skull base surgeries - 7.5 cm preferred.</p> <ul style="list-style-type: none"> <li>• The handpiece should be compatible to trans-nasal skull base burrs which should be attachable directly to motor.</li> <li>• High quality store box for all items.</li> </ul>			
	<p><b>Stapes Drill</b></p> <ul style="list-style-type: none"> <li>• The burs should be color coded</li> <li>• Technical Specs of drill</li> <li>• Weight: 56.8 Grams</li> <li>• Diameter: 2mm and angled at 15° from the plane of the handpiece.</li> <li>• Power: (3) 9-Volt batteries housed in the accelerator foot control.</li> <li>• Speed : Upto 12,000 r/min</li> <li>• Skeeter burs -3 each ( 2mm, 2.5 mm, 3mm)</li> <li>• Storage box for all</li> </ul>			
	<p><b>Note</b></p> <ul style="list-style-type: none"> <li>• 5 year CMC and 5 years warranty is required including all accessories</li> <li>• System should be BIS/ISO/European CE/USFD Approved certificate of both standards should be attached with tender.</li> <li>• All the equipment including all accessories should be provided with detailed data sheet with the original brochure of the quoted model (Photocopies/Print outs are not accepted)</li> <li>• List of major installation (minimum 5) in last two years should be provided.</li> <li>• Training for user department, doctors and staff should be provided. As per satisfaction of buyer.</li> <li>• Storage box for all</li> </ul>			1



# Medtronic

**India Medtronic Pvt. Ltd.**

CIN: U33110MH1993PTC204814  
1261, Solitaire Corporate Park, Bldg.  
No. 12, 6th Floor, Andheri - Ghatkopar  
Link Road, Andheri (East),  
Mumbai - 400093, India.  
www.medtronic.co.in

tel +91-22 48810700/701

fax +91-22 48810704

16

February 19, 2021

To,  
H.O.D (ENT Department)  
AIIMS Raipur

**Subject: : Letter of Authorization for submission of Quotation in AIIMS Raipur**

Dear Sir/Ma'am,

We, India Medtronic Private Limited, having our registered office address at 1261, Solitaire Corporate Park, Bldg. No. 12, 6th Floor, Andheri - Ghatkopar Link Road, Andheri (East), Mumbai - 400093, India. ("India Medtronic"), are the owned subsidiary of Medtronic Inc, with address at 710 Medtronic Parkway, Minneapolis, Minnesota 55432, USA., who are the manufacturers of Medtronic range of medical devices, hereby certify that Bagree Enterprise having its principal place of business at Near City Kotwal, Chhottapara, Raipur - 492001 is as on today and until 30/04/2021 authorized distributor of India Medtronic Private Limited for the resale of "Medtronic ENT Range of Products" to your institution.

This Letter of authorization has been issued to Bagree Enterprise only for the purpose of submission of Quotation Ref No: BE/322 Dated 12.02.2021 with your Institution by Bagree Enterprise.

Yours faithfully

**For India Medtronic Private Limited**

**MOHAMMED  
TAREQUE**

Digitally signed by  
MOHAMMED TAREQUE  
Date: 2021.02.19 14:21:02  
+05'30'

Mohammed Tareque  
Authorized Signatory