



अखिलभारतीयआयुर्विज्ञानसंस्थान, रायपुर(छत्तीसगढ़)
All India Institute of Medical Sciences, Raipur (Chhattisgarh)
Tatibandh, GE Road,
Raipur-492 099 (CG)
www.aiimsraipur.edu.in

AIIMS/R/HS/NEPH/5008s/2022/PAC

Date: 15/06/2022

Call for Objections

Sub:- Inviting Comments/objections (if any), regarding procurement of Diasafe Plus Filter and Blood Tubing Set for FMC 5008s Hemodialysis Machine for Nephrology Department, AIIMS Raipur under Proprietary Article.

Nephrology Department, AIIMS Raipur has to procure the Diasafe Plus Filter and Blood Tubing Set for FMC 5008s Hemodialysis Machine for Nephrology Department through Proprietary Article Basis under GFR -166. The proposal submitted by M/s Fresenius Medical Care India Private Limited who is sole manufacturer and M/s Balaji Surgical, Raipur is the local Agent of this item along with Proprietary Article Certificate are attached and same upload on AIIMS Raipur website.

The above documents are being uploaded for open information to submit objections/comments (if any) from any manufacturer/supplier before declaring proprietary article for the said equipment/items to be procured, within 10 days (i.e. 25/06/2022) from the date of issuance/uploading of the notifications. The objection should be raised on the technical compliance sheet as enclosed, if any firm claiming suitability of their product with respect to specification mentioned.

The comments should be sent to the office of Stores Office (Hospital), Room No. 329, C C1 Block, Gate No. 01, AIIMS, Raipur in sealed envelope or through mail storesofficer.hp@aiimsraipur.edu.in on or before 25/06/2022 up to 3.00 PM failing which it will be presumed that any other vendor having no comments to offer and case will be decided on merits.

**Stores Officer- H
AIIMS Raipur**

Encl:-

1. Proprietary Letter by Manufacturer
2. Proprietary Letter by Department
3. Authentication letter of vendor
4. Technical Specification



अखिल भारतीय आयुर्विज्ञान संस्थान, रायपुर (छ. ग.)
All India Institute of Medical Sciences, Raipur (Chhattisgarh)
 Tatibandh, GE Road, Raipur-492 099 (C.G.)
 Email: nephrology@aiimsraipur.edu.in
गुर्दा रोग विभाग
DEPARTMENT OF NEPHROLOGY



स्वामित्वप्रमाणपत्र
Proprietary Article Certificate

फाइल संख्या औरसंदर्भ File Number and Reference		
1	सामग्रीकाविवरण Description of article	① Blood Tubing Set ② Diysate
2	पूर्वानुमानितमात्रा / वार्षिकआवश्यकता Forecast of quantity/annual requirement	① Blood Tubing - 6000 pieces per year ② Diysate = 90 piece per year
3	उपरोक्तमात्रा हेतुअनुमानितमूल्य Approximate estimated value for above	
4	निर्माताकानाम एवं पता Maker's name and address	M/s Fresenius Medical Care
5	अधिकृतडीलर / स्टॉकिस्टकानाम Name(s) of authorised dealers/stockists	M/s Balaji Surgicals
6	<p>मैंपी ए सी के आधारपरउपरोक्त खरीदकोस्वीकारकरताहूँऔर यह प्रमाणित करताहूँकि:</p> <p>नोट- (बी), (सी-1) या (सी-2) में से केवल एक कोबनाए रखने के लिए टिककरें, जोभीतागूहोऔरदूसरोकोकाट दें।कृपया (ए) टिककरपुष्टिकरेंइसकेबिनापीएसप्रमाणपत्र अवैध होगा</p> <p>I approve the above purchase on PAC basis and certify that:-</p> <p>Note- Tick to retain only one out of (b), (c-1) or (c-2) whichever is applicable and cross out others. Please do confirm (a) by ticking it - without which PAC certificate will be invalid.</p>	
6 (a)	यह एकमात्र फर्महैजोइसमदकानिर्माण / संग्रहणकररहाहै। और This is the only firm who is manufacturing /stocking this item. AND	✓
6 (b)	किसीअन्य फर्म द्वारासमरूपनदनिर्मित / विक्रय नहीं कियाजाताहै, जिसकाउपयोगइसकेबदलो कियाजासकताहै।अथवा A similar article in not manufacturing/sold by any other firm, which could be used in lieu OR	✓
6 (c-1)	कोईअन्य मेक / ब्रांडनिम्नलिखितकारणों (जैसेओईएम / वारंटी के) के लिए उपयुक्त नहीं होगा।अथवा No other make/brand will be suitable for following tangible reasons (like OEM/warranty spares): OR	✓

6 (c)	<p>कोई अन्य मक / ब्रांड निम्नलिखित कारणों से उपयुक्त नहीं होगा (अगर पीएससीमिछले खरीदमेंभीदियागयाथा, तोकृपयाइसकेबाद से अधिकस्रोतोंका चत्ता लगाने के लिए प्रयास करें)तथा</p> <p>No other make/brand will be suitable for following intangible reasons (if PAC was also given in the last procurement cycle, please also bring out efforts made since then to locate more sources): OR</p>	✓

	<p>प्रस्ताव के लिए वित्त शाखा की सहमतिकसंदर्भ(कार्रवाई मेंडारऔरलेखाविभाग द्वारा की जायेगी)</p> <p>Reference of concurrence of finance wing to the proposal (Action will be taken by stores & Account Department)</p>	✓

<p>पिछलेतीनसालोंमेंइसमद की पीएससी खरीदकाइतिहासनीचेदियाजासकताहै (यदि कोई हो)History of PAC purchase of this Item for past three years may be given below (if any)</p>			
प्रदायकका नाम	_____		
Name of the Supplier	_____		
आदेश / निविदासंदर्भ औरदिनांक	आदेशितमात्रा	आदेशपरमूलदर	प्रतिफलप्रदर्शनरिपोर्टअगरकोईहो
Order/Tender reference & Date	Quantity Ordered	<input type="checkbox"/> Basic Rate on order (₹)	Adverse Performance Reported If any

अनुमोदनकरनेवालेप्राधिकारीकाहस्ताक्षर.....

दिनांक.....

डॉ. विनय राठौर
Dr. Vinay Rathore

अधिविशेषज्ञ (किडनी रोग विभाग)
Assistant Professor (Nephrology)
अखिल भारतीय आयुर्विज्ञान संस्थान, रायपुर (A.G.)
A. India Institute of Medical Sciences, Raipur (A.G.)



FRESENIUS MEDICAL CARE

INDIA

Proprietary Certificate

To whom it may concern:

We, Fresenius Medical Care AG & Co. KGaA, 61346 Bad Homburg, Germany hereby certify that below mentioned product is proprietary item of Fresenius Medical Care AG & Co. KGaA.

The below mentioned products are manufactured by and only for Fresenius Medical Care AG & Co. KGaA.

Sr.No.	Article No.	Items
01	F00000384	AV Set Online Plus 5008-R
02	F00001068	AV Set Online Plus BVM Pead 5008-R
03	5008201	Diasafe Plus
04	F00006672	Bibag 5008 650 gm
05	F00006673	Bibag 5008 900 gm
06	6321551	Filter Cartridge GAC5 9 3/4" 5-micron ACT. CAR.
07	6308801	Filter Cartridge GX10 9 3/4" 10 micron Aqua WTU
08	F00005158	Citrosteril
09	5085621	Puristril 340

M/s Fresenius Medical Care India Private Limited is an affiliated company and ultimately owned by Fresenius Medical Care AG & Co. KGaA.

Only M/s Fresenius Medical Care India Private Limited or any other authorized representative as appointed by M/s Fresenius Medical Care India Pvt Ltd. is authorized to bid, negotiate, sell and conclude the contract in regard to this business in India.

Date – 01/09/2021

For and on behalf of Fresenius Medical Care AG & Co. KGaA

DocuSigned by:
Mark Stanley
F855E09FF624A2...
Senior Vice President &
General Counsel
for Asia Pacific Region

DocuSigned by:
Chris Chuang
8C35C2C74C79438
Vice President
Quality & Regulatory Affairs
Asia Pacific

डॉ. विनय राठौर
Dr. Vinay Rathore
सहायक प्राध्यापक (किडनी रोग विभाग)
Assistant Professor (Nephrology)
अखिल भारतीय आयुर्विज्ञान संस्थान, रायपुर (छ.ग.)
All India Institute of Medical Sciences, Raipur (C.G.)

Page 1 / 1

Fresenius Medical Care AG & Co. KGaA, 61346 Bad Homburg, Germany, T +49 6172 609-0, Registered Office and Commercial Register: Hof an der Saale, HRB 4019, VAT-ID No.: DE 811127677, Chairman of Supervisory Board: Dr. Dieter Schenk
General Partner: Fresenius Medical Care Management AG, Registered Office and Commercial Register: Hof an der Saale, HRB 3894
Management Board: Rice Powell (Chairman), Helen Giza, Franklin W. Maddux MD, Dr. Katarzyna Mazur-Hofsäb, Dr. Olaf Schermeier,
William Valle, Kent Wanzek, Harry de Wit, Chairman of Supervisory Board: Stephan Sturm
Bank Account: Commerzbank AG, Frankfurt/Main, IBAN: DE23 5008 0000 0711 6731 00, SWIFT/BIC: DRESDEFF501



**FRESENIUS
MEDICAL CARE**

एमएस रायपुर

Manufacturer's Letter

Dated: 23-05-2022

To,
Director,
AIIMS Raipur.

Sub : Authorization Letter to supply consumables on PAC basis

Dear Sir,

This is to inform you that M/s Balaji Surgical is our authorized distributor for supply of dialysis machines and consumables manufactured by Fresenius Medical Care and/or its affiliates. The address and other contact details of M/s Balaji Surgical as per our records are as follows:

Contact Person Name : Mr. Vivek Tripathi

Address : M/s Balaji Surgical
Gol Chowk, Back side of Sai Mandir, Rohnipuram, Raipur, CG.


Contact Number : Mob: 9009323292

Email Address : balajisurgicalraipur@gmail.com


Note: This letter is valid till 31st March 2023 or till further communication, whichever is earlier.

Yours faithfully,

For and on behalf of Fresenius Medical Care India Pvt. Ltd.


Krishnendu Mahato
Asst. Manager Tender Management




डॉ. विनय राठौर
Dr. Vinay Rathore

सहायक प्राध्यापक (निदान रोग विभाग)
Assistant Professor (Medicine)
अखिल भारतीय आयुर्विज्ञान संस्थान, रायपुर (छ.ग.)
All India Institute of Medical Sciences, Raipur (C.G.)

Fresenius Medical Care India Private Limited

CIN: U24231DL2006FTC147436

Regd. Office : S-21, Second Floor, Star City, District Centre, Mayur Palace, Mayur Vihar Phase-1, New Delhi-110091 Web: www.freseniusmedicalcare.asia

Corporate Office Gurgaon :
Ground Floor, SAS Tower-A,
The Medicity, Sec-38,
Gurgaon-122001
Tel : 0124 - 6642500
Fax : 0124 - 6642506

Branch Office Mumbai :
Kanakia Wallstreet, B-207/208,
2nd Floor, Andheri - Kurla Road,
Andheri - 400093
Tel : 022 - 4217 6500
Telefax : 022 - 4217 6501
Telefax : 022 - 4217 6501

Branch Office Kolkata :
Unit No. T2-14A, 14th Floor, Millennium City IT Park,
Tower-II, Block-DN, Plot No.62, Salt Lake Sector V,
Bidhan Nagar, Kolkata-700091
Tel : 033 - 3090 9500
Telefax : 033 - 3090 9506
Telefax : 033 - 3090 9506

Branch Office Chennai :
2nd Floor (VTB Center), Old No-4,
New No-7, South Boag Road,
T. Nagar, Chennai - 600 017
Tel : 044 - 4396 8300
Fax : 044 - 4396 8300
Fax : 044 - 4396 8315

SPECIFICATION OF CONSUMABLE

1. Blood tubing (Compatible for FMC 5008 S machine)

1.1. Tubing should be made up of Medical grade PVC.

It should be Phthalate free, non DEHP, Non pyrogenic, Non toxic with Filling volume 132 to 169 ml.

1.2 It should be enabled to perform exact measurement of arterial pressure via inbuilt dome.

1.3. Trouble free air detection should be enabled directly on system tube.

1.4. Sterilization –beta sterilization by accelerated electrons (e-beam), fewer material alterations of comparison to gamma irradiation, no radioactive source of radiation.

1.5. Inner diameter of pump segment should be 8 mm, Diameter of venous bubble catcher 22 mm.

2. DIASAFE Plus filter


2.1. Dialysis fluid filters, material: Polysulfone: Extended surface area of at least 2.2 m²

2.2. Silicone sealing: with DIAFIX lock system.

2.3. Housing material: Polypropylene

2.4. Potting material: Polyurethane

2.5. Filtration rate of minimum: 5 mL/min x mm Hg (3.75 L/min bar; max. 2 bar)


डॉ. विनय राथोर
Dr. Vinay Rathore

सहायक प्राध्यापक (किडनी रोग विभाग)
Assistant Professor (Nephrology)
अखिल भारतीय आयुर्विज्ञान संस्थान, रायपुर (उ.प्र.)
All India Institute of Medical Sciences, Raipur (U.G.)