



अखिल भारतीय आयुर्विज्ञान संस्थान, रायपुर (छत्तीसगढ़)  
All India Institute of Medical Sciences, Raipur (Chhattisgarh)  
Tatibandh,  
Raipur-492 099 (CG)

No. AIIMS/R/HS/Surg\_Gast/2022/PAC/225

Date: 11/01/2023

विषय/ Sub: Inviting Quotations for procurement of "EEA Circular stapler with tristaple technology" for Department of Surgical Gastroenterology at AIIMS Raipur under Proprietary Article.

### **QUOTATION NOTICE**

Sealed quotations are invited from intending registered Stockiest/Distributors/Traders having GST Number and relevant documents for "EEA Circular stapler with tristaple technology" for Department of Surgical Gastroenterology at AIIMS Raipur under Proprietary Article. The quotation should be submitted to office of Stores Officer (Hospital), Room No. 146, C-C1 Block, Gate No 1, AIIMS Raipur up to 3:00 pm on **18/01/2023**. The quotations will be opened on the same day at 3:30pm. Details of item are given as under:-

Sr. No.	NAME	Qty.	UNIT	Make & HSN Code	UNIT Rate in Rs.	GST in %	Unit Rate with GST	TOTAL PRICE in RS.
1	EEA Circular Stapler for Medium thick tissue with tristaple technology diatmeter 31 mm	05	Nos.					
2	EEA Circular Stapler for Extra thick tissue with tristaple technology diatmeter 31 mm	03	Nos.					
<b>Specification of the product attached at Annexure-A</b>								

### **Terms & Conditions:-**

1. Rate should be mentioned in words & figures both.
2. **GST rates applicable** on your quoted item may please be confirmed.
3. Delivery Schedule- within 15 days from the date of issue of P.O.
4. No additional documents related to this NIQ will be entertained after NIQ opened.
5. Price should be F.O.R. for Destination basis (i.e. Concerned Department).
6. LD@0.5% of delayed supply per week or part week for delay of supply of material subject to maximum upto 10% of delayed supply should be deducted.
7. **Quotation No./Name and Due date of opening must be written on top of the envelop.**
8. **The GST registration details may please be furnished.**
9. Please confirm if there any change (Upward/Reduction) in your **Basic Price** structure and you are also requested to pass the Input Credit as per the following **Anti Profiteering Clause** of GST. **“Upon Implementation of GST, any reduction in the rate of tax on supply of goods or service or the benefit of input tax credit shall be passed on to AIIMS Raipur by way of commensurate reduction in the prices”**

10. Brand & Make should be clearly mentioned in offers
11. Validity of offer should not be less than 90 days.
12. Supply, Installation and Commissioning will be done by firm (if applicable).
13. The quantity shown in above column are totally tentative, it may be increase and decrease at the time of placement of purchase order.
14. Firm has to submit the latest Purchase Order copy of supplied of above items in other Government Institute with the quotation.

**Stores Officer (H)**  
**AIIMS Raipur (C.G)**

## Annexure-A

### Technical Specification:-

#### Specification for:

**A) EEA Circular Stapler for Medium thick Tissue with tristaple technology diameter Size 31 mm:**

Quantity: 5 units

1. Graduated Height 3 Rows Circular Stapler for End to End Anastomosis with Extra Security
2. For Medium thick Tissue
3. Having 3, 3.5 and 4 mm of staple heights (purple color)
4. 31 mm head diameter size
5. Should have 45 staple
6. Must have audible and tactile feedback during firing
7. It should have detachable Anvil Head Profile with Tilt-Top Anvil
8. Shaft size of 35cm
9. It has to be with integrated trocar including ready to Fire Green indicator
10. USFDA, CE, ISO certified

**B) EEA Circular Stapler for Extra thick Tissue with tristaple technology diameter Size 31 mm:**

Quantity: 3 units

1. Graduated Height 3 Rows Circular Stapler for End to End Anastomosis with Extra Security
2. For Extra thick Tissue
3. Having 4,4.5 and 5 mm of staple heights (black color)
4. 31 mm head diameter size
5. Should have 45 staple
6. Must have audible and tactile feedback during firing
7. It should have detachable Anvil Head Profile with Tilt-Top Anvil
8. Shaft size of 35cm
9. It has to be with integrated trocar including ready to Fire Green indicator
10. USFDA, CE, ISO certified

#### Terms and conditions:

1. All the items should be from the same manufacturing company
2. Item wise selection as L-1 shall not be done
3. The bidder who has not quoted for all the instruments i.e. the entire set shall not be entertained and bid shall be rejected

**Stores Officer (H)  
AIIMS Raipur (C.G)**