



No. AIIMS/R/CS/T & E/17/85/LPC

Dated: 20/06/2017

Notice for inviting Quotation for supply of Dual Chamber Temporary Pacemaker

QUOTATION NOTICE

Sealed quotations are invited from intending registered Stockiest/
Distributors having TIN/ relevant documents for supply of the following items to be
used in Department of Trauma & Emergency at AIIMS, Raipur and should be
submitted to Medical College Building, 2nd Floor, Gate no. 05 office of Store Officer
up to 3:00 pm on 28th June 2017, item description as per detailed bellow.

| S. N. | Description of Item | Qty. | Unit Price in ₹ | VAT/Tax in % | Unit Price Incl. Tax. | Total Price in ₹. |
|-------|--|----------|-----------------|--------------|-----------------------|-------------------|
| 1. | Dual Chamber Temporary Pacemaker: Specification : 1. Should have Knob facility for ease to operation. 2. Should have an additional internal battery back up to facilitate quick change of normal AA battery. 3. Should be able to be attached easily to patient arm, let or IV pole. 4. Should have rapid atrial pacing to manage atrial flutter safe reliable. 5. Should have safety awareness ranges to alert clinician to area of caution. 6. Should have modes AOO, AAI,AAT,VOO,VVI,VVT, DDD,DDI,DVI,DAI,DOO, DDT, VDD, AND VAT. 7. Should have basic pacing rate to 30 to 180ppm which is continuously adjustable. 8. Should have rapid atrial pacing rate 40 to 1000ppm. 9. Should have rapid atrial pacing rate: 40 to 120ppm in steps of 2ppm increments, 120 to 200ppm in steps of 5ppm increments, 200 to 400ppm in steps of 10ppm increments 400 to 1000ppm in steps of 50 increments. 10. Should have an output amplitude 0.1 to 20mv. 11. Should have Atrium pulse width 0.7ms Ventricle 0.5ms and high rate 1.0ms. 12. Should have sensitivity 0.1 to 20mv which is incrementally adjustable. 13. Should have refractory 400ms for AXX modes and 250ms for VXX modes. 14. Should have blanking pace 125ms and sense 75ms. 15. Should have work on 2,1.5 AA batteries which are very easily available in the market. 16. Should have runaway protection depending on rate settings. 17. Should have back up battery life > 150hrs. 18. Should have AV/PV delay 1.5ms to 300ms. Setting automatic depending on rate manual. 19. Should have PVARP 100 TO 500ms setting automatic depending on rate or normal. 20. Should have TARP AV Delay +PVARP with minimum rate + 60ppm. 21. Should have AV/VA blanking 70ms. | 01 No | | | | |

| | | | | | |
|--|--|--|--|--|--|
| 22. Should have UTR/MTR rate + 40ppm. 23. Should have noise detection at 125ms in automatic mode switch to asyndronous mode. 24. Pacing cable should be provided. 25. It should be CE approved. 26. Adequate service backup should be available within the state. 27. It should have warranty of 2 years. | | | | | |
|--|--|--|--|--|--|

OTHER COMMERCIAL TERMS

1. Rate should be mentioned in word and figures both.
2. Taxes, if any (Should be clearly mentioned).
3. Delivery Schedule – within 15 days from the date of issue of PO.
4. Price should be F.O.R. Destination basis (Trauma & Emergency Department).
5. LD @ 0.5% of delayed supply per week or part of week for delay of supply of material subject to maximum up to 10% of delayed supply should be deducted.
6. Quotation No/Name and Due date of opening must be written on top of the envelop.
7. VAT Reg. (i.e. From 11) must be attached with the quotation.
8. 100% payment against received and acceptance of material.
9. Brand and make should be clearly mentioned in offer.
10. AIIMS Raipur reserves the right to place the order for full or part quantity to one or more items.
11. RTGS detail required for payment purpose.
12. Validity of offer should not be less than 90 days.

**Stores Officer
AIIMS, Raipur (CG)**