



क्रमांक /No. AIIMS/R/CS/BB/18/04/LPC/A

दिनांक: 02/08/2018

Notice for inviting Quotation for Portable Autoclave for Transfusion Medicine & Blood Bank at AIIMS Raipur.

कोटेशन सुचना
QUOTATION NOTICE

Sealed quotations are invited from intending registered Stockiest/ Distributors having GST/ relevant documents for supply Portable Autoclave for Transfusion Medicine & Blood Bank at AIIMS, Raipur should be submitted to Medical College Building, 2nd Floor, Gate no. 05 office of Store Officer up to 3:00 pm on **08st August 2018**, item description as per detailed bellow.

स. क्र. S. N.	सामग्री का विवरण Description of Item	मात्रा Qty.	HSN Code	ब्रांड Make/ Brand/ Model	इकाई दर Unit Price in ₹	कर GST in %	कर सहित इकाई दर Unit Price Incl. GST
1.	Portable Autoclave विनिर्देश: (Specification): Detail technical specification as per Annexure-I	01 No					

नियम व शर्तें

Terms & Condition

1. Rate should be mentioned in word and figures both.
2. GST (Kindly mention in above table) should be clearly mentioned in the offer.
3. Delivery Schedule – within 15 days from the date of issue of PO.
4. Price should be F.O.R. Destination basis (Transfusion Medicine & Blood Bank).
5. LD @ 0.5% of delayed supply per week or part of week for delay of supply of material subject to maximum up to 10% of delayed supply should be deducted.
6. Quotation No/Name and Due date of opening must be written on top of the envelop.
7. 100% payment against received and acceptance and satisfactory installation of material.
8. Brand and make should be clearly mentioned in offer.
9. AIIMS Raipur reserves the right to place the order for full or part quantity to one or more items.

10. RTGS detail required for payment purpose.
11. Validity of offer should not be less than 90 days.
12. Supply, Installation and commissioning will be done by firm (if applicable).
13. Firm to submit documentary evidence in support of claim for GST at the time of submission of Bills.
14. The quantity shown in above column are totally tentative, it can be increase and decrease at the time of placement of order.
15. The bidder must quote for 1 year onsite warranty for Portable Autoclave as per the above mentioned Specification from the date of completion of the satisfactory installation as certified by the department.
16. The bidder should submit their quote for subsequent 1 year onsite AMC for Portable Autoclave, as per annexure-B but it should not be more than 4% of the quoted unit price.
17. Failure to comply this condition will entail the rejection of the quotation/bids. The Price comparison shall be made taking into account on basis price and post warranty AMC. The Rate Contracting Authority reserve the right to award AMC shall be quoted for equipment. So the price of AMC should be quoted according to the cost of equipment.
18. The amount of AMC would be released to the supplier on successful completion of the maintenance of that particular year duly certified by the department.
19. Quoting price for Portable Autoclave and AMC is mandatory. Lowest one will be deciding i.e. **L1 = Item price with 01 year warranty + 01 year AMC**
20. The 10% of total contract value may be with held as security deposit during the warranty period of assets (if applicable).

(सुशील सोनबेर)
भंडार अधिकारी
एम्स रायपुर (छ.ग.)

Portable Autoclave

1. Portable microprocessor-controlled fully automatic, rapid autoclave suitable for sterilization of unwrapped instrument, wrapped instruments, packs and special cycle for liquids.
2. Capacity of minimum of 100 litres.
3. Have 2 more programmable cycles to allow custom creation of different cycle parameters for special sterilization needs. The programmable cycle should allow change of Time, Temperature, Dry time and Vent.
4. Should have three or more Air removal purges, prior to sterilization.
5. Pre-programmed cycles should have selection of at least 5 programs.
6. Have a microprocessor-based fault detection circuit for monitoring all functions of sterilizer during a cycle, giving necessary signals to alert operator.
7. Should meet the requirements of ASME boiler and Pressure Vessel Code; Pressure Equipment Directive (PED) & EN 13445 norms of equivalent.
8. Should be BIS and CPCB standards of US FDA or European CE-class II A certified.
9. Electrical safety standards such as IEC class 1 or ISO standards.
10. Printer should be provided for recording cycle, cycle time, temperatures and pressure.
11. Should have Front filling and Drain facility behind front door, for safety. Should incorporate all safety features including fault detection circuit, continuous monitoring of chamber temperature to prevent overheat condition; audible alarms upon detection of a Fault with program.
12. Display should indicate cycle selected, cycle temp and exposure time for selected cycle.
13. During cycle, display should reflect messages indication status of cycle and errors.
14. During sterilization mode, LCD display should show—remaining cycle time, temperature and pressure.

(सुशील सोनबेर)
भंडार अधिकारी
एम्स रायपुर

**PRICE SCHEDULE FOR
ANNUAL MAINTENANCE CONTRACT (A.M.C.)**

(RATES SHOULD BE QUOTED IN INDIAN RUPEES ONLY)

Sl. No.	Name of the item	Qty.	For 1 year AMC after warranty (In Rs.)	GST @ %	Net Amount with GST	Total Amount with GST charges
(1)	(2)		(3)	(4)	(5)	(6)
1.	Portable Autoclave	01 No				

Note:-

1. Firm should quote the rate for AMC for equipment at relevant column and it should not be more than 4% per year of unit price of the quoted price.
2. Firm should upload this sheet after filling of relevant column as attachment of quotation.
3. In case of discrepancy between unit price and total prices, THE LOWER ONE shall prevail.
4. The cost of Annual Maintenance Contract (AMC) after satisfactory completion of Warranty period which include preventive maintenance including testing & calibration as per technical/ service/ operational, should be quoted along taxes applicable on the date of tender opening. The taxes to be paid extra, to be specifically stated. In the absence of any such stipulation the price will be taken inclusive of such taxes and no claim for the same will be entertained later.
5. AMC charges will be negotiable with respect to lower charges quoted for AMC.
6. The uptime warranty will be 98% on 24 (hrs) x 365 (days) basis or as stated in Technical Specification of the TE document.

Place:

Date

Signature of bidder with Seal