



क्र.:AIIMS/R/CS/Gen. Surgery/018/09-022/LPC/A

Dated:- 10/09/2018

कोटेशन सूचना
QUOTATION NOTICE

विषय / Sub:- Quotation inviting for Purchase of **Semi Rigid Uretero-Renoscope** for Department of General Surgery at AIIMS Raipur.

Sealed quotations are invited from intending registered Stockiest/ Distributors having GST number and relevant documents for Supply/Purchase of **Semi Rigid Uretero - Renoscope** for Department of General Surgery. The vendor should send the quotation with copy of certificate of GST & other documents should be submitted to office of Store Officer, AIIMS, Raipur, up to **03.00 pm** on dt: **17/09/2018**. The quotation will be opened on the same day at **03.30 pm** in the Store Office, Medical College Building, 2nd Floor, Gate No. 05, AIIMS-Raipur, Tatibandh GE Road, Raipur (C.G.). Details of item are as under:-

क्र. सं. S.N.	समग्री का विवरण Description of Items	एच.एन. एन. नं. HSN Code	मेक एवं ब्रांड Make / Brand	मात्रा Qty	ईकाई Unit	प्रति ईकाई दर Unit Rate in Rs.	कर Tax In GST@ %	कर के साथ ईकाई की दर Unit Rate with TAX	कुल मूल्य Total Price
01	Semi Rigid Uretero - Renoscope			01	Set				

Technical Specification

- 1- The scope should be having a viewing angle of 5°.
- 2- The scope should have caliber of 6/7.5 Fr or more.
- 3- The caliber of distal tip of sheath should be 6 Fr.
- 4- Working length should be 430 mm or more.
- 5- Total length of the instrument should be 600mm or more.
- 6- Should have incorporated Marberger type lateral offset eyepiece.
- 7- Should have incorporated fiber optics for light transmission.
- 8- Should have two lateral Irrigation ports with leur type connectors.
- 9- Should have main instrument channel of caliber at least 4.2 x 4.6 Fr.
- 10- Should have auxiliary instrument channel for use of 1x4 Fr or 2x2.4 Fr accessory instrument.
- 11- Should have automatic valve facility for inserting instruments.
- 12- The complete set should be fully autoclaveable.
- 13- Should be provided with the following accessories.
- 14- The quoted scope should be fully compatible and functional with the existing camera processor and light source/light cable (Richard wolf make) of the Department of General Surgery.
- 15- The instrument should have BIS/Euro. CE/USFDA approved certification.
- 16- Manufacture/supplier should have ISO certification.
- 17- Electrical safety standards such as IEC 60601-1 or equivalent BIS standard should be confirmed.
- 18- Should meet internationally recognized standard for electromagnetic compatibility (EMC) for electro medical equipment's such as 60601-1-2, IEC 60601-1-2 or equivalent BIS standards.
- 19- **Standard manufacture warranty against wear and tear for at least 1 year should be provided.**

नियम एवं शर्तें

Terms & Condition

1. Rate should be mentioned in words & figure both.
2. Taxes, if any (should be clearly mention).
3. Delivery Schedule – within 10 days from the date of issue of PO.
4. Price should be FOR Destination basis.(i.e. concerned department)
5. LD @ 0.5% of delayed supply per week or part of week for delay of supply of material subject to maximum up to 10%. After expiry of delivery period material cannot be accepted without extension of delivery period.
6. Quotation No/Name and Due date of opening must be written on top of envelop.

7. **GST** rates applicable on your quoted item may please be confirmed. **HSN** code for each item should be clearly mention.
8. Please confirm if there any change (Upward/Reduction) in your Basic Price structure. And you are also requested to pass the Input Credit as per the following Anti Profiteering Clause of GST. "Upon Implementation of GST, any reduction in the rate of tax on supply of goods or service or
9. The benefit of input tax credit shall be passed on to AIIMS Raipur by way of commensurate reduction in the prices.
10. In the event of increase in price, detailed justification and supporting evidence may be submitted for our consideration.
11. The GST registration details may please be furnished.
12. 100% payment against receipt and acceptance of material.
13. Validity of offer should not be less than 90 days
14. No Part supply or Part Payment will be entertained.
15. RTGS detail required for payment purpose.
16. Expenditure will be debitable to GIA-48.
17. Brand & Make should be clearly mentioned in offer (If require).
18. The Quantity of above column is totally tentative. It can be increased or decreased at the time of placement of order.
19. AIIMS Raipur reserve the right to place the order for full or part quantity to one or more items.

(सुशील सोनबेर)

भंडार अधिकारी

अखिल भारतीय आयुर्विज्ञान संस्थान

रायपुर (छ.ग.)-492099