

अखिल भारतीय आयुर्विज्ञान संस्थान, रायपुर(छत्तीसगढ़) All India Institute of Medical Sciences, Raipur (Chhattisgarh) Tatibandh, GE Road, Raipur-492 099 (CG) <u>www.aiimsraipur.edu.in</u>

AIIMS/R/CS /Dent/17/50/LPC/C

Date: 25.10.2018

## कोटेशन सूचना <u>QUOTATION NOTICE</u>

Inviting Quotations for Purchase of <u>Implant micromotor with physiodispenser and contra-</u> <u>angle surgical handpiece for</u> Department of Dentistry at, AIIMS Raipur.

Sealed quotations are invited from intending registered Stockist / Distributors having GST and relevant documents for Purchase of **Implant micromotor with physiodispenser and contra- angle surgical handpiece for** Department of Dentistry at, AIIMS Raipur. The quotation with copy of certificate of GST & other documents should be submitted to office of **Store Officer Medical College Building** up to 02/11//2018 before 3:00 pm. The quotations will be opened on the same day at 3:30pm. Details of item are given as under:-

क्र. सं.	सामग्री का विवरण	एचएसएन	मेक एवं	मात्रा	इकाई दर	टेक्स	कर के साथ	कुल मुल्य
S.No.	Description of	कोड	ब्रांड	Qty	रू. में	GST	दर	<b>Total Price</b>
	Items	HSN	Make /	_	Unit		Unit	
		Code	Brand		Rate in		Rate	
					Rs.		with	
							TAX	
1	<u>Implant</u>			01				
	<u>micromotor</u>							
	<u>with</u>							
	<u>physiodispenser</u>							
	and contra-							
	angle surgical							
	<b>handpiece</b>							

<u>Technical Specifications for Implant micromotor with physiodispenser and contra- angle</u> <u>surgical handpiece</u>

- Must be USFDA or European CE or Indian BIS Quality standard certified
- Central Control console unit with touch control adjustment for speed, torque, irrigation, direction, program
- Foot pedal unit for On/Off, direction, Irrigation controls
- Adjustable flow rate or irrigation by peristaltic pump with minimum 90 ml per minute flow rate
- Up to 75 NCM torque at the surgical handpiece
- Selectable reduction ratios
- Contra angle handpiece
- Audible alerts
- Full LCD display

- <70dB noise level
- Be supplied with 10 sets of sterile irrigation tubing
- All attachments and saline stands necessary to run the equipment
- Ready to plug and play to run out of the box
- AC input 220 -240 V
- With Transport and storage case
- Compatible with all implant surgical drills
- Have bur/drill internal cooling system as well as external spray for cooling
- Console, cord and handpiece must be fibreoptic for excellent illumination
- Speed variable from 5 ot 50000 RPM
- Sterilizable micromotor and cable
- Handpiece compatible with both external and internal drill irrigation
- Detachable handpiece which should be capable of full disassembly for maintence
- One can of handpiece lubrication spray

## OTHER COMMERCIAL TERMS

- 1. Rate should be mentioned in words & figure both.
- 2. Taxes, if any (should be clearly mention).
- 3. Delivery Schedule within 10 days from the date of issue of PO.
- 4. Price should be FOR Destination basis.(i.e. concerned department)
- 5. LD @ 0.5% of delayed supply per week or part of week for delay of supply of material subject to maximum up to 10%. After expiry of delivery period material cannot be accepted without extension of delivery period.
- 6. Quotation No/Name and Due date of opening must be written on top of envelop.
- 7. **GST** rates applicable on your quoted item may please be confirmed. **HSN** code for each item should be clearly mention.
- 8. Please confirm if there any change (Upward/Reduction) in your Basic Price structure. And you are also requested to pass the Input Credit as per the following Anti Profiteering Clause of GST. "Upon Implementation of GST, any reduction in the rate of tax on supply of goods or service or
- 9. the benefit of input tax credit shall be passed on to AIIMS Raipur by way of commensurate reduction in the prices.
- 10. In the event of increase in price, detailed justification and supporting evidence may be submitted for our consideration.
- 11. The GST registration details may please be furnished.
- 12. 100% payment against receipt and acceptance of material.
- 13. Validity of offer should not be less than 90 days
- 14. No Part supply or Part Payment will be entertained.
- 15. RTGS detail required for payment purpose.
- 16. Expenditure will be debitable to GIA-48.
- **17. Brand & Make should be clearly mentioned in offer as well as tender/quotation specific authorization may be submit with the offer/Bid.**
- 18. The Quantity of above column is totally tentative. It can be increased or decreased at the time of placement of order.
- **19.** AIIMS Raipur reserve the right to place the order for full or part quantity to one or more items.

**(सुशील सोनबेर)** भंडार अधिकारी अखिल भारतीय आयुर्विज्ञान संस्थान रायपुर (छ.ग.)