



क्र.:AIIMS/R/CS/Gen. Surgery/018/09-048/LPC/

Dated:- 29/11/2018

कोटेशन सूचना

QUOTATION NOTICE

विषय/Sub:- Quotation inviting for Purchase of **Laparoscopic Grasping Forceps** for Department of General Surgery at AIIMS Raipur.

Sealed quotations are invited from intending registered Stockiest/ Distributors having GST number and relevant documents for Supply/Purchase of **Laparoscopic Grasping Forceps** for Department of General Surgery. The vendor should send the quotation with copy of certificate of GST & other documents should be submitted to office of Store Officer, AIIMS, Raipur, up to **03.00 pm** on dt: **07/12/2018**. The quotation will be opened on the same day at **03.30 pm** in the Store Office, Medical College Building, 2nd Floor, Gate No. 05, AIIMS-Raipur, Tatibandh GE Road, Raipur (C.G.). Details of item are as under:-

क्र. सं. S.N.	समग्री का विवरण Description of Items	एच.एन. एन. नं. HSN Code	मेक एवं ब्रांड Make / Brand	मात्रा Qty	ईकाई Unit	प्रति ईकाई दर Unit Rate in Rs.	कर Tax In GST @ %	कर के साथ ईकाई की दर Unit Rate with TAX	कुल मूल्य Total Price
01	CROCE-OLMI Forceps			01	Nos				
02	Bowl Grasping Forceps			01	Nos.				
03	Kelly Type Forceps			01	Nos.				
04	Roddick OSLEN Forceps			01	Nos				

Specification: As per Annexure - I

नियम एवं शर्तें

Terms & Condition

1. Rate should be mentioned in words & figure both.
2. Taxes, if any (should be clearly mention).
3. Delivery Schedule – within 10 days from the date of issue of PO.
4. Price should be FOR Destination basis.(i.e. concerned department)
5. LD @ 0.5% of delayed supply per week or part of week for delay of supply of material subject to maximum up to 10%. After expiry of delivery period material cannot be accepted without extension of delivery period.
6. Quotation No/Name and Due date of opening must be written on top of envelop.
7. **GST** rates applicable on your quoted item may please be confirmed. **HSN** code for each item should be clearly mention.
8. Please confirm if there any change (Upward/Reduction) in your Basic Price structure. And you are also requested to pass the Input Credit as per the following Anti Profiteering Clause of GST. “Upon Implementation of GST, any reduction in the rate of tax on supply of goods or service or
9. The benefit of input tax credit shall be passed on to AIIMS Raipur by way of commensurate reduction in the prices.
10. In the event of increase in price, detailed justification and supporting evidence may be submitted for our consideration.
11. The GST registration details may please be furnished.
12. 100% payment against receipt and acceptance of material.
13. Validity of offer should not be less than 90 days
14. No Part supply or Part Payment will be entertained.
15. RTGS detail required for payment purpose.
16. Expenditure will be debitable to GIA-48.
17. Brand & Make should be clearly mentioned in offer (If require).

//2//

18. The Quantity of above column is totally tentative. It can be increased or decreased at the time of placement of order.
19. AIIMS Raipur reserve the right to place the order for full or part quantity to one or more items.

(सुशील सोनबेर)
भंडार अधिकारी
अखिल भारतीय आयुर्विज्ञान संस्थान
रायपुर (छ.ग.)-492099

Annexure- I

Technical specifications of laparoscopic grasping forceps

1. The grasping forceps should be of dismantling and rotating type with three individual parts: handle, outer sheath and jaw insert.
2. The grasping forceps outer sheath should have diameter of 5 mm.
3. The outer sheath should be fully insulated with leur lock connector and plug
4. A rotating knob should be present allowing 360 degree rotation of the jaw.
5. A flushing port with should be present for proper cleaning of the instrument.
6. The working length of the forceps should be 35 -36 cm.
7. The plastic handle of the grasping forceps should be without ratchet mechanism.
8. The handle should have a large contact area at the finger ring for easy operation.
9. The plastic handle as well as the instrument tips should be interchangeable.
10. The instrument assembly should have one touch release button for quick disassembly.
11. All the forceps must have a connector pin for use with monopolar electro-cautery and fully insulated outer tube.
12. The items should be reusable and fully autoclavable.
13. Jaw design of the grasping forceps
 - a. CROCE-OLMI type forceps having atraumatic, fenestrated, curved, single action jaw, jaw length of 27-30 mm - 01 in number
 - b. Bowel grasping forceps having atraumatic, fenestrated, curved double action jaws, jaw length of 18-20 mm - 01 in number
 - c. Kelly type Maryland forceps having non-fenestrated, curved double action jaws, jaw length of 15-16 mm - 01 in number
 - d. Reddick Oslen type forceps having non-fenestrated, straight double action jaws, jaw length of 13-15 mm - 01 in number
14. The instruments should have BIS/European CE and US FDA approval certification.
15. Manufacture/Supplier should have ISO certification.
16. Electrical safety standards such as IEC 60601-1 or equivalent BIS standard should be confirmed.
17. Should meet internationally recognized standard for electromagnetic compatibility (EMC) for electro medical equipment's such as 60601-1-2, IEC 60601-1-2 or equivalent BIS standards.
- 18. Standard manufacture warranty against wear and tear for at least 1 year should be provided.**
19. The manufacturer must have direct repair facility in India.