



क्रमांक /No. AIIMS/R/CS/URO/19/33/LPC

दिनांक: 01/04/2019

Notice for inviting Quotation for OT Trolley, Bowl Stand & Irrigation stand for Urology Department at AIIMS Raipur.

कोटेशन सुचना  
**QUOTATION NOTICE**

Sealed quotations are invited from intending registered Stockiest/ Distributors having GST/ relevant documents for supply of OT Trolley, Bowl Stand & Irrigation stand for Urology Department at AIIMS, Raipur should be submitted to Medical College Building, 2nd Floor, Gate no. 05 office of Store Officer up to 3:00 pm on **10<sup>th</sup> April 2019**, item description as per detailed bellow.

स. क्र. S. N.	सामग्री का विवरण Description of Item	मात्रा Qty.	HSN Code	ब्रांड Make/ Brand/ Model	इकाई दर Unit Price in ₹	कर GST in %	कर सहित इकाई दर Unit Price Incl. GST	कुल रकम Total Price in ₹.
1.	<b>Stainless Steel OT Instrument Trolley</b> (For Endourology & Laparoscopy) विनिर्देश: (Specification): Detail technical specification as per Annexure-I	02 Nos						
2.	<b>Stainless Steel Bowl Stand</b> विनिर्देश: (Specification): Detail technical specification as per Annexure-I	04 Nos						

स. क्र. S. N.	सामग्री का विवरण Description of Item	मात्रा Qty.	HSN Code	ब्रांड Make/ Brand/ Model	इकाई दर Unit Price in ₹	कर GST in %	कर सहित इकाई दर Unit Price Incl. GST	कुल रकम Total Price in ₹.
3.	<b>Stainless Steel Irrigation Bottle stand</b> (For Endourology) विनिर्देश: (Specification): Detail technical specification as per Annexure-I	04 Nos						

### नियम एवं शर्तें:

#### **Terms & Condition:**

1. Rate should be mentioned in word and figures both.
2. Taxes/GST, if any (Should be clearly mentioned).
3. Delivery Schedule – within 15 days from the date of issue of PO.
4. No any additional documents related to this NIQ will be entertained after opening of NIQ.
5. Price should be F.O.R. Destination basis (Urology Department).
6. LD @ 0.5% of delayed supply per week or part of week for delay of supply of material subject to maximum up to 10% of delayed supply should be deducted.
7. Quotation No/Name and Due date of opening must be written on top of the envelop.
8. 100% payment against received and acceptance of material.
9. No claim will be entertained regarding interest on any payment.
10. Where there is a statutory requirement for tax deduction at source, such deduction towards income tax and other tax as applicable will be made from the bills payable to the Supplier at rates as notified from time to time.
11. No payment shall be made for rejected stores. Rejected equipment's must be removed by the supplier within two weeks of the date of issue of rejection advice at their own cost & replace immediately. In case these are not removed these will be auctioned at the risk and responsibility of the suppliers without notice.
12. **Brand and make should be clearly mentioned in offer as well as tender/quotation specific authorization may be submit with the offer/Bid.**
13. GST rate applicable on your quoted item may please be confirmed. HSN code for each item should be clearly mentioned.
14. Please confirm if there any change (Upward/Reduction) in your Basic Price structure. And you are also requested to pass the Input Credit as per the following Anti Profiteering Clause of GST.  
**“Upon Implementation of GST, any reduction in the rate of tax on supply of goods or service or the benefit of input tax credit shall be passed on to AIIMS Raipur by way of commensurate reduction in the prices”.**
15. The GST registration details may please be furnished.

16. AIIMS Raipur reserves the right to place the order for full or part quantity to one or more items.
17. RTGS detail required for payment purpose.
18. No part supply of part payment will be entertained.
19. Validity of offer should not be less than 90 days.
20. Supply, installation and commissioning will be done by firm (if applicable).
21. The quantity shown in above requirement column are totally tentative, it can be increase and decrease at the time of placement of purchase order.

भंडार अधिकारी  
एम्स रायपुर (छ.ग.)

**Detailed of Technical Specifications of Urology Drape:**

**Specifications for Stainless Steel OT Instrument Trolley (For Endourology):**

- A. Frame work should be made of fully Stainless-Steel tubes mounted on 100 mm swivelling castors.
- B. Chrome plated pressed sheet frame.
- C. Frame work Stainless-Steel tubes should be made of atleast 1-inch diameter tubes & of atleast 18G thickness steel.
- D. Two stainless steel shelves with four side railings on both shelves. One shelf on top and one at bottom above casters.
- E. All stainless-steel parts should be matt finished.
- F. There shouldn't be any sharp edges.
- G. Material of trolley: Stainless Steel 304
- I. Material of Top Shelf & Bottom Shelf: 22G thick non-magnetic Stainless-Steel sheet bent atleast 2.5 cm downward all around.
- J. Railings at Top & Bottom Shelf: Made of 6 mm or more diameter non-magnetic Stainless-Steel pipe atleast 50 mm above the shelves. Sheet thickness of 6 mm diameter Stainless Steel pipe should be 1.5 mm or more.
- K. Length of Trolley: 120 cm
- L. Width of Trolley: 60 cm
- M. Height of Trolley: 75 cm
- N. Ground clearance: 16 cm
- O. 100 mm diameter rubber bonded four swivelling castors with brakes.
- P. Weight carrying capacity of the shelves (in kg): 10 Kg or more
- Q. Warranty: 2 years for manufacturing defects, rust & corrosion.



**Pic: 1 Stainless Steel OT Instrument Trolley**

### Specifications for Stainless Steel OT-Bowl Stand

- A. Frame work should be made of fully Stainless-Steel tubes mounted on three swivelling castors.
- B. Chrome plated pressed sheet frame.
- C. All stainless-steel parts should be matt finished.
- D. There shouldn't be any sharp edges.
- E. Material of trolley: Stainless Steel 304
- F. Height of Trolley: 75 cm
- G. 50 mm diameter rubber bonded three swivelling castors with brakes.
- H. Weight carrying capacity of the stand (in kg): atleast 6 Kg
- I. Warranty: 2 years for manufacturing defects, rust & corrosion.



**Pic: 2 Stainless Steel OT-Bowl Stand**

**Specifications for Irrigation Bottle Stand (For Endourology):**

- A. Frame work should be made of fully Stainless-Steel tubes mounted on base having swivelling castors.
- B. Five legs base mounted on 50 mm size swivelling castors.
- C. Casters should be rubber bonded with brakes.
- D. Telescopic tubes for height adjustments.
- E. Maximum Height: 2600 mm
- F. Minimum Height 1500 mm
- G. Fixed vertical base SS Tube diameter: 25 mm, made of 15G thickness steel.
- H. Telescopic SS Tube diameter: 16 mm, made of 18G thickness steel.
- I. No. of Hooks for hanging irrigation bottle: 4 Nos
- J. Hooks must not be pigtail type. It should be of hook type for easy mounting of 3Litres irrigation bottles.
- K. Weight bearing capacity of each hook: atleast 3 Kg
- L. All stainless-steel parts should be matt finished.
- M. There shouldn't be any sharp edges.
- N. Material of frame: Stainless Steel 304
- O. Warranty: 2 years for manufacturing defects, rust & corrosion.



**Pic: 3 Irrigation Bottle Stand (For Endourology)**