



अखिलभारतीयआयुर्विज्ञानसंस्थान, रायपुर(छत्तीसगढ़)  
**All India Institute of Medical Sciences, Raipur (Chhattisgarh)**  
Tatibandh, GE Road,  
Raipur-492 099 (CG)  
[www.aiimsraipur.edu.in](http://www.aiimsraipur.edu.in)

AIIMS/R/H.S/2018/9622/155

Date: 08-04-2018

**Inviting Quotations for Purchase of Urine collection beg from ECL for IPD and OPD services, in Hospital, at AIIMS Raipur.**

**कोटेशन सूचना**

**QUOTATION NOTICE**

Sealed quotations are invited from intending registered Stockist / Distributors having GST and relevant documents for Purchase of **Urine collection beg from ECL for IPD and OPD services, AIIMS Raipur**. The quotation with copy of certificate of GST & other documents should be submitted to office of **Store Officer Room No. – 51, 2nd floor Ayush Building, Gate No. 1 up to 11-04-2019 before 3:00 pm**. The quotations will be opened on the same day at 3:30pm. Details of item are given as under:-

**Hospital Requirement for IPD, OPD Services, AIIMS, Raipur**

Sr No क्र.	Name of Item समग्री विवरण	Quantity requir ed आवश्य क मात्रा	Unit इकाई	Brand/ Make निर्माण	Product HSN code प्रोडक्ट HSN कोड	Unit rate (INR) इकाई दर	Applicable GST (INR) लागु GST			Total Amo unt (INR) कुल दर
							CG ST	SGST	IGS T	
1	<b>Technical Specification of Urine Collection Bag:</b> 1. Single use urine collection bag compatible for fully closed urine drainage system to eliminate the risk of contamination. 2. Collection bag should be made of medical grade transparent Polyvinyl Chloride (PVC) or alike material for easy observation of urine. 3. The bag should have 2000 ml capacity with clear graduation marking at every 100 ml for easy measurement reading. 4. Top outlet with closure	4000	Nos							

<p>mechanism or bottom outlet with T-valve mechanism (leak proof) for discarding urine.</p> <p>5. Tapered universal inlet connector with stepped ridges for secure and leak-proof connection with Foley catheter. Connector should be with a cap inside the sterile package.</p> <p>6. <b>Wide bore (6 to 8 mm inner diameter bore), flexible and kink resistant transparent PVC inlet tubing</b> for unrestricted urine flow and milking of catheter if needed.</p> <p>7. Length of tubing should be at least <b>100 cm long</b>.</p> <p>8. Should have (needle/needle-free) urine sampling port facility for un-contaminated urine sampling.</p> <p>9. Efficient Anti-reflux valve to prevent backflow of urine from bag to inlet tubing.</p> <p>10. Modulated handle (T-type) for easy handling of bag and holds the inlet tubing upright to prevent kinking at junction of tubing with the bag.</p> <p>11. Bedside hanging mechanism: Double hook hanger or string with hook for secure and easy bedside hanging [No plastic rope which to be tied for hanging].</p> <p>12. Each unit must be individually packed in a sterile pouch.</p>										
---	--	--	--	--	--	--	--	--	--	--

### Terms & Condition

1. Firm to mention Make/Brand name in their quotation.
2. Taxes, if any (Kindly mention in above table) should be clearly mentioned in the offer.
3. Document relating to registration of firm i.e. GST number and relevant document should be submitted along with quotation.
4. Products are certified from ISI/FDA/CE0434/GMP certified as applicable.
5. Supply should be done within 15 days after Placement of PO.

6. Price should be FOR Destination basis (i.e. concerned department).
7. Payment will be released after certification from department of pharmacy.
8. **Quotation Name/No. and due date of opening must be mentioned on top of envelopes.**
9. LD @ 0.5% of delayed supply per week or part of week for delay of supply of material subject to maximum up to 10% of delayed supply to be deducted.
10. AIIMS Raipur reserves the right to place order for full or part quantity to one or more firms. The AIIMS, Raipur reserves the right to increase/decrease the number of required quantity All other terms & condition as per GCC applicable.
11. Material to be delivered at **Gate No-04, Hospital Pharmacy Store D-D1 block Lower ground floor.**
12. **Validity of the quotation should be 90 days from the date of opening.**
13. **Sample to be submitted as and when required.**
14. **Quotation should be in the given format.**
15. **Minimum 1 year expiry from the date of supply.**

**Rishi Gupta**  
**Stores Officer (H)**  
AIIMS, Raipur (C.G.)