



क्र./ AIIMS/R/CS /Patho/19/142/LPC/B

दिनांक: 30.04.2019

कोटेशन सूचना
QUOTATION NOTICE

Inviting Quotations for Purchase **Semi Automated Urine Strip Analyzer** for Department of Pathology & Lab Medicine, AIIMS Raipur.

Sealed quotations are invited from intending registered Stockist / Distributors having GST and relevant documents for Purchase of **Semi Automated Urine Strip Analyzer** for Department of Pathology & Lab Medicine, AIIMS Raipur. The quotation with copy of certificate of GST & other documents should be submitted to office of **Store Officer, 2nd floor, Medical College Building Gate No. 05, Tatibandh, G.E. Road, AIIMS, Raipur** up to 08/05/2019 before 3:00 pm. The quotations will be opened on the same day at 3:30pm. Details of item are given as under:-

क्र. सं. S. No.	सामग्री का विवरण Description of Items	एचएसएन कोड HSN Code	मेक एवं ब्रांड Make / Brand	मात्रा Qty	इकाई दर रु. में Unit Rate in Rs.	टेक्स GST@ %	कर के साथ दर Unit Rate with TAX	कुल मूल्य Total Price
1	Semi Automated Urine Strip Analyzer			01 set				

Technical Specification for Semi Automated Urine Strip Analyzer:

1. The unit should work of reflectance photometer to evaluate the color intensity of each test zone
2. The unit should have three wavelengths viz 470nm, 525nm and 625nm.
3. The unit should have incubation period of 60 seconds.
4. The unit should have throughput of around 600 samples/hr.
5. The unit should have strip position check system.
6. The unit should have concealed waste container for used strips disposal.
7. The unit should give results in SI. conventional and arbitrary units.
8. Company should provide 10 sets of two level ready to use liquid stable urine controls free of cost along with the instrument.
9. Company should provide tri-level control strip for instrument QC free of cost along with the machine.
10. Instrument should be touch screen with color LCD display.
11. The unit should have built in printing capability .
12. Instrument should have option to connect to bar code reader via USB port.

13. Unit should have memory capacity for 2000 measurements with all details like data by date and sample ID.
14. Unit should have external ports RS 232 and also 2 USB ports for LIS connection.
15. Power supply : 220-240 V; 50-60 HZ.
16. Should have European CE- USFDA certification or BIS approved.
17. The calibration, IQ, OQ and PQ of the instrument should be performed at the time of installation and certificates should be provided
18. UPS backup adequate for the duration of one cycle of processing or half an hour UPS backup should be provided.
19. Start up kit for at least 5000 test strips of 10 parameters should be provided free of cost along with the machine.
20. Document supporting track record and satisfactory performance form institutes of national importance (minimum one) should be provided.
21. Support for induction and follow up training of technical staff, on – site standardization and troubleshooting of procedures/ test to be provided by the company.
22. Three (03) years warranty should be provided.

OTHER COMMERCIAL TERMS

1. Rate should be mentioned in words & figure both.
2. Taxes, if any (should be clearly mention).
3. Delivery Schedule – within 10 days from the date of issue of PO.
4. Price should be FOR Destination basis.(i.e. concerned department)
5. LD @ 0.5% of delayed supply per week or part of week for delay of supply of material subject to maximum up to 10%. After expiry of delivery period material cannot be accepted without extension of delivery period.
6. Quotation No/Name and Due date of opening must be written on top of envelop.
7. **GST** rates applicable on your quoted item may please be confirmed. **HSN** code for each item should be clearly mention.
8. Please confirm if there any change (Upward/Reduction) in your Basic Price structure. And you are also requested to pass the Input Credit as per the following Anti Profiteering Clause of GST. “Upon Implementation of GST, any reduction in the rate of tax on supply of goods or service or
9. the benefit of input tax credit shall be passed on to AIIMS Raipur by way of commensurate reduction in the prices.
10. In the event of increase in price, detailed justification and supporting evidence may be submitted for our consideration.
11. The GST registration details may please be furnished.

12. 100% payment against receipt and acceptance of material.
13. Validity of offer should not be less than 90 days
14. No Part supply or Part Payment will be entertained.
15. RTGS detail required for payment purpose.
16. Expenditure will be debitable to GIA-48.
17. Brand & Make should be clearly mentioned in offer (If require).
18. The Quantity of above column is totally tentative. It can be increased or decreased at the time of placement of order.
19. AIIMS Raipur reserve the right to place the order for full or part quantity to one or more items.

भंडार अधिकारी
अखिल भारतीय आयुर्विज्ञान संस्थान,
रायपुर (छ.ग.)