

अखिल भारतीय आयुर्विज्ञान संस्थान, रायपुर (छत्तीसगढ़)

All India Institute of Medical Sciences, Raipur (Chhattisgarh)

Tatibandh, GE Road, Raipur-492 099 (CG) www.ailmsraipur.edu.in

화:AIIMS/R/CS/Gen. Surgery/018/09-053/LPC/B

Dated:-24/06/2019

कोटेशन सूचना QUOTATION NOTICE

Quotation inviting for Purchase of **Digital Intracorporeal Pneumatic Lithotriptor** for Department of General Surgery at AIIMS Raipur.

Sealed quotations are invited from intending registered Stockiest/ Distributors having GST number and relevant documents for Supply/Purchase of **Digital Intracorporeal Pneumatic Lithotriptor** for Department of General Surgery. The vendor should send the quotation with copy of certificate of GST & other documents should be submitted to office of Store Officer, AIIMS, Raipur, up to **03.00 pm** on dt: **01/07/2019**. The quotation will be opened on the same day at **03.30 pm** in the Store Office, Medical College Building, 2nd Floor, Gate No. 05, AIIMS-Raipur, Tatibandh GE Road, Raipur (C.G.). Details of item are as under:-

क्रं. सं. S.N.	समग्री का विवरण Description of Items	एच.एन. एन. नं. HSN Code	मेक एवं ब्रांड Make / Brand	मात्रा Qty	ईकाई Unit	प्रति ईकाई दर Unit Rate in Rs.	कर Tax In GST@ %	कर के साथ ईकाई की दर Unit Rate with TAX	कुल मूल्य Total Price
01	Digital Intracorporeal Pneumatic Lithotripter			01	Set				

- 1. Technical Specification as per Annexure-I.
- 2. Warranty: 05 Year onsite warranty from the date of supply of Material.

नियम एवं शर्ते

Terms & Condition

- 1. Rate should be mentioned in words & figure both.
- 2. Taxes, if any (should be clearly mention).
- 3. Delivery Schedule within 10 days from the date of issue of PO.
- 4. Price should be FOR Destination basis. (i.e. concerned department)
- 5. LD @ 0.5% of delayed supply per week or part of week for delay of supply of material subject to maximum up to 10%. After expiry of delivery period material cannot be accepted without extension of delivery period.
- 6. Quotation No/Name and Due date of opening must be written on top of envelop.
- 7. **GST** rates applicable on your quoted item may please be confirmed. **HSN** code for each item should be clearly mention.
- 8. Please confirm if there any change (Upward/Reduction) in your Basic Price structure. And you are also requested to pass the Input Credit as per the following Anti Profiteering Clause of GST. "Upon Implementation of GST, any reduction in the rate of tax on supply of goods or service or
- 9. The benefit of input tax credit shall be passed on to AIIMS Raipur by way of commensurate reduction in the prices.
- 10. In the event of increase in price, detailed justification and supporting evidence may be submitted for our consideration.
- 11. The GST registration details may please be furnished.
- 12. 100% payment against receipt and acceptance of material.
- 13. Validity of offer should not be less than 90 days
- 14. No Part supply or Part Payment will be entertained.
- 15. RTGS detail required for payment purpose.
- 16. Expenditure will be debitable to GIA-General
- 17. Brand & Make should be clearly mentioned in offer (If require).
- 18. The Quantity of above column is totally tentative. It can be increased or decreased at the time of placement of order.
- 19. AIIMS Raipur reserve the right to place the order for full or part quantity to one or more items.

भंडार अधिकारी अखिल भारतीय आयुर्विज्ञान संस्थान रायपुर (छ.ग.)—492099



Technical Specification of Digital Intracorporeal Pneumatic Lithotripter

1. Lithotriptor Control Unit

Should be made from rust free plastic body

Should be light weight having weight of less than 2 kilograms

Should have maximum dimension less than 25 cm in any one aspect

Should be compatible with Indian standard supply voltage of 230 VA/ 50Hz/A± 10 % or

110 VA C @60 Hz/A \pm 10 % Should have power consumptions of 30 watt.

Input Gas should be oxygen or compressed dry air

In let pressure range to control unit should be in the range of 4 - 6 Kg / cm2

Out let pressure indication should be 0 - 5 Kg / cm2 by digital display pressure gauge

Should have single or multiple mode operations

Frequency of impacts should be in the range of 0 pulse per second to 12 pulse per second

Should be accompanied by suitable and compatible air compressor and trolley

2. Hand Piece

Should be made from high quality non toxic alluminium alloy Should be light weight in the range of 100 gms \pm 10 gms

3. Probes for ureteric calculi

Length	Quantity
610 mm	05
610 mm	05
610 mm	05
	610 mm 610 mm

4. Probes for Kidney Stones

I LODGO IOI	Itianoj starre	
Size	Length	Quantity
2.5 mm	450 mm	05

5. Probes for Bladder Stones By Using Litho Bridge attachment

Size Length Quantity 1.5 mm 450 mm 05

- 6. The quoted model should be fully compatible and functional with the existing cystoscope/uretroscope (Richard wolf make) of the Department of General Surgery
- 7. The instrument should have BIS/European CE and US FDA approval certification
- 8. Manufacture/Supplier should have ISO certification
- 9. Electrical safety standards such as IEC 60601-1 or equivalent BIS standard should be confirmed
- 10. Should meet internationally recognized standard for electromagnetic compatibility (EMC) for electro medical equipments such as 60601-1-2,IEC 60601-1-2 or equivalent BIS standards
- 11. Standard manufacture warranty against wear and tear for at least 1 year should be provided