



अखिल भारतीय आयुर्विज्ञान संस्थान ,रायपुर) छत्तीसगढ़
(All India Institute of Medical Sciences, Raipur (Chhattisgarh)

Tatibandh, GE Road, Raipur-492 099 (CG)

www.aiimsraipur.edu.in

No. AIIMS/R/CS/409/BPS/19/LPC

Dated: 25/06/2019

Notice for inviting Quotation for supply of **Non Consumable items i.e. Fluid Warmer for Department of Burn & Plastic Surgery, AIIMS, Raipur.**

कोटेशन सूचना

QUOTATION NOTICE

Sealed quotations are invited from intending registered Stockiest/ Distributors having GST No./relevant documents for supply of Non Consumable item i.e. Fluid Warmer for Department of Burns & Plastic Surgery to be used in AIIMS, Raipur and should be submitted to Medical College Building, 2nd Floor, Gate no. 05 office of Store Officer up to 3:00 pm on 03.07.2019 The quotations will be opened on the same day at 3:30pm, item description is detailed bellow:

S.no	Item Detail	Qty. Req.	Make	Unit	HSN Code	Unit Price	GST @	Unit Price with Tax	Total Amount
1.	Fluid Warmer (Specification as - Annexure A)	01							

नियम और शर्त01:

Terms and Condition:

1. Rate should be mentioned in word and figures both.
2. Delivery Schedule – within 15 days from the date of issue of PO.
3. Price should be F.O.R. Destination basis (Department of Burn & Plastic Surgery, AIIMS Raipur).
4. LD @ 0.5% of delayed supply per week or part of week for delay of supply of material subject to maximum up to 10% of delayed supply should be deducted. After expiry of delivery period material cannot be accepted without extension of delivery period.
5. Quotation No./Name and Due date of opening must be written on top of envelop.
6. GST rates applicable on your quoted item may please be confirmed. HSN code should be clearly mentioned.
7. Please confirm if there any change (Upward/Reduction) in your Basic Price structure. And you are also requested to pass the Input Credit as per the following Anti Profiteering Clause of GST.
 - a. “Upon Implementation of GST, any reduction in the rate of tax on supply of goods or service or the benefit of input tax credit shall be passed on to AIIMS Raipur by way of commensurate reduction in the prices”.
8. The GST registration details may please be furnished.
9. RTGS details required for payment purpose.
10. 100% payment against received and acceptance of material.
11. No part supply or part payment will be entertained.
12. No claim will be entertained regarding interest on any payment of any circumstance.
13. Supply, installation and commissioning will be done by firm (if applicable).
14. Where there is statutory requirement for tax deduction at source, such deduction towards income tax and other tax as applicable will be made from the bills payable to the supplier at rates as notified from time to time.

15. No payment shall be made for rejected stores. Rejected equipment's must be removed by the supplier within two weeks of the dated of issue of rejection advice at their own cost & replace immediately. In case these are not removed these will auctioned at the risk and responsibility of the suppliers without notice.
16. Brand and Make should be clearly mentioned in offer as well as tender/quotation specific authorization may be submit with the offer/Bid.
17. AIIMS Raipur reserves the right to place the order for full or part quantity to one or more items.
18. Validity of offer should not be less than 90 days.
19. The Quantity of above column is totally tentative. It can be increased or decreased at the time of placement of order.

भंडार अधिकारी
अखिल भारतीय आयुर्विज्ञान संस्थान,
रायपुर (छ.ग.)

(Annexure –A)
Technical Specifications for Fluid Warmer

1. It should Deliver blood and intravenous fluid to the patient at normothermic temperature at wide range of flow rates from gravity flow rates to 5000 ml/hr
2. It should have dedicated disposable triple lumen tubing that eliminates patient line cool down of infusate
3. It should have Single step programming of warmer
4. It should Displays set point of recirculating reservoir
5. It should Meets AABB standards for blood warming
6. All the Disposables must be latex free
7. It should have built in safety audio and visual alarms for Check disposables, Add recirculating solution and Over Temperature
8. It should be mounted on IV pole.
9. It should have built in over temperature test button and alarm test button
10. The product should be “US-FDA& CE” / BIS and ISO Certified
11. Company should have Service Centre in India.
12. Warmer Should be IV stand mountable type model
13. Certificates to be enclosed
14. Product Safety should be EN 60601-1, UL 2601-1
15. EMC should be EN 60601-1-2, FCC 47 CFR Part 15, Class B
16. Enclosure Protection should be IEC 60529 IP Code: IPX1
17. Fluid Warmers should be ASTM F2172-02
18. Height, Overall should be 24.1cm(9.5 inches)+/- 2%
19. Width, Overall should be 21.0 cm (8.3 inches)+/- 2%
20. Depth, Overall should be 17.8 cm (7.0 inches)+/- 2%
21. Weight, Dry should be 3.5 Kg (7.6 lbs)+/-2%.
22. Weight, Wet (with recirculating solution) should be 5.0 Kg (11.0 lbs)+/- 2%
23. Weight, Shipping should be 3.6 Kg (7.95 lbs)+/- 2%
24. Recirculating Solution Capacity should be 1.4 L (0.37 gallons)+/- 2%
25. Maximum Height on I.V. Pole should be 107 cm (42 inches)+/- 2%
26. Operation should be temperature 10°C to 45°C and humidity 10 to 95
27. Transportation should be temperature -18°C to 60°C and humidity 5 to 90 %
28. Storage should be tem. 18°C to 60°C and humidity 5 to 90%
29. Temperature set point should be 41.9oC +/- 0.1oC
30. Over temperature set point should be 43.1oC
31. Protection Against Electrical Shock should be Class 1 Equipment, Type BF
32. Mode of Operation should be continuous
33. Type of current should be alternating
34. Ingress protection rating should be IPX1
35. Recirculating Solution temperature should be performance recirculating solution temperature reaches 37°C from ambient in about 4 minutes
36. Normothermic Flow Rates should be At gravity flow rates to 5,000 ml per hour
37. Warranty 01 Year.

Following Accessories Need To Be Supplied Per Unit Of Fluid Warmer:

38. 10 units of Triple Lumen Tubing.

