



क्रमांक /No. AIIMS/R/CS/URO/19/95/LPC

दिनांक: 26/07/2019

Notice for inviting Quotation for LED Spot Light for minor procedures for Urology Department at AIIMS Raipur.

कोटेशन सुचना
QUOTATION NOTICE

Sealed quotations are invited from intending registered Stockiest/ Distributors having GST/ relevant documents for supply of LED Spot Light for minor procedures for Urology Department at AIIMS, Raipur should be submitted to Medical College Building, 2nd Floor, Gate no. 05 office of Store Officer up to 3:00 pm on **07th August 2019**, item description as per detailed bellow.

स. क्र. S. N.	सामग्री का विवरण Description of Item	मात्रा Qty.	HSN Code	ब्रांड Make/ Brand/ Model	इकाई दर Unit Price in ₹	कर GST in %	कर सहित इकाई दर Unit Price Incl. GST	कुल रकम Total Price in ₹.
1.	LED Spot Light for minor procedures विनिर्देश: (Specification): Detail technical specification as per Annexure-I	02 Set						

नियम एवं शर्तें:

Terms & Condition:

1. Rate should be mentioned in word and figures both.
2. Taxes/GST, if any (Should be clearly mentioned).
3. Delivery Schedule – within 15 days from the date of issue of PO.
4. No any additional documents related to this NIQ will be entertained after opening of NIQ.
5. Price should be F.O.R. Destination basis (Urology Department).
6. LD @ 0.5% of delayed supply per week or part of week for delay of supply of material subject to maximum up to 10% of delayed supply should be deducted.
7. Quotation No/Name and Due date of opening must be written on top of the envelop.
8. 100% payment against received and acceptance of material.
9. No claim will be entertained regarding interest on any payment.

10. Where there is a statutory requirement for tax deduction at source, such deduction towards income tax and other tax as applicable will be made from the bills payable to the Supplier at rates as notified from time to time.
11. No payment shall be made for rejected stores. Rejected equipment's must be removed by the supplier within two weeks of the date of issue of rejection advice at their own cost & replace immediately. In case these are not removed these will be auctioned at the risk and responsibility of the suppliers without notice.
12. **Brand and make should be clearly mentioned in offer as well as tender/quotation specific authorization may be submit with the offer/Bid.**
13. GST rate applicable on your quoted item may please be confirmed. HSN code for each item should be clearly mentioned.
14. Please confirm if there any change (Upward/Reduction) in your Basic Price structure. And you are also requested to pass the Input Credit as per the following Anti Profiteering Clause of GST.
"Upon Implementation of GST, any reduction in the rate of tax on supply of goods or service or the benefit of input tax credit shall be passed on to AIIMS Raipur by way of commensurate reduction in the prices".
15. The GST registration details may please be furnished.
16. AIIMS Raipur reserves the right to place the order for full or part quantity to one or more items.
17. RTGS detail required for payment purpose.
18. No part supply of part payment will be entertained.
19. Validity of offer should not be less than 90 days.
20. Supply, installation and commissioning will be done by firm (if applicable).
21. The quantity shown in above requirement column are totally tentative, it can be increase and decrease at the time of placement of purchase order.

भंडार अधिकारी
एम्स रायपुर (छ.ग.)

Technical Specification of LED Spot Light for Minor Procedures:

1. Central illumination intensity at a distance of 1 meter: **minimum 30000 Lux**
2. Central illumination intensity at a distance of 0.5 meter: **minimum 70000 Lux**
3. Light field diameter at a distance of 1 meter: **minimum 165mm**
4. Light field diameter at a distance of 0.5 meter: minimum 110mm
5. Color Temperature: **minimum 4500 K**
6. Color Rendering index (Ra) in every color temperature setting: 90 or more
7. LED Life time: minimum 50000 hours
8. Should be mounted on stable mobile stand with at least five wheels.
9. Stand should have hinged arm on which the light is mounted.
10. Light should be height adjustable and angle adjustable.
11. **Class II** Medical Electrical Equipment Protection (IEC 60601-1)
12. Light Housing Ingress Protection Class (IEC 60601-1 & IEC 60601-2-2): **IP20 or better**
13. **System should have BIS/ US FDA/ European CE with four-digit no. certificate for the intended use.**
14. **Warranty: One Year**

भंडार अधिकारी
एम्स रायपुर (छ.ग.)