



अखिल भारतीय आयुर्विज्ञान संस्थान ,रायपुर) छत्तीसगढ़
(All India Institute of Medical Sciences, Raipur (Chhattisgarh)

Tatibandh, GE Road, Raipur-492 099 (CG)

www.aiimsraipur.edu.in

No. AIIMS/R/CS/281/BPS/18/LPC/

Dated: 30/10/2019

Notice for inviting Quotation for supply of **Non Consumable items i.e. DVT/VTE Pump with Battery for Department of Burn & Plastic Surgery, AIIMS, Raipur.**

कोटेशन सूचना

QUOTATION NOTICE

Sealed quotations are invited from intending registered Stockiest/ Distributors having GST No./relevant documents for supply of Non Consumable item i.e. **DVT/VTE Pump with Battery** for Department of Burns & Plastic Surgery to be used in AIIMS, Raipur and should be submitted to Medical College Building, 2nd Floor, Gate no. 05 office of Store Officer up to 3:00 pm on **06.11.2019** The quotations will be opened on the same day at 3:30pm, item description is detailed below:

S.no.	Item Detail	Qty. Req.	Make	Unit	HSN Code	Unit Price	GST @	Unit Price with Tax	Total Amount
1.	DVT/VTE Pump with Battery								
	Description of Equipment :- 1. Vena flow with battery version – S mode 2. Tube Assembly 3. Power cord 4. DVT Elite Calf Cuff (Specification as -Annexure A)	01 01 01 10							

नियम और शर्त01:

Terms and Condition:

1. Rate should be mentioned in word and figures both.
2. Delivery Schedule – within 15 days from the date of issue of PO.
3. Price should be F.O.R. Destination basis (Department of Burn & Plastic Surgery, AIIMS Raipur).
4. LD @ 0.5% of delayed supply per week or part of week for delay of supply of material subject to maximum up to 10% of delayed supply should be deducted. After expiry of delivery period material cannot be accepted without extension of delivery period.
5. Quotation No./Name and Due date of opening must be written on top of envelop.
6. GST rates applicable on your quoted item may please be confirmed. HSN code should be clearly mentioned.

8. Please confirm if there any change (Upward/Reduction) in your Basic Price structure. And you are also requested to pass the Input Credit as per the following Anti Profiteering Clause of GST.
 - a. **“Upon Implementation of GST, any reduction in the rate of tax on supply of goods or service or the benefit of input tax credit shall be passed on to AIIMS Raipur by way of commensurate reduction in the prices”.**
9. The GST registration details may please be furnished.
10. RTGS details required for payment purpose.
11. 100% payment against received and acceptance of material.
12. No part supply or part payment will be entertained.
13. **No claim will be entertained regarding interest on any payment of any circumstance.**
14. Supply, installation and commissioning will be done by firm (if applicable).
15. **Where there is statutory requirement for tax deduction at source, such deduction towards income tax and other tax as applicable will be made from the bills payable to the supplier at rates as notified from time to time.**
16. **No payment shall be made for rejected stores. Rejected equipment’s must be removed by the supplier within two weeks of the dated of issue of rejection advice at their own cost & replace immediately. In case these are not removed these will auctioned at the risk and responsibility of the suppliers without notice.**
17. **Brand and Make should be clearly mentioned in offer as well as tender/quotation specific authorization may be submit with the offer/Bid.**
18. AIIMS Raipur reserves the right to place the order for full or part quantity to one or more items.
19. Validity of offer should not be less than 90 days.
20. The Quantity of above column is totally tentative. It can be increased or decreased at the time of placement of order.

Sr.Administrative Officer
AIIMS Raipur

(Annexure –A)
Technical Specifications for DVT/VTE Pump (With Battery)

- 1) The device must be able to provide compression patterns on the lower limbs, mimicking normal ambulation and reduce DVT/VTE.
- 2) Must provide choices between rapid inflation and slow inflation technologies.
- 3) Must provide pre-set gradient compressions of 73-63-45 mmHg for thigh & calf sleeves (*and uniform compressions of 130-45 mmHg for foot sleeves.*)
- 4) With rapid inflation, the compression of 73-63 mmHg must be rapid lasting only for 6 seconds.
- 5) Must be automatically identifying and adjusting the pressures based on the sleeves attached, prior to turning the pump on.
- 6) Must have a digital display with audible and visible alarm prompts for faults and low battery power.
- 7) Must have a compliance monitor (timer) reset facility
- 8) Must be supplied with inbuilt battery back-up
- 9) Must have an telescoping hook/hanger to hand the device to bed/trolley sides
- 10) Must be able to operate even on single leg if required
- 11) Must be easy to use, portable and weight must not exceed 2.50 kgms
- 12) Power input: 220V-240V, 50-60 Hz, 200-500 mA
- 13) It should be “USFDA & European CE” / BIS approved product Certificates to be enclosed.
- 14) Standards: IEC 60601-1, IEC 60601-1-2
- 15) TUV device classification must be Class 1BF
- 16) Sleeves must be made of brushed nylon & polyester and must be free of latex
- 17) Must have following sleeves: Calf cuff.

Scope of Supply: For every 1 unit of DVT pump, 10 units of Calf Cuff have to be supplied.