



क्र./ AIIMS/R/CS /Pharma/18/013/LPC/

दिनांक: 08.11.2019

कोटेशन सूचना
QUOTATION NOTICE


Inviting Quotations for Purchase of instrument for clinical and PG research laboratories for Department of Pharmacology, AIIMS Raipur.

Sealed quotations are invited from intending registered Stockist / Distributors having GST and relevant documents for Purchase of instrument for clinical and PG research laboratories for Department of Pharmacology, AIIMS Raipur. The quotation with copy of certificate of GST & other documents should be submitted to office of **Store Officer, 2nd floor, Medical College Building Gate No. 05, Tatibandh, G.E. Road, AIIMS, Raipur** up to 18/11/2019 before 3:00 pm. The quotations will be opened on the same day at 3:30pm. Details of item are given as under:-

क्र. सं. S.N o.	सामग्री का विवरण Description of Items	एचएसएन कोड HSN Code	मेक एवं ब्रांड Make / Brand	मात्रा Qty	इकाई दर रु. में Unit Rate in Rs.	टेक्स GST@ %	कर के साथ दर Unit Rate with TAX	कुल मूल्य Total Price
1	Digital Memory Drum Specification as per Annexure-I			01 Nos				
2	Human Maze Learning Electronics with Digital reset error counter and Timer Specification as per Annexure-I			01 Nos				
3	Flicker Fusion Apparatus Specification as per Annexure-I			01 Nos				

Sr. No.	Equipment	Detail	Quantity
1.	Digital Memory Drum	<p>Purpose: It can be used in most of the Learning, Memory and Forgetting Experiments like Serial Verbal Learning, Retro-active Inhibition.</p> <p>Specification:</p> <ol style="list-style-type: none"> 1. Should be micro-computer based and should have memory storage of at least 200 words 2. Should have normal keyboard attached for easy operation. 3. Instrument should have arrangement to change or edit words 4. Should have button for changing display of word. 5. Instrument should be able to display score and error score. 6. Each stimulus can be exposed to the subject at the desired speed like 1, 2, 4, 8 with the Exposure Time 1, 2, 4, 8 Sec. respectively 7. Minimum 3-year warranty after installation. 8. Operating manual to be provided. 9. Instrument should be ISO certified. The product should be ISI marked. 10. Manufacturer should provide warranty of 3 years. 	1

Sample Image:



Sr. No.	Equipment	Detail	Quantity
2	Human Maze Learning Electronics with Digital reset error counter and Timer	<p>Purpose: To analyse psychomotor activity in humans</p> <p>Specification:</p> <ol style="list-style-type: none"> 1. Instrument should micro-computer based. 1. Stylus in maze should have enough length of cable so that subject can move stylus comfortably over maze 2. Should have Digital Timer for recording Time in seconds and its decimal parts. 3. Should have Digital Error Counter which could automatically records Errors in Digital Form. 4. Time and Error both should be recorded automatically. 5. It should have facility to modify the maze path as per the need of experimenter or atleast 5 – 6 different maze paths shall be provided. 6. Reset button should be in instrument to reset Digital Timer and Digital Error Counter 7. Should able to count and display total time and contact time in seconds digitally 8. Also should display number of contacts to the maze path digitally. 9. The Instrument should be ISO certified. The product should be ISI marked. 10. Should have 3 years warranty. 11. Operating manual should be provided. 	1

Sample
Image:



Sr. No.	Equipment	Detail	Quantity
3	Flicker Fusion Apparatus	<p>Purpose: To measure the Flicker-Fusion Threshold of a human subject</p> <p>Specification:</p> <ol style="list-style-type: none"> 1. Instrument to be micro-computer based 2. The apparatus should have two working sides for experimenter and subject. 3. Experimenter unit will stable and subject will be movable with attached enough length of wires for convenience of subject. 4. The Experimenter unit should consist of – <ol style="list-style-type: none"> a. Light colour and intensity selection switch. b. A switch to select increasing or decreasing rate of flicker of a light of variable Intensity. b. Flicker rate stimulus: 5 to 50 minimum flashes per seconds c. LCD display of frequency/flicker rate d. Flicker value to be displayed in 3 digits. 5. The Subject's unit should consist of switch for increasing or decreasing frequency. 6. Operating Manual should be provided 7. Minimum 3 years warranty 8. The Instrument should be ISO certified/or the product should be ISI marked. 	1

Sample Image:



OTHER COMMERCIAL TERMS

1. Rate should be mentioned in words & figure both.
2. Taxes, if any (should be clearly mention).
3. Delivery Schedule – within 10 days from the date of issue of PO.
4. Price should be FOR Destination basis.(i.e. concerned department)
5. LD @ 0.5% of delayed supply per week or part of week for delay of supply of material subject to maximum up to 10%. After expiry of delivery period material cannot be accepted without extension of delivery period.
6. Quotation No/Name and Due date of opening must be written on top of envelop.
7. **GST** rates applicable on your quoted item may please be confirmed. **HSN** code for each item should be clearly mention.
8. Please confirm if there any change (Upward/Reduction) in your Basic Price structure. And you are also requested to pass the Input Credit as per the following Anti Profiteering Clause of GST. “Upon Implementation of GST, any reduction in the rate of tax on supply of goods or service.
9. The benefit of input tax credit shall be passed on to AIIMS Raipur by way of commensurate reduction in the prices.
10. The GST registration details may please be furnished.
11. 100% payment against receipt and acceptance of material.
12. Validity of offer should not be less than 90 days
13. No Part supply or Part Payment will be entertained.
14. RTGS detail required for payment purpose.
15. Expenditure will be debitable to GIA-48.
16. Brand & Make should be clearly mentioned in offer (If require).
17. The Quantity of above column is totally tentative. It can be increased or decreased at the time of placement of order.
18. AIIMS Raipur reserve the right to place the order for full or part quantity to one or more items.
19. Warranty mention on specification.

Sr. Administrative Officer
All India Institute of Medical Sciences,
Raipur