



AIIMS/R/CS /Patho/19/171/LPC/

Date: 29.11.2019

**QUOTATION NOTICE**

Inviting Quotations for Purchase of **General Incubator** for use in emergency lab in Department of Pathology & Lab Medicine, AIIMS Raipur.

Sealed quotations are invited from intending registered Stockist / Distributors having GST and relevant documents for Purchase of **General Incubator** for use in emergency lab in Department of Pathology & Lab Medicine,, AIIMS Raipur. The quotation with copy of certificate of GST & other documents should be submitted to office of **Store Officer, 2<sup>nd</sup> floor, Medical College Building Gate No. 05, Tatibandh, G.E. Road, AIIMS, Raipur** up to 09/12/2019 before 3:00 pm. The quotations will be opened on the same day at 3:30pm. Details of item are given as under:-

| क्र. सं.<br>S.No. | सामग्री का विवरण<br>Description of<br>Items                                     | एचएसए<br>न कोड<br>HSN<br>Code | मेक एवं ब्रांड<br>Make / Brand | मात्रा<br>Qty | इकाई<br>दर रु.<br>में<br>Unit<br>Rate<br>in Rs. | टेक्स<br>GST | कर के साथ<br>दर<br>Unit Rate<br>with TAX | कुल मूल्य<br>Total<br>Price |
|-------------------|---|-------------------------------|--------------------------------|---------------|---|--------------|--|-----------------------------|
| 1                 | General Incubator<br><b>Technical<br/>Specification as<br/>per given below-</b> |                               |                                | 01<br>Set     |   |              |  |                             |

**Technical Specification for Incubator:**

- Inner Chamber should be made up of Stainless Steel and Outer abrasion resistance and durable finish by heat cured epoxy coating on steel exterior.
- Should have capacity around 400 Liters or better.
- Temperature range should be ambient+5°C to 70°C.
- Control Accuracy should be  $\pm 0.2^\circ\text{C}$  or better.
- Uniformity should be  $\pm 1^\circ\text{C}$  or better
- Should have eye level door mounted Controller for easy access and mounting.
- Should have long life glass sheathed heaters.
- Should have excellent control accuracy with the PID temperature Controller.
- Should have timer up-to 9999 mins.
- Should have over temperature Safety.
- Should have electric overload safety with MCB.
- Should have crevice free frameless tempered glass door.
- Should have rounded corners in stainless steel chamber.
- Should have high level of uniformity due to the gentle internal circulation.
- Should have rust proof and efficient stainless steel wire mesh shelves.
- Should have silicon gasket in door.

- Shelves height should be adjustable in steps of 25 mm.
- 5 KVA stabilizer with high and low voltage protection should be provided.
- Oven should be CE certified or BIS approved
- The calibration, IQ, OQ and PQ of the instrument should be performed at the time of installation and certificates should be provided.
- Support for induction and follow up training of technical staff, on-site standardization and troubleshooting of procedures/ tests to be provided by the company.
- Three (03) years **warranty** should be provided

#### **OTHER COMMERCIAL TERMS**

1. Rate should be mentioned in words & figure both.
2. Taxes, if any (should be clearly mention).
3. Delivery Schedule – within 10 days from the date of issue of PO.
4. Price should be FOR Destination basis.(i.e. concerned department)
5. LD @ 0.5% of delayed supply per week or part of week for delay of supply of material subject to maximum up to 10%. After expiry of delivery period material cannot be accepted without extension of delivery period.
6. Quotation No/Name and Due date of opening must be written on top of envelop.
7. **GST** rates applicable on your quoted item may please be confirmed. **HSN** code for each item should be clearly mention.
8. Please confirm if there any change (Upward/Reduction) in your Basic Price structure. And you are also requested to pass the Input Credit as per the following Anti Profiteering Clause of GST.  
“Upon Implementation of GST, any reduction in the rate of tax on supply of goods or service or
9. The benefit of input tax credit shall be passed on to AIIMS Raipur by way of commensurate reduction in the prices.
10. The GST registration details may please be furnished.
11. 100% payment against receipt and acceptance of material.
12. Validity of offer should not be less than 90 days
13. No Part supply or Part Payment will be entertained.
14. RTGS detail required for payment purpose.
15. Expenditure will be debitable to GIA-48.
16. Brand & Make should be clearly mentioned in offer (If require).
17. The Quantity of above column is totally tentative. It can be increased or decreased at the time of placement of order.
18. AIIMS Raipur reserve the right to place the order for full or part quantity to one or more items.

**Sr. Administrative Officer  
AIIMS, Raipur (C.G.)**