



अखिल भारतीय आयुर्विज्ञान संस्थान, रायपुर(छत्तीसगढ़)
All India Institute of Medical Sciences, Raipur (Chhattisgarh)
Tatibandh, GE Road,
Raipur-492 099 (CG)
www.aiimsraipur.edu.in

AIIMS/Raipur/HS/General/2019-20/155/09-081

Date: 27-12-2019

Inviting Quotations for procurement of Laparoscopic Tacking Device for Department of General Surgery at AIIMS, Raipur.

QUOTATION NOTICE

Sealed quotations are invited from intending registered Stockist / Distributors having GST and relevant documents for procurement of Laparoscopic Tacking Device for Department of General Surgery at AIIMS, Raipur. The quotation with copy of certificate of GST & other documents should be submitted to office of **Store Officer, Room No. – 51, 2nd floor, Ayush Building, Gate No. 1, AIIMS, Raipur up to 02-01-2020 before 03:00 pm**. The quotations will be opened on the same day at 03:30 pm. Details of item are given as under:-

Sl. No.	Name of Item	Quantity required	Unit	HSN Code	Brand /Make	Unit rate (INR)	GST Applicable			Amount (INR)
1	Laparoscopic Tacking Device Specification : Attached at Annexure-1	10	Nos							
Grand Total										

Terms & Condition :

1. Firm to mention Make/Brand name in their quotation.
2. Taxes, if any (Kindly mention in above table) should be clearly mentioned in the offer.
3. Document relating to registration of firm i.e. GST and relevant document should be submitted along with quotation.
4. Products are certified from US FDA/European CE (Four digit notification number) / BIS as applicable, the Certificate to this effect should be attached.
5. Supply should be done within 15 days after Placement of PO.
6. Price should be FOR Destination basis (i.e. concerned department).
7. 100% Payment will be released after certification from concerned department.
8. **Quotation Name/No. and due date of opening must be mentioned on top of envelopes.**
9. LD @ 0.5% of delayed supply per week or part of week for delay of supply of material subject to maximum up to 10% of delayed supply to be deducted.
10. AIIMS Raipur reserves the right to place order for full or part quantity to one or more firms. The AIIMS, Raipur reserves the right to increase/decrease the number of required quantity.

11. All other terms & condition as per GCC applicable.
12. Material to be delivered at **Department of General Surgery at AIIMS, Raipur.**
13. **Validity of the quotation should be 90 days from the date of opening.**
14. **Firm to submit documentary evidence in support of their claim for GST at the time of submission of bills.**
15. **Firm to submit sample if required.**

Pro. (Dr.) Karan Peepre
Medical Superintendent
AIIMS, Raipur (CG)

Annexure-1

Specification of Laparoscope tacking device

1. The tacker device should have an ergonomic pistol grip handle.
2. There should be a ratchet mechanism to provide tactile feed back during the firing of the tacks.
3. The shaft length of the device should be 36 cm or more.
4. The shaft diameter should be 5 mm for introduction through 5 mm trocars.
5. The tacker device should contain at least 30 tacks.
6. The device should not have any sharp pointed tip following delivery of all the tacks.
7. Absorbable tacks:
 - a. Should be made up of absorbable polyglycolic acid and lactic acid.
 - b. The design should have smooth head with proximal wings.
 - c. Dimension should be less than 6 mm in total length and less than 4.5 mm in inserted length.
 - d. Should be amenable to unscrewing in case of necessity
8. The device should have BIS/US FDA/European CE (four digit notification number) product certification.