



No.: AIIMS/R/CS/PACU/Capital/51001/19/

Date: 25/01/2020

**QUOTATION NOTICE**

Inviting quotations for purchase of “**Patient Convective Warming Device**” for PACU (MOT) 4<sup>th</sup> Floor A1 Block at AIIMS Raipur.

Sealed quotations are invited from intending registered Stockiest / Distributors having GST and relevant documents for above said. The quotation should be submitted to, Medical College Building, 2<sup>nd</sup> Floor, gate no.-5, AIIMS, Raipur office of the Stores Officer up to **31/01/2020** before **3:00 pm**. The quotations will be opened on the same day at **3.30 pm**.

**Terms & Condition:**

1. Quotation should be in **two- bid system** i.e. **Technical Bid** (as per **Annexure-I**) and **Financial Bid** (as per **Annexure-II**).
2. The technical bid and the financial bid should be sealed by the bidder in separate covers which will be super-scribed “**Technical Bid for Patient Convective Warming Device**” and “**Financial Bid for Patient Convective Warming Device**”. Both Sealed Envelopes should be kept in a main/ bigger envelope super-scribed as “**Quotation for supply of Patient Convective Warming Device**”.
3. **All the required documents should be enclosed in the envelope of technical bid and financial bid should be enclosed in the envelope of financial bid.**
4. Rate should be mentioned in words & figure both.
5. Taxes/GST, if any (Kindly mention in table)
6. Delivery schedule: within 15 days from the date of issue of P.O.
7. No additional documents related to this NIQ will be entertained after opening of NIQ.
8. Price should be FOR Destination basis (i.e. PACU (MOT) 4<sup>th</sup> Floor A1 Block, AIIMS Raipur).
9. LD @ 0.5% of delayed supply per week or part thereof for delay of supply of material subject to maximum up to 10% should be deducted. After expiry of delivery period material cannot be accepted without extension of delivery period.
10. Quotation No/Name and due date of opening must be written on top of envelop.
11. **Brand, Make & Warranty should be clearly mentioned in offer as well as tender/ quotation. Specific authorization may be submitted with the offer/bid.**
12. GST rates applicable on your quoted item may please be confirmed. **HSN code for each item should be clearly mentioned.**
13. Any change (Upward/Reduction) in the Basic Price structure post submission of bid shall be promptly intimated to AIIMS Raipur. You are also required to pass the Input Credit as per the following **Anti Profiteering Clause of GST**.  
“**Upon Implementation of GST, any reduction in the rate of tax on supply of goods or service or the benefit of input tax credit shall be passed on to AIIMS Raipur by way of commensurate reduction in the prices**”.
14. The GST registration details may be furnished.
15. RTGS details required for payment purpose.
16. 100% payment against receipt and acceptance of material.
17. No part supply or part payment will be entertained.
18. Validity of offer should not be less than 90 days.
19. Supply, installation and commissioning will be done by firm (if applicable).
20. The quantity in above column is totally tentative. It can be increased or decreased at the time of placement of order.
21. AIIMS, Raipur reserve the right to place the order for full or part quantity to one or more firms.

**Sr. Administrative Officer,  
AIIMS, Raipur (C.G.)**



ANNEXURE-I

Ref. No. & Date: -  
Quotation Notice No.: -  
Brand/Make Name:-  
Model No.:-  
Warranty-

**TECHNICAL BID**

Purchase of “**Patient Convective Warming Device**” for PACU (MOT) 4<sup>th</sup> Floor A1 Block at AIIMS Raipur. Details of item are given as under:

S. N.	Description of item(s)	Qty./ Unit	Offered Specification	Specify Technically Complied / Not Complied/ if any Deviations
1.	<b>Patient Convective warming Device :</b> 1. The warmer should use a forced air technology or convective warming technology. 2. Should have 4 set temperatures with ambient setting: 34 <sup>0</sup> C, 40 <sup>0</sup> C, 45 <sup>0</sup> C, & 47 <sup>0</sup> C. 3. Should have hose and nozzle clip & sheet Clip for easy use of blankets and reduce drag of hose pipe. 4. Should have low maintenance requirements. 5. Should have short warm up time of less than 40 seconds. 6. Should have accuracy of +/- 1 <sup>0</sup> C, supported by authenticated paper. 7. Can be mounted to the infusion pole or bed rails or placed on the floor or cart Mount facility. 8. Should have temperature in LCD Display. 9. Should have automatically step – down facility from high setting to medium setting after 45 mins. 10. Should have filter expiry display for easy maintenance of machine and filter life should be at least 2000 hrs. 11. Should have corrugated hose pipe. 12. Must meet all convective warming standards. 13. Should be easy to operate in two steps. 14. Should have safety alarm like over temperature, power disconnection. 15. Airflow of approx 23L / sec for better and fresh warming. 16. Should be light weight less than 6 kg.	02 Nos.		



अखिल भारतीय आयुर्विज्ञान संस्थान, रायपुर (छत्तीसगढ़)  
All India Institute of Medical Sciences, Raipur (Chhattisgarh)

Tatibandh, GE Road, Raipur-492 099 (CG)

[www.aiimsraipur.edu.in](http://www.aiimsraipur.edu.in)

17. Should deliver warm air uniformly to the patient through disposable blankets.			
18. Should have blankets options like Full body, Upper and Lower body, Pediatric & sterile blankets for surgical access.			
19. Blankets must have strong 2Ply layer to resist tear, puncture and fluid with comfortable draping on patient.			
20. Should have USFDA and CE certificate.			

Sign of bidder :-  
Date :-  
Name of the bidder :-  
Seal of the bidder :-



ANNEXURE-II

Ref. No. & Date: -

Quotation Notice No.: -

**FINANCIAL BID**

Purchase of “Patient Convective Warming Device” for PACU (MOT) 4<sup>th</sup> Floor A1 Block at AIIMS Raipur. Details of item are given as under:

S. N.	Description of item(s)	Qty./ Unit	HSN Code	Make/ Brand/ Model	Unit Price in ₹.	GST Per Unit @	Total Unit Price with Tax in ₹.	Total Amount in ₹.
1.	<b>Patient Convective warming Device :</b> 1. The warmer should use a forced air technology or convective warming technology. 2. Should have 4 set temperatures with ambient setting: 34 <sup>0</sup> C, 40 <sup>0</sup> C, 45 <sup>0</sup> C, & 47 <sup>0</sup> C. 3. Should have hose and nozzle clip & sheet Clip for easy use of blankets and reduce drag of hose pipe. 4. Should have low maintenance requirements. 5. Should have short warm up time of less than 40 seconds. 6. Should have accuracy of +/- 1 <sup>0</sup> C, supported by authenticated paper. 7. Can be mounted to the infusion pole or bed rails or placed on the floor or cart Mount facility. 8. Should have temperature in LCD Display. 9.Should have automatically step – down facility from high setting to medium setting after 45 mins. 10. Should have filter expiry display for easy maintenance of machine and filter life should be at least 2000 hrs. 11. Should have corrugated hose pipe. 12. Must meet all convective warming standards. 13. Should be easy to operate in two steps. 14. Should have safety alarm like over temperature, power disconnection. 15. Airflow of approx 23L / sec for better and fresh warming. 16. Should be light weight less than 6 kg. 17. Should deliver warm air uniformly to the patient through disposable blankets. 18. Should have blankets options like Full body ,Upper and Lower body , Pediatric & sterile blankets for	02 Nos.						



**अखिल भारतीय आयुर्विज्ञान संस्थान, रायपुर (छत्तीसगढ़)**  
**All India Institute of Medical Sciences, Raipur (Chhattisgarh)**

Tatibandh, GE Road, Raipur-492 099 (CG)

[www.aiimsraipur.edu.in](http://www.aiimsraipur.edu.in)

surgical access. 19. Blankets must have strong 2Ply layer to resist tear, puncture and fluid with comfortable draping on patient. Should have USFDA and CE certificate.							
--	--	--	--	--	--	--	--

Sign of bidder :-

Date :-

Name of the bidder :-

Seal of the bidder :-