



अखिलभारतीयआयुर्विज्ञानसंस्थान, रायपुर(छत्तीसगढ़)
All India Institute of Medical Sciences, Raipur (Chhattisgarh)
Tatibandh, GE Road,
Raipur-492 099 (CG)
www.aiimsraipur.edu.in

AIIMS/R/HS/2019-20/OW/15519/155/

Date: 06/02/2020

Inviting Quotations for Purchase of X-Ray view box (4-screen) for OPD/IPD services at AIIMS Raipur.

QUOTATION NOTICE

Sealed quotations are invited from intending registered Stockist / Distributors having GST and relevant documents for Purchase of X-Ray view box (4-screen) for OPD/IPD services at AIIMS Raipur. The quotation with copy of certificate of GST & other documents should be submitted to office of **Store Officer Room No. – 51, 2nd floor AYUSH Building, Gate No. 1 up to 08/02/20 before 12:00 pm**. The quotations will be opened on the same day at 12:30pm. Details of item are given as under:-

S. No.	Item description	Qty. reqd.	Unit	HSN Code	Brand	Unit Rate	GST Applicable			Total Amount (Inclusive of All taxes)
							CG ST	SG ST	IG ST	
1	<u>X-ray View Box (4-Screen)</u> <u>Specification:-</u> <ul style="list-style-type: none">• Slim, Durable and effective• Energy efficient: Savings range from 82% to 93%• No UV emission• Unparallel latest LED backlight technology.• 4-5 times longer than regular CCFL/EEFL• Power efficient: LED light with long life – up to 100,000 hours• Reliable magnetic nipper Initial brightness can be set• Film activated switch, automatic shutdown if no film inside for 2 sec• Dimensions Frame 12 step digital dimmer to get the perfect brightness• (L x H x T) 1631 x 503 x 29 MM• It should have a thickness of 29mm.• It should have an digital feather touch on-off switch along with adjustment range of approximately 90%.• 8-10 Step digital dimmer facility with step	05	Nos.							

up/step down intensity of 400 LUX. • Automatic film sensor. • For 2/3/4 Film viewers, facility to switch on only the section where the film needs to be viewed. • Screen must have 8 LED strip lights. • Power consumption 60W • Light source LED • Power supply Input AC 100 – 240 V 50/60 Hz • Brightness (LUX) > 10,000 • Net weight (KGS) 12.4 • Viewing Area 1488 x 440 • ISO & CE certified LED X ray viewers. • The stockist/distributor must have authorization certificate from company.									
Grand Total									

Terms & Condition

1. Firm to mention Make/Brand name in their quotation.
2. Taxes, if any (Kindly mention in above table) should be clearly mentioned in the offer.
3. Document relating to registration of firm i.e. GST and relevant document should be submitted along with quotation.
4. Products are certified from ISO & CE approved as applicable, the Certificate to this effect should be attached.
5. The stockist/distributor must have authorization certificate from company.
6. Supply should be done within 15 days after Placement of PO.
7. Price should be FOR Destination basis (i.e. concerned department).
8. 100% Payment will be released after certification from Department of Pharmacy.
9. **Quotation Name/No. and due date of opening must be mentioned on top of envelops.**
10. LD @ 0.5% of delayed supply per week or part of week for delay of supply of material subject to maximum up to 10% of delayed supply to be deducted.
11. AIIMS Raipur reserves the right to place order for full or part quantity to one or more firms. The AIIMS, Raipur reserves the right to increase/decrease the number of required quantity.
12. All other terms & condition as per GFR-2017 applicable.
13. Material to be delivered at **Hospital Store, AIIMS Raipur(CG).**
14. **Validity of the quotation should be 90 days from the date of opening.**

Stores Officer (H)
AIIMS, Raipur (C.G)