# अखिल भारतीय आयुर्विज्ञान संस्थान, रायपुर (छतीसगढ़)



#### All India Institute of Medical Sciences, Raipur (Chhattisgarh)

Tatibandh, GE Road, Raipur-492 099 (CG) www.aiimsraipur.edu.in

क्रं:AIIMS/R/CS/Gen. Surg/19/08-17/LPC

Dated: - 23/03/2020

#### **QUOTATION NOTICE**

<u>विषय/Sub:-</u> Quotation Inviting for Purchase of **Reusable Laparoscopic Bipolar Forceps with Curved Jaws** for Department of General Surgery at AIIMS Raipur.

Sealed quotations are invited from intending registered Stockiest/ Distributors having GST number and relevant documents for Supply/Purchase of **Reusable Laparoscopic Bipolar Forceps** with Curved Jaws for Department of General Surgery. The vendor should send the quotation with copy of certificate of GST & other documents should be submitted to office of Store Officer, AIIMS, Raipur, up to 03.00 pm on dt: 30/04/2020. The quotation will be opened on the same day at 03.30 pm in the Store Office, Medical College Building, 2<sup>nd</sup> Floor, Gate No. 05, AIIMS-Raipur, Tatibandh GE Road, Raipur (C.G.). Details of item are as under:-

# नियम एवं शर्ते

#### **Terms & Condition**

- 1. Rate should be mentioned in words & figure both.
- 2. Quotation should be in **two- bid system** i.e. **Technical Bid** (as per **Annexure-I**) and **Financial Bid** (as per **Annexure-I**).
- 3. The technical bid and the financial bid should be sealed by the bidder in separate covers which will be super-scribed **Technical Bid for printing and supply of Reusable Laparoscopic Bipolar Forceps with Curved Jaws "Spandan**" and **Financial Bid for printing**". Both Sealed Envelopes should be kept in a main/ bigger envelope super-scribed as **Quotation for printing and supply of Hindi Magazine** "Spandan".
- 4. All the required documents should be enclosed in the envelope of technical bid and financial bid should be enclosed in the envelope of financial bid.
- 5. Rate should be mentioned in words & figure both.
- 6. Taxes/GST, if any (Kindly mention in above table)
- 7. Delivery Schedule within 14 days from the date of issue of PO.
- 8. Price should be FOR Destination basis.(i.e. concerned department)
- 9. LD @ 0.5% of delayed supply per week or part of week for delay of supply of material subject to maximum up to 10%. After expiry of delivery period material cannot be accepted without extension of delivery period.
- 10. Quotation No/Name and Due date of opening must be written on top of envelop.
- 11. **GST** rates applicable on your quoted item may please be confirmed. **HSN** code for each item should be clearly mention.
- 12. Please confirm if there any change (Upward/Reduction) in your Basic Price structure. And you are also requested to pass the Input Credit as per the following Anti Profiteering Clause of GST. "Upon Implementation of GST, any reduction in the rate of tax on supply of goods or service or
- 13. The benefit of input tax credit shall be passed on to AIIMS Raipur by way of commensurate reduction in the prices.
- 14. In the event of increase in price, detailed justification and supporting evidence may be submitted for our consideration.
- 15. The GST registration details may please be furnished.
- 16. 100% payment against receipt and acceptance of material.
- 17. Validity of offer should not be less than 90 days
- 18. No Part supply or Part Payment will be entertained.
- 19. RTGS detail required for payment purpose.
- 20. Expenditure will be debitable to GIA-48 (General)
- 21. Brand & Make should be clearly mentioned in offer (If require).
- 22. The Quantity of above column is totally tentative. It can be increased or decreased at the time of placement of order.
- 23. AIIMS Raipur reserve the right to place the order for full or part quantity to one or more items.

वरि.प्रशासनिक अधिकारी

अखिल भारतीय आयुर्विज्ञान संस्थान रायपुर (छ.ग.)—492099

#### **ANNEXURE-I**

Ref. No. & Date: Quotation Notice No.: -

# **TECHNICAL BID**

Printing and Supply of Reusable Laparoscopic Bipolar Forceps with Curved Jaws for General Surgery Department at AIIMS, Raipur. Details of item are given as under:

S.	Description of item(s)	Qty./	Offered Specification	Specify Technically
N.		Unit		Complied / Not
				Complied/ if any
				Deviations
1.	Reusable Laparoscopic Bipolar Forceps with	01 Set		
	Curved Jaws.			
	Specification as per Annexure -III			

Sign of bidder :Date :Name of the bidder :Seal of the bidder :-



# अखिल भारतीय आयुर्विज्ञान संस्थान, रायपुर (छत्तीसगढ़)

# All India Institute of Medical Sciences, Raipur (Chhattisgarh)

Tatibandh, GE Road, Raipur-492 099 (CG)

www.aiimsraipur.edu.in

#### **ANNEXURE-II**

Ref. No. & Date: -Quotation Notice No.: -

### **FINANCIAL BID**

Printing and Supply of Reusable Laparoscopic Bipolar Forceps with Curved Jaws for General Surgery Department at AIIMS, Raipur. Details of item are given as under:

S. N	Description of item(s)	Qty./ Unit	HSN Code	Unit Price in ₹	GST Per Unit @	Total Amount in ₹.
1.	Reusable Laparoscopic Bipolar Forceps with Curved Jaws.	01 Set				
	i. Specification as per Annexure -III					

Sign of bidder :Date :Name of the bidder :-

Seal of the bidder :-

# TECHNICAL SPECIFICATIONS FOR REUSABLE BIPOLAR FORCEPS IN LAPAROSCOPIC SURGERY

- 1. Bipolar Maryland forceps for dissection and grasping: 01 set
  - a) The diameter of shaft of the instrument should be 5mm for use through 5.5 mm ports.
  - b) The working shaft length should be 350mm or longer.
  - c) The instrument should have double action NON-STICKY jaws.
  - d) The blade/jaw should be of minimum of 18mm length.
  - e) The bipolar instrument handle should be detachable having pistol grip but without rachet mechanism
  - f) Connector-pin in handle should be available for attachment of bipolar high frequency cable
  - g) The instrument should have rotational shaft tube with insulated sheath
  - h) The instrument should have fully autoclavable parts sustaining a maximum temperature of equal to or more than 138 degree centigrade
  - i) The detachable parts should be amenable to cleaning in standard disinfecting solutions
  - j) A bipolar high frequency cable with length of 3.5 meters or longer compatible with Olympus/Erbe high frequency generator unit should be provided
  - k) The instrument should offer Class IIb as well as EN 60 601-1 protection
  - 1) The complete instrument should be approved for clinical use by BIS/US FDA/ European CE (4 digit notification number to be mentioned)
  - m) Standard manufacturer warranty of at least 01 year should be provided against wear and tear of the equipment.