



No.: AIIMS/R/CS/Patho/20/23/LPC/

Date: 22.05.2020

QUOTATION NOTICE

Inviting quotations for purchase of “**Laboratory Vertical Autoclave**” for Department of Pathology & Lab Medicine at AIIMS Raipur.

Sealed quotations are invited from intending registered Stockiest / Distributors having GST and relevant documents for above said. The quotation should be submitted to, Medical College Building, 2nd Floor, gate no.-5, AIIMS, Raipur office of the Stores Officer up to **29.05.2020** before **3:00 pm**. The quotations will be opened on the same day at **3.30 pm**.

Terms & Condition:

1. Quotation should be in **two- bid system** i.e. **Technical Bid** (as per **Annexure-I**) and **Financial Bid** (as per **Annexure-II**).
2. The technical bid and the financial bid should be sealed by the bidder in separate covers which will be super-scribed “**Technical Bid for Laboratory Vertical Autoclave**” and “**Financial Bid for Laboratory Vertical Autoclave**”. Both Sealed Envelopes should be kept in a main/ bigger envelope super-scribed as “**Quotation for supply of Laboratory Vertical Autoclave**”.
3. **All the required documents should be enclosed in the envelope of technical bid and financial bid should be enclosed in the envelope of financial bid.**
4. Rate should be mentioned in words & figure both.
5. Taxes/GST, if any (Kindly mention in above table)
6. Delivery schedule: within 15 days from the date of issue of P.O.
7. No additional documents related to this NIQ will be entertained after opening of NIQ.
8. Price should be FOR Destination basis (i.e. Department of Pathology, AIIMS Raipur).
9. LD @ 0.5% of delayed supply per week or part thereof for delay of supply of material subject to maximum up to 10% should be deducted. After expiry of delivery period material cannot be accepted without extension of delivery period.
10. Quotation No/Name and due date of opening must be written on top of envelop.
11. **Brand, Make & Warranty should be clearly mentioned in offer as well as tender/ quotation. Specific authorization may be submitted with the offer/bid.**
12. GST rates applicable on your quoted item may please be confirmed. **HSN code for each item should be clearly mentioned.**
13. Any change (Upward/Reduction) in the Basic Price structure post submission of bid shall be promptly intimated to AIIMS Raipur. You are also required to pass the Input Credit as per the following **Anti Profiteering Clause of GST.**
“Upon Implementation of GST, any reduction in the rate of tax on supply of goods or service or the benefit of input tax credit shall be passed on to AIIMS Raipur by way of commensurate reduction in the prices”.
14. The GST registration details may be furnished.
15. RTGS details required for payment purpose.
16. 100% payment against receipt and acceptance of material.
17. No part supply or part payment will be entertained.
18. Validity of offer should not be less than 90 days.
19. Supply, installation and commissioning will be done by firm (if applicable).
20. The quantity in above column is totally tentative. It can be increased or decreased at the time of placement of order.
21. AIIMS, Raipur reserve the right to place the order for full or part quantity to one or more firms.
22. Due of COVID-19 soft copy of quotation notice is acceptable on the company letterhead with seal and signature in storesofficer.cp@aiimsraipur.edu.in mail id.
23. The product should be of superior quality and highly standard.

Stores Officer,
AIIMS, Raipur (C.G.)



ANNEXURE-I

Ref. No. & Date: -

Quotation Notice No.: -

Brand/Make Name:-

Model No.:-

Warranty-

TECHNICAL BID

Purchase of “**Laboratory Vertical Autoclave**” for Department of Pathology & Lab Medicine at AIIMS Raipur. Details of item are given as under:

S. N.	Description of item(s)	Qty./Unit	Offered Specification	Specify Technically Complied/Not Complied / if any Deviations
1.	The Vertical Autoclave should have Chamber Capacity of 70 or more liter.			
2.	The chamber of Vertical Autoclave should be manufactured as per ASME standards.			
3.	The Vertical Autoclave should work on the domestic power supply of : 230 V AC, 50 HZ, Single phase.			
4.	The unit's internal chamber, cover lid and all wetted parts should be fabricated from stainless steel of 304 grade.			
5.	The Vertical Autoclave's all joints should be smooth finished for crevice free internals.			
6.	The outer body should be made up of MS sheet with heat cured epoxy coating on both sides.			
7.	The lid should be equipped with single lever lock mechanism and lever handle moulded from industrial plastic.			
8.	The lid should be provided with auto purge cum vacuum breaker valve and a manually operable valve for exhaust.			
9.	The unit should have a solenoid valve for auto purging of air & normal exhaust.			
10.	The Vertical Autoclave should have stainless steel pressure gauge with dual range dial display in KPA and PSI along with a co-related temperature scale for steam in degrees Celsius.			
11.	The operations of the unit should be controlled by a microprocessor based controller.			
12.	The user should be able to set the temperature up to 122°C.			
13.	The autoclave should have timer facility.			
14.	In case of abrupt power failure / switching off, the Last Cycle which was set should remain in the memory.			
15.	The display for the parameters should be Two lines Alpha-Numeric digital display.			
16.	The unit should be equipped with Low Water Detection unit and should give Audio- Visual alarm in case of Low Water in the chamber and cut off the supply to the heater.			
17.	The unit should have safety valve to protect the equipment in case of over pressurization.			
18.	The Lid should be equipped with pressure interlock device. Also the heater should not start if the Lid is open.			
19.	The unit should be provided with safety cut-out for high temperature.			
20.	The unit should give indication by audio-visual alarm on completion of set autoclave cycle.			
21.	The electrical safety should be ensured by inbuilt MCB.			
22.	The unit should be mounted on 04 Nos. PU coated castors out of which at least 2 should have locking mechanism			
23.	Warranty of 03 years should be provided.			
24.	After sale, service should be available promptly.			
25.	Should be CE or FDA or BIS approved product.			
26.	Support for induction and follow up training of technical staff, on-site standardization and troubleshooting of procedures/ tests to be provided by the company.			

Sign of bidder :-

Date :-

Name of the bidder :-

Seal of the bidder :-



ANNEXURE-II

Ref. No. & Date: -

Quotation Notice No.: -

FINANCIAL BID

Purchase of “**Laboratory Vertical Autoclave**” for Department of Pathology & Lab Medicine at AIIMS Raipur. Details of item are given as under:

S. N.	Description of item(s)	Qty./ Unit	Offered Specification	Specify Technically Complied/Not Complied / if any Deviations
1.	The Vertical Autoclave should have Chamber Capacity of 70 or more liter.			
2.	The chamber of Vertical Autoclave should be manufactured as per ASME standards.			
3.	The Vertical Autoclave should work on the domestic power supply of : 230 V AC, 50 HZ, Single phase.			
4.	The unit's internal chamber, cover lid and all wetted parts should be fabricated from stainless steel of 304 grade.			
5.	The Vertical Autoclave's all joints should be smooth finished for crevice free internals.			
6.	The outer body should be made up of MS sheet with heat cured epoxy coating on both sides.			
7.	The lid should be equipped with single lever lock mechanism and lever handle moulded from industrial plastic.			
8.	The lid should be provided with auto purge cum vacuum breaker valve and a manually operable valve for exhaust.			
9.	The unit should have a solenoid valve for auto purging of air & normal exhaust.			
10.	The Vertical Autoclave should have stainless steel pressure gauge with dual range dial display in KPA and PSI along with a co-related temperature scale for steam in degrees Celsius.			
11.	The operations of the unit should be controlled by a microprocessor based controller.			
12.	The user should be able to set the temperature up to 122°C.			
13.	The autoclave should have timer facility.			
14.	In case of abrupt power failure / switching off, the Last Cycle which was set should remain in the memory.			
15.	The display for the parameters should be Two lines Alpha-Numeric digital display.			
16.	The unit should be equipped with Low Water Detection unit and should give Audio- Visual alarm in case of Low Water in the chamber and cut off the supply to the heater.			
17.	The unit should have safety valve to protect the equipment in case of over pressurization.			
18.	The Lid should be equipped with pressure interlock device. Also the heater should not start if the Lid is open.			
19.	The unit should be provided with safety cut-out for high temperature.			
20.	The unit should give indication by audio-visual alarm on completion of set autoclave cycle.			
21.	The electrical safety should be ensured by inbuilt MCB.			
22.	The unit should be mounted on 04 Nos. PU coated castors out of which at least 2 should have locking mechanism			
23.	Warranty of 03 years should be provided.			
24.	After sale, service should be available promptly.			
25.	Should be CE or FDA or BIS approved product.			
26.	Support for induction and follow up training of technical staff, on-site standardization and troubleshooting of procedures/ tests to be provided by the company.			

Sign of bidder :-

Date :-

Name of the bidder :-

Seal of the bidder :-